



# National Center for Mental Health

June 26, 2024

## NOTICE OF AWARD

**OMNIBUS BIO-MEDICAL SYSTEMS, INC.**

4<sup>th</sup> Floor Wilson Corporate Center  
225 Wilson St., Greenhills, San Juan  
Tel No: 8727-1058; 8722-4607  
Email: sales@omnibus.ph

Dear Sir/Madam,

Please be advised that as a result of **Public Bidding** conducted by National Center for Mental Health for the **Supply, Delivery, Installation, Testing and Commissioning of Various Brand-New Equipment for Clinical Laboratory Section CY 2024** and is hereby awarded to your company the following:

NO	ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
1	CL-08-2024	<p><b>Item No. 8. Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Incubator, Microbiological/General Incubator</b></p> <p>Brand / Model Offered / Country of Origin: <b>PHCbi / MIR-H263-PK / JAPAN</b></p> <p><b>SPECIFICATIONS:</b></p> <ol style="list-style-type: none"> <li>153 liters capacity</li> <li>Passive air movement</li> <li>Digital control display of temperature and time</li> <li>Adjustable temperature from ambient +5°C - 80°C</li> <li>Tempered safety glass</li> <li>Touch button operating panel</li> <li>Stainless steel chamber and shelves material</li> <li>With 3 shelves</li> <li>Alarms and self diagnostic function</li> <li>Power: Single phase, 100V-240V, 50Hz/60Hz</li> <li>With 2KVA Automatic Voltage Regulator</li> </ol> <p><b>STANDARD REQUIREMENTS:</b></p> <ol style="list-style-type: none"> <li>Current and valid certificate of authorized or exclusive distributorship;</li> <li>Notarized current and valid certificate of Manufacturer's compliance with ISO 13485 and/or from distributor;</li> </ol>	1	UNIT	Php 212,000.00	Php 212,000.00

*"There is no Health without Mental Health"*

Send dated July 09, 2024

	<ol style="list-style-type: none"> <li>3. Notarized certification from the manufacturer stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall;</li> <li>4. Notarized certification guaranteeing the availability on the supply of spare parts ten (10) years from the end of production. A certification must be issued by the manufacturer for the bidder;</li> <li>5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years;</li> <li>6. Notarized certification that the system or machine is not a retrofit solution;</li> <li>7. Notarized manufacturer's approved US FDA Permanent Notification (PMN) or CE mark approval certificate;</li> <li>8. The winning bidder must provide current and valid calibration certificate for each equipment during delivery;</li> <li>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder;</li> <li>10. Notarized certification from the manufacturer authenticated by the Philippine Consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor;</li> <li>11. Notarized warranty Certification: Certification that the bidder / supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user;</li> <li>12. Notarized certification that the bidder shall provide preventive maintenance and/or calibration schedule within warranty period;</li> <li>13. Notarized certification that the supplier / bidder shall provide applications training for the user and maintenance personnel of the hospital;</li> <li>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment;</li> <li>15. Notarized certification that the brand has been in the local market for at least five (5)</li> </ol>					
--	--	--	--	--	--	--

		<p>years with at least five (5) current installations;</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier;</p> <p>17. Certification that the supplier / bidder shall provide free installation of the equipment;</p> <p>18. With 24/7 Technical Support.</p> <p>19. Delivery period – 60 calendar days.</p>				
	CL-08-2024	<p><b>Item No. 9. Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Incubator, CO2)</b></p> <p>Brand / Model Offered / Country of Origin: <b>PHCbi / MCO-230AIC-PK / JAPAN</b></p> <p><b>SPECIFICATIONS:</b></p> <ol style="list-style-type: none"> <li>1. 230 liters capacity</li> <li>2. LCD Touch Panel Controller/ Touchscreen controller (deliver full control over different protocols)</li> <li>3. Integrated tray catches to accommodate more culture containers</li> <li>4 Self- diagnostic system for errors with visual and audible alarms</li> <li>5. Notification and alarm Functions: <ul style="list-style-type: none"> <li>• Temperature setpoint deviations (with notification and alarm for temperature deviation)</li> <li>• CO2 concentration setpoint deviations (with notification and alarm for CO2 deviation)</li> <li>• Other deviations outside normal machine operations (with notification and alarm for other abnormal operations)</li> </ul> </li> <li>6. With temperature safety device</li> <li>7. With chamber controller for controls of the Temperature in °C and Carbon Dioxide concentration</li> <li>8. Tempered inner safety glass door</li> <li>9. Stainless steel chamber and shelves material</li> <li>10. Adjustable temperature from ambient +5C to 50C</li> <li>11. Temperature control uniformity - as per manufacturer's standard</li> <li>12. CO2 control range and deviation - Up to 53% CO2 range</li> <li>13. Vertical/natural air flow (Passive air movement)</li> <li>14. Automatic decontamination protocol</li> </ol>	1	UNIT	Php 570,000.00	Php 570,000.00

15. Must come with CO2 tank filled, regulator and tubing
16. Power: Single phase, 100V-240V, 50Hz/60Hz
17. With 2KVA Automatic Voltage Regulator

**STANDARD REQUIREMENTS:**

1. Current and valid certificate of authorized or exclusive distributorship;
2. Notarized current and valid certificate of Manufacturer's compliance with ISO 13485 and/or from distributor;
3. Notarized certification from the manufacturer stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall;
4. Notarized certification guaranteeing the availability on the supply of spare parts ten (10) years from the end of production. A certification must be issued by the manufacturer for the bidder;
5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years;
6. Notarized certification that the system or machine is not a retrofit solution;
7. Notarized manufacturer's approved US FDA Permanent Notification (PMN) or CE mark approval certificate;
8. The winning bidder must provide current and valid calibration certificate for each equipment during delivery;
9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder;
10. Notarized certification from the manufacturer authenticated by the Philippine Consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor;
11. Notarized warranty Certification: Certification that the bidder / supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user;
12. Notarized certification that the bidder shall provide preventive maintenance and/or

	<p>calibration schedule within warranty period;</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the user and maintenance personnel of the hospital;</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment;</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations;</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier;</p> <p>17. Certification that the supplier / bidder shall provide free installation of the equipment;</p> <p>18. With 24/7 Technical Support.</p> <p>19. Delivery period – 60 calendar days.</p>			
<b>GRAND TOTAL</b>				<b>Php 782,000.00</b>

You are hereby required to provide the following within ten (10) calendar days:


- Notarized Contract Agreement Form (legal size) for the Procurement of **Supply, Delivery, Installation, Testing and Commissioning of Various Brand-New Equipment for Clinical Laboratory Section CY 2024;**  
*Note: Signatory: NOEL V. REYES, MD, FPPA, MMHoA, Medical Center Chief II;* and
- Performance Security in any forms and amount stipulated below:

A. Cash / Cashier's Manager's Check Bank Guarantee	Five Percent (5%) of the Total Contract Price
B. Surety Bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission / GSIS	Thirty Percent (30%) of the Total Contract Price

Failure to provide any of the above shall constitute sufficient ground for cancellation of the Award and forfeiture of the Bid Security.

Truly yours,  


**NOEL V. REYES, MD, FPPA, MMHoA**  
 Medical Center Chief II

CONFORMED BY:   
 DATE AND TIME: JULY 9, 2024