


REPUBLIC OF THE PHILIPPINES Department Of Health





July 22, 2024

NOTICE OF AWARD

FIA MEDICAL SUPPLY INC.

Lot 11&12 Block 8 Road 6-A Area 2 UPS 5

Brgy. San Isidro, Parañaque City

Tel. No.: (02) 8840-3552

Email: mikee.flamed@gmail.com

Dear Sir/Madam,

Please be advised that as a result of Public Bidding conducted by National Center for Mental Health for the Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Bio Refrigerator for the use of Pharmacy Section CY 2024 (LOT) and is hereby awarded to your company the following:

NO.	ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
1	PS- 01- 2024	BIO REFRIGERATOR SINGLE DOOR	1	unit	₱ 280,000.00	₱ 280,000.00
	2024	Country of Origin / Brand / Model: China / HAIER / HYC-509R				
		SPECIFICATIONS: - WITH UPS - 509 LITERS - TEMP RANGE 2-8 C - HGT 78 INCH (77.7 exterior) - WIDTH 27.3INCH (27.1 exterior) - DEPTH 32 INCH (31.9 exterior) - WEIGHT 215KG - RFID TECHNOLOGY - GLASS DOOR - Delivery and installation to Pharmacy Pavilion 7. - Inverter				
		STANDARD REQUIREMENTS 1) Notarized/CTC of current and valid certificate of manufacturer's compliance with ISO 13485 and/or CE certificate or its equivalent. 2) Notarized certification that the unit being offered shall be brand new. 3) Notarized/CTC of certificate of exclusive or authorized			S	

"There is no Health without Mental Health"





	distributorship from the manufacturer.
	4) Notarized certification that the
	supplier/bidder shall provide
	applications training for users and
	maintenance personnel of the
	hospital.
	5) Notarized certification that the
	supplier/bidder shall be responsible
	for the transportation to site,
	delivery, installation, and testing on
	the site and expenses for such will
	on the account of the supplier.
	6) Notarized certification that the
	supplier/bidder has the capability
	for the corrective and preventive
	maintenance of the unit, to include
	service engineers and technicians.
	7) Notarized/CTC of bidder's
	certification that the parts shall be
1	available at the authorized
	Philippine service centers for a
	period of five (5) years after the
	warranty period.
	8) Notarized warranty certificate: one
	(1) year for parts and two (2) years
	for service from date of delivery,
	inspection and acceptance.
	9) Notarized certification that the
	supplier/bidder shall provide free
	installation of the equipment.
	10) Notarized certification that the
	brand has been in the local market
	for at least two (2) years with at
	least one (1) installation of the
	equipment in the local market.
	11) Notarized certification to provide
	user's Manual in English language
	A. Operations Manual
	B. Service Manual
2	BIO REFRIGERATOR TWO- 1 unit ₱ 418,000.00 ₱ 418,000.00
	DOOR
	Country of Origin / Brand / Model:
	China / HAIER / HYC-1378
	SPECIFICATIONS:
	- WITH UPS
	- 1440 x 830-925 x 1950-1980 mm
	(78 inches / 1990mm height
	880mm / 1450 mm width)
	- 1364-1378 liters
	1501 1570 Htels

- 248-360 kg
- +2 to +14 °C (at fridge part center, AT -5 to 0°C)
- +2 to +23 °C (at fridge part center, AT 0 to 35°C)
- Microprocessor with non-volatile memory (Safety Lock with keypad)
- LED Thermistor
- Forced cool air circulation
- Forced type (cycle defrost)
- Fully automatic
- HFC and HFO combined (CFC-Free)
- Rigid polyurethane foamed insulation (CFC-free)
- = 2 doors
- double pane glass
- self-closing
- 8-12 wire shelves (polyethylene-coated, 530 x 604 mm, adjustable)
- max load 50 kg
- Delivery and installation to Pharmacy Pavilion 2.
- Inverter

STANDARD REQUIREMENTS

- Notarized/CTC of current and valid certificate of manufacturer's compliance with ISO 13485 and/or CE certificate or its equivalent.
- 2) Notarized certification that the unit being offered shall be brand new.
- Notarized/CTC of certificate of exclusive or authorized distributorship from the manufacturer.
- Notarized certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital.
- 5) Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation, and testing on the site and expenses for such will on the account of the supplier.
- 6) Notarized certification that the supplier/bidder has the capability for the corrective and preventive maintenance of the unit, to include service engineers and technicians.

.,,	GRAND TOTAL: ₱ 6	98,000.00
	B. Service Manual	
	A. Operations Manual	
	user's Manual in English language	
	11) Notarized certification to provide	
	equipment in the local market.	
	least one (1) installation of the	
	for at least two (2) years with at	
	brand has been in the local market	
	10) Notarized certification that the	
	installation of the equipment.	
	supplier/bidder shall provide free	
	9) Notarized certification that the	
	inspection and acceptance.	
	for service from date of delivery,	
	(1) year for parts and two (2) years	
	8) Notarized warranty certificate: one	
	warranty period.	
	period of five (5) years after the	
	Philippine service centers for a	
	available at the authorized	
	certification that the parts shall be	
	7) Notarized/CTC of bidder's	

You are hereby required to provide the following within ten (10) calendar days:

Notarized Contract Agreement Form (legal size) for the Procurement of Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Bio Refrigerator for the use of Pharmacy Section CY 2024 (LOT);
Note: Signatory: NOEL V. REYES, MD, FPPA, MMHoA, Medical Center Chief II; and

Performance Security in any forms and amount stipulated below:

A. Cash / Cashier's Manager's Check Bank	Five Percent (5%) of the Total Contract Price			
Guarantee				
B. Surety Bond callable upon demand issued by a surety or insurance company	Thirty Percent (30%) of the Total Contract Price			
duly certified by the Insurance Commission / GSIS				

Failure to provide any of the above shall constitute sufficient ground for cancellation of the Award and forfeiture of the Bid Security.

Truly yours,

NOEL V. REYES, MD, FPPA, MMHoA

Medical Center Chief II

CONFORMED BY:

DATE AND TIME: VILY 30, 2624