



National Center for Mental Health

NOTICE OF AWARD

May 29, 2023

MRL CYBERTEC CORP

3/F MRL Tower No. 124 Malakas St., Central District,
Diliman, Quezon City, Philippines 1100
Tel. Nos.: (632) 441-0398 / 441-0399 / 441-0400
Fax No.: (632) 441-0478 / 441-0429 / 434-7628
P.O. Box AC 550 – ACPO Cubao

Sir / Madam:

Please be advised that an Alternative Methods of Procurement – Direct Contracting is used by the National Center for Mental Health – Bids and Awards Committee for Contract Services and Consultancy for the **PREVENTIVE MAINTENANCE AND CALIBRATION OF VARIOUS ANATOMICAL LABORATORY EQUIPMENT CY 2023** and is hereby awarded to your company the following:

ITEM / DESCRIPTION	QTY	UNIT OF MEASURE	UNIT PRICE	TOTAL PRICE
<p>One (1) Year Preventive Maintenance / Calibration of Laboratory (ANATOMICAL) Equipment CY 2023 – Semi-Annual visit (Twice a year, every six (6) months).</p> <p>“Leica” TP 1020 Tissue Processor with S/N: 4258....</p> <p>Scope of Work:</p> <ol style="list-style-type: none"> 1. Checking of wax batch 11, 12 temperature calibrations. 2. Lubrication of mechanical moving parts. 3. Checking carousel alignment/calibrations. 4. Functionality/performance test of the control panel. 5. Cleaning of exterior / interior of the unit. 	1	Unit	19,000.00	19,000.00
<p>“Leica” RM2125 Microtome with S/N: 11981....</p> <p>Scope of Work:</p> <ol style="list-style-type: none"> 1. Mechanical parts lubrication. 2. Checking of micrometer and interior of the unit 	1	Unit	23,000.00	23,000.00

“There is no Health without Mental Health”



<ul style="list-style-type: none"> 3. Cleaning of exterior and interior of the unit 4. Functionality / Performance test of <ul style="list-style-type: none"> 1.1 Smoothness of operation of handwheel / positioning wheel 1.2 Display 1.3 Mechanical course feed and trimming steps 1.4 Clamping mechanisms of knife holder assembly 				
<p>"Leica" CM 1850 Cryostat with S/N: 2068....</p> <p>Scope of Work:</p> <ul style="list-style-type: none"> 1. Checking of Calibrations: <ul style="list-style-type: none"> 1.1 Cabinet Temperature 1.2 Mechanical System 1.3 Cutting system of micrometer 2. Functionality test of display, peltier button, temp-set buttons, clock set buttons, defrost buttons, lock button, light, coarse feed buttons. 3. Cleaning and checking of refrigeration systems: <ul style="list-style-type: none"> 1.1 Interior and exterior of condenser 1.2 Main Unit 	1	Unit	21,000.00	21,000.00
<p>"MORTECH" Morgue Ref with S/N: 7467....</p> <p>Scope of Work:</p> <ul style="list-style-type: none"> 1. Checking, Alignment and lubrications of mechanical system of bed carriers. 2. Cleaning, Checking of calibration of refrigeration system (condenser unit, evaporation unit, blower). 3. Cleaning of outside surface (Disinfection). 	1	Unit	37,000.00	37,000.00
<p>"MORTECH" Morgue Ref with S/N: 8103....</p> <p>Scope of Work:</p> <ul style="list-style-type: none"> 1. Checking, Alignment and lubrications of mechanical system of bed carriers. 2. Cleaning, Checking of calibration of refrigeration system (condenser unit, evaporation unit, blower). 3. Cleaning of outside surface (Disinfection). 	1	Unit	36,960.00	36,960.00

	AMOUNT:	Php136,960.00
	DELIVERY:	7 DAYS
	TERMS:	30 DAYS
	VALIDITY:	30 DAYS
	TOTAL A MOUNT:	Php136,960.00


Truly yours,



NOEL V. REYES, MD, FPPA, MMH0A
 Medical Center Chief II

↑

CONFORME:



EDGARDO S. BARCOZO
 Authorized Representative

Date and Time: June 15, 2023 @ 4:48pm