



REPUBLIC OF THE PHILIPPINES
Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
Nueve de Febrero Street, Mandaluyong City, Philippines
BIDS AND AWARDS COMMITTEE



Telephone No. 531-9001 loc. 239 Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

Section VI

Schedule of Requirements

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF VARIOUS BRAND-NEW EQUIPMENT FOR ANATOMICAL LABORATORY SECTION CY 2024 (LINE ITEM)
IB No. E-002-2024-PB

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item No.	Description	Qty	Delivery Site			Delivery Period and Terms of Payment
			OFFICE	FACILITY	ADDRESS	
1	<p>Supply, Installation, Testing and Delivery, Commissioning of a Brand-New Tissue Embedding Machine</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> Two (2) piece modular system consisting of a cold plate and paraffin dispensing station. Cold plate can be moved left or right, with at least sixty (60) embedded blocks With standard magnifying glass and equipped with LED working light. Programmable touchscreen display with easy to navigate onscreen control. With interchangeable cassette and mold tray. At least four (4) liters paraffin tank With foot pedal switch for paraffin dispensing function. 	1	Anatomical Laboratory Section	NCMH New Laboratory Building (Histopathology Unit)	National Center for Mental Health	<ol style="list-style-type: none"> One-time delivery within thirty (30) calendar days upon receipt of Notice to Deliver Terms of payment at least 30 days from receipt of Sales invoice

	<p>8. Power supply; 220V,50/60Hz</p> <p>9. With manual operation</p> <p>10. Installation and initial Operation onsite</p> <p>11. With manpower training</p> <p>12. Reliable, prompt technical support with trained engineers.</p> <p>STANDARD ACCESSORIES:</p> <p>1. AVR (Automatic Voltage Regulator)</p> <p>2. With at least twenty-four (24) stainless embedding molds</p> <p>3. Two (2) sets of Tissue forceps consist of 1mm tip, 2mm tip, 4mm tip</p> <p>STANDARD REQUIREMENTS:</p> <p>1. Current and valid certificate of manufacturer's compliance with ISO 13485 and/or CE certificate or its equivalent</p> <p>2. Certification that the unit being offered shall be brand new</p> <p>3. Certificate of exclusive or authorized distributorship from the manufacturer.</p> <p>4. Certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital.</p> <p>5. Certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be</p>									
--	--	--	--	--	--	--	--	--	--	--

2	<p>on the account of the supplier/bidder.</p> <p>6. Certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit to include service engineers and technicians.</p> <p>7. Certification to provide users' Manual in English language.</p> <p>a. Operations Manual - 2 copies</p> <p>b. Service Manual - 2 copies</p> <p>8. Bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period.</p> <p>9. Warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance.</p> <p>10. Certification that the supplier/bidder shall provide free installation of the equipment.</p> <p>11. Certification that the brand has been in the local market for at least two (2) years with at least one (1) installation of the equipment in the local market.</p> <p>12. Delivery period - 30 calendar days</p> <p>13. To be Installed at NCMH New Laboratory Building (Histopathology Unit)</p>	1	Anatomical Laboratory Section	NCMH New Laboratory Building (Histopathology Unit)	National Center for Mental Health	a. One-time delivery within thirty (30) calendar days upon receipt of
---	---	---	-------------------------------	--	-----------------------------------	--

	<p>1. Floor type ductless ventilated hood system</p> <p>2. Exhaust ventilation system is equipped with the following:</p> <p>a.) Centrifugal Blower fan with adjustable speed</p> <p>b.) Air face velocity should be within 0.4 to 0.8 m/s</p> <p>c.) Activated carbon filter</p> <p>3. Chemically resistant steel frame and working plane with polypropylene spillage tray.</p> <p>4. Movable tempered glass front shield</p> <p>5. Four (4) heavy duty casters, 2 front wheels are swivel type with brake.</p> <p>6. Built-in electronic control panel with LCD touch screen display.</p> <p>7. Fluorescent Lamp or LED lighting</p> <p>8. UV light for decontamination</p> <p>9. Water proof electric socket</p> <p>10. Audio and visual alarm for charging filter.</p> <p>11. Dimension:</p> <p>Width - 1.0 to 1.2 m</p> <p>Depth- 0.6 to 0.8 m</p> <p>Height - 2.0 to 2.2 m</p> <p>12. Power Supply: 220-240V/50-60Hz</p> <p>STANDARD ACCESSORIES:</p> <p>1. AVR (Automatic Voltage Regulator)</p>		<p>Notice to Deliver</p> <p>b. Terms of payment at least 30 days from receipt of Sales invoice</p>
--	--	--	---

	<p>STANDARD REQUIREMENTS:</p> <ol style="list-style-type: none"> 1. Current and valid certificate of manufacturer's compliance with ISO 13485 and/or CE certificate or its equivalent 2. Certification that the unit being offered shall be brand new 3. Certificate of exclusive or authorized distributorship from the manufacturer. 4. Certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital. 5. Certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder. 6. Certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit to include service engineers and technicians. 7. Certification to provide users' Manual in English language. <ol style="list-style-type: none"> a. Operations Manual - 2 copies b. Service Manual - 2 copies 8. Bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 9. Warranty certificate: one (1) year for parts and two (2) years for service from date of 				
--	--	--	--	--	--

	<p>delivery, inspection and acceptance.</p> <p>10. Certification that the supplier/bidder shall provide free installation of the equipment.</p> <p>11. Certification that the brand has been in the local market for at least two (2) years</p> <p>12. Delivery period - 30 calendar days</p> <p>13. To be Installed at NCMH New Laboratory Building (Histopathology Unit)</p>				
3	<p>Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Automatic Tissue Processor</p> <p>SPECIFICATIONS:</p> <p>1. Carousel type with twelve (12) stations consisting of ten (10) reagent stations and two (2) paraffin wax bath stations.</p> <p>2. All containers of the twelve (12) stations: a) shall be made of Aluminum metal with handle. b) shall have a capacity of 1.5 liters up to 2.5 liters. c) must be able to withstand vacuum with maximum pressure difference of 500 hpa (approximate 0.5 bar).</p> <p>3. The tissue baskets shall be made of aluminum or stainless steel with a minimum capacity of 100 cassettes.</p> <p>4. Shall have a microprocessor control with LCD screen displaying the following: indication of station, program parameters such as numbers of infiltration time, and real-time over-all run time.</p>	1	Anatomical Laboratory Section	NCMH New Laboratory Building (Histopathology Unit)	National Center Mental Health
					<p>a. One-time delivery within thirty (30) calendar days upon receipt of Notice Deliver to</p> <p>b. Terms of payment at least 30 days from receipt of Sales invoice</p>

5. Shall have vacuum and agitation functions for better tissue processing.

6. Shall have manual and automatic operating modes.

7. Shall have a minimum of 9 freely selectable programs.

8. Shall have audible alarm and error messages.

9. With fume extraction system, enclosed type which includes the following components:

a) fume containment shields.

b) two activated carbon filters dedicated for formalin and for xylene/alcohol.

c) fume extraction fan with exhaust air tube.

10. Power supply; 220V,50/60Hz

11. With manpower training and certification.

12. With manual of operation.

13. Installation and initial Operation onsite

STANDARD ACCESSORIES:

1. AVR (Automatic Voltage Regulator)

2. Two (2) Boxes of tissue cassettes

STANDARD REQUIREMENTS:

1. Current and valid certificate of manufacturer's compliance with ISO 13485 and/or CE certificate or its equivalent

	<p>2. Certification that the unit being offered shall be brand new</p> <p>3. Certificate of exclusive or authorized distributorship from the manufacturer.</p> <p>4. Certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital.</p> <p>5. Certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder.</p> <p>6. Certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit to include service engineers and technicians.</p> <p>7. Certification to provide users' Manual in English language.</p> <p>a. Operations Manual - 2 copies</p> <p>b. Service Manual - 2 copies</p> <p>8. Bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period.</p> <p>9. Warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance.</p> <p>10. Certification that the supplier/bidder shall provide free installation of the equipment.</p>				
--	---	--	--	--	--

	<p>11. Certification that the brand has been in the local market for at least two (2) years with at least one (1) installation of the equipment in the local market.</p> <p>12. Delivery period - 30 calendar days</p> <p>13. To be Installed at NCMH New Laboratory Building (Histopathology Unit)</p>					
CONFORME:						
<i>(Company Name)</i>						
<i>(Name and Signature of Authorized Representative)</i>						