



**BIDS AND AWARDS COMMITTEE**

Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

**LIST OF ITEMS**

PUBLIC RE-BIDDING FOR THE SUPPLY, DELIVERY INSTALLATION, TESTING AND COMMISSIONING OF VARIOUS BRAND-NEW EQUIPMENT FOR CLINICAL LABORATORY SECTION CY 2024 (LINE ITEM)

ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
RCL01-2024	<p><b>Hemoglobinometer</b>  <b>Specifications:</b>            1. Compact and lightweight            2. Portable            3. With digital display            4. Finger prick, spot testing analyzer with micro cuvette technology            5. Should have LED/LCD display of hemoglobin in g/l or g/dl            6. Instrument should be able to work in hot climate - as per manufacturer's standard            7. The winning bidder shall provide startup kits for at least 50 tests</p> <p><b>Standard Requirements:</b>            1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship.            2. Notarized/CTC certificate of current and valid of Manufacturer's compliance with ISO 13485.            3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.            4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be issued by the manufacturer for the bidder.            5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.</p>	1	unit	₱ 80,000.00	₱ 80,000.00

	<p>6. Notarized certification that the system or machine is not a retrofit solution</p> <p>7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.</p> <p>8. The winning bidder shall provide current and valid calibration certificate for each equipment during delivery</p> <p>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.</p> <p>10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship</p> <p>11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>18. With 24/7 Technical Support</p>				
--	---	--	--	--	--

RCL02-2024	<p>19. Delivery period – 90 calendar days</p> <p><b>Multiplex PCR system capable of panel testing</b></p> <p><b>Specifications:</b></p> <ol style="list-style-type: none"> <li>1. Molecular Multiplex PCR Machine</li> <li>2. Capable of running Pneumonia, blood culture, meningitis, gastrointestinal, respiratory and joint infection panels</li> <li>Respiratory panel: Can detect Chlamydia pneumoniae.</li> <li>Meningitis Panel: Can detect Cytomegalovirus (CMV)</li> <li>Pneumonia Panel: Can run sputum/ETA/BAL samples, includes detection of AMR genes</li> <li>Blood Culture: Includes detection of yeast, includes detection of AMR genes</li> <li>3. Panels are FDA approved.</li> <li>4. Integrates sample preparation, nucleic acid extraction, amplification, and detection in 60 minutes testing time</li> <li>5. Can detect bacteria, viruses, yeasts and antimicrobial resistant genes</li> <li>6. LIS ready</li> </ol> <p><b>Standard Requirements:</b></p> <ol style="list-style-type: none"> <li>1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship.</li> <li>2. Notarized/CTC certificate of current and valid of Manufacturer's compliance with ISO 13485.</li> <li>3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.</li> <li>4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be issued by the manufacturer for the bidder.</li> <li>5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.</li> <li>6. Notarized certification that the system or machine is not a retrofit solution</li> <li>7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.</li> <li>8. The winning bidder shall provide current and valid calibration certificate for each equipment during delivery</li> </ol>	1	unit	₱ 4,300,000.00	₱ 4,300,000.00
------------	--	---	------	----------------	----------------

	<p>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.</p> <p>10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship</p> <p>11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>18. With 24/7 Technical Support</p> <p>19. Delivery period – 90 calendar days</p>				
RCL03-2024	<p><b>Agglutination viewer</b>  <b>Specifications:</b></p> <p>1. Designed for use in blood typing, crossmatching, prothrombin time, kahn and other agglutination, flocculation, and serodiagnostic test.</p> <p>2. Illuminates test tube contents through the use of at least 5 watt bulb from above and a magnifying mirror from below.</p>	1	unit	₱ 45,000.00	₱ 45,000.00

3. Flexible neck allows for easy manipulation of light source.
4. Include Bulb, magnifying mirror, Flexible mirror, side mounted on/off switch, flexible neck, mainly operator.
5. Hertz: 50/60Hz
6. Voltage: 220v
7. Weight: 3kg
8. Amps: 36mA
9. Temperature: 18C to 33C10.

**Standard Requirements:**

1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship.
2. Notarized/CTC certificate of current and valid of Manufacturer's compliance with ISO 13485.
3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.
4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be issued by the manufacturer for the bidder.
5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.
6. Notarized certification that the system or machine is not a retrofit solution
7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.
8. The winning bidder shall provide current and valid calibration certificate for each equipment during delivery
9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.
10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship
11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive

	<p>maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>18. With 24/7 Technical Support</p> <p>19. Delivery period – 90 calendar days</p>				
			<b>GRAND TOTAL:</b>		<b>₱ 4,425,000.00</b>

Submitted by: 

**FRANK ANTHONY E. PANGANIBAN, MD, MMHoA, DPSP**  
 OIC, Clinical Laboratory

Recommending Approval:

**-SGD-**

**ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA**  
 Chairperson, BAC for Equipment

Approved By:

**-SGD-**

**NOEL V. REYES, MD, FPPA, MMHoA**  
 Medical Center Chief II