



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

Section VI

Schedule of Requirements

**PUBLIC RE-BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF VARIOUS BRAND-NEW EQUIPMENT FOR CLINICAL LABORATORY SECTION CY 2024 (LINE ITEM)
IB No. E-003-2024-PRB**

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item No.	Description	Qty	Delivery Site			Delivery Period and Terms of Payment
			OFFICE	FACILITY	ADDRESS	
1	<p>Hemoglobinometer</p> <p>Specifications:</p> <ol style="list-style-type: none"> 1. Compact and lightweight 2. Portable 3. With digital display 4. Finger prick, spot testing analyzer with micro cuvette technology 5. Should have LED/LCD display of hemoglobin in g/l or g/dl 6. Instrument should be able to work in hot climate - as per manufacturer's standard 7. The winning bidder shall provide startup kits for at least 50 tests <p>Standard Requirements:</p> <ol style="list-style-type: none"> 1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship. 2. Notarized/CTC certificate of current and valid of 	1	Clinical Laboratory Section	NCMH New Laboratory Building	National Center for Mental Health	One-time delivery within ninety (90) calendar days upon receipt of Notice to Deliver

Handwritten signature

<p>Manufacturer's compliance with ISO 13485.</p> <p>3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.</p> <p>4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be issued by the manufacturer for the bidder.</p> <p>5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.</p> <p>6. Notarized certification that the system or machine is not a retrofit solution</p> <p>7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.</p> <p>8. The winning bidder shall provide current and valid calibration certificate for each equipment during delivery</p> <p>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.</p> <p>10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship</p> <p>11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and</p>					
--	--	--	--	--	--

	<p>services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>18. With 24/7 Technical Support</p> <p>19. Delivery period – 90 calendar days</p>					
2	<p>Multiplex PCR system capable of panel testing</p> <p>Specifications:</p>	1	Clinical Laboratory Section	NCMH New Laboratory Building	National Center for Mental Health	One-time delivery within ninety (90) calendar days upon receipt of

<p>1. Molecular Multiplex PCR Machine</p> <p>2. Capable of running Pneumonia, blood culture, meningitis, gastrointestinal, respiratory and joint infection panels Respiratory panel: Can detect Chlamydia pneumoniae. Meningitis Panel: Can detect Cytomegalovirus (CMV) Pneumonia Panel: Can run sputum/ETA/BAL samples, includes detection of AMR genes Blood Culture: Includes detection of yeast, includes detection of AMR genes</p> <p>3. Panels are FDA approved.</p> <p>4. Integrates sample preparation, nucleic acid extraction, amplification, and detection in 60 minutes testing time</p> <p>5. Can detect bacteria, viruses, yeasts and antimicrobial resistant genes</p> <p>6. LIS ready</p> <p>Standard Requirements:</p> <p>1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship.</p> <p>2. Notarized/CTC certificate of current and valid of Manufacturer's compliance with ISO 13485.</p> <p>3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.</p> <p>4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be</p>					<p>Notice Deliver to</p>
--	--	--	--	--	---------------------------------

<p>issued by the manufacturer for the bidder.</p> <p>5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.</p> <p>6. Notarized certification that the system or machine is not a retrofit solution</p> <p>7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.</p> <p>8. The winning bidder shall provide current and valid calibration certificate for each equipment during delivery</p> <p>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.</p> <p>10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship</p> <p>11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training</p>								
---	--	--	--	--	--	--	--	--

	<p>for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>18. With 24/7 Technical Support</p> <p>19. Delivery period – 90 calendar days</p>					
3	<p>Agglutination viewer</p> <p>Specifications:</p> <p>1. Designed for use in blood typing, crossmatching, prothrombin time, kahn and other agglutination, flocculation, and serodiagnostic test.</p> <p>2. Illuminates test tube contents through the use of at least 5 watt bulb from above and a magnifying mirror from below.</p> <p>3. Flexible neck allows for easy manipulation of light source.</p> <p>4. Include Bulb, magnifying mirror, Flexible mirror, side mounted on/off switch,</p>	1	Clinical Laboratory Section	NCMH New Laboratory Building	National Center for Mental Health	One-time delivery within ninety (90) calendar days upon receipt of Notice to Deliver

<p>flexible neck, mainly operator.</p> <p>5. Hertz: 50/60Hz</p> <p>6. Voltage: 220v</p> <p>7. Weight: 3kg</p> <p>8. Amps: 36mA</p> <p>9. Temperature: 18C to 33C10.</p> <p>Standard Requirements:</p> <p>1. Notarized/CTC certificate of current and valid of authorized or exclusive distributorship.</p> <p>2. Notarized/CTC certificate of current and valid of Manufacturer's compliance with ISO 13485.</p> <p>3. Notarized certificate from the manufacturer or distributor stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.</p> <p>4. Notarized/CTC bidder's certificate guaranteeing the availability on the supply of spare parts five (5) years from the end of production. A certification must be issued by the manufacturer for the bidder.</p> <p>5. Notarized certification that the bidder must be in the local market for a minimum of ten (10) years.</p> <p>6. Notarized certification that the system or machine is not a retrofit solution</p> <p>7. Notarized manufacturer's approved US FDA Premarket Notification (PMN) or CE mark approval certificate. The machine should have a US FDA approval.</p> <p>8. The winning bidder shall provide current and valid calibration certificate for</p>					
---	--	--	--	--	--

<p>each equipment during delivery</p> <p>9. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. The certificate must be issued by the manufacturer to the bidder.</p> <p>10. Notarized certification from the manufacturer or distributor that the warranty should not be affected with change of distributorship</p> <p>11. Notarized Warranty Certification: Certification that the bidder/supplier shall provide a three (3) year warranty for parts and services that includes corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end-user</p> <p>12. Notarized certification that the bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>13. Notarized certification that the supplier / bidder shall provide applications training for the users and maintenance personnel of the hospital.</p> <p>14. Notarized certification to provide manuals: Two (2) sets of service manual in English Language and Two (2) sets of user manual in English Language upon delivery of the equipment.</p> <p>15. Notarized certification that the brand has been in the local market for at least five (5) years with at least five (5) current installations.</p> <p>16. Notarized certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site</p>					
--	--	--	--	--	--

<p>(hospital / health facility) expenses for such will be on the account of the supplier.</p> <p>✓ 17. Notarized certification that the supplier / bidder shall provide free installation of the equipment</p> <p>✓ 18. With 24/7 Technical Support</p> <p>✓ 19. Delivery period - 90 calendar days</p>					
CONFORME:					
<i>(Company Name)</i>					
<i>(Name and Signature of Authorized Representative)</i>					