



National Center for Mental Health

LIST OF ITEMS

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND – NEW ELECTROENCEPHALOGRAM (EEG) MACHINE FOR PAVILION 11- NEUROPHYSIOLOGY LABORATORY CY 2024

ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
EG01-2024	<p>Public Bidding for the Supply, Delivery, Installation, Testing and Commissioning of Brand-New Electroencephalogram (EEG) Machine CY 2024 for Pavilion 11-Neurophysiology Laboratory CY 2024</p> <p>SPECIFICATIONS: A. Software: EEG System/ Program <input type="checkbox"/> Can create significant information and clinical history of the patient <input type="checkbox"/> Can perform calibration and bio-calibration in the beginning and end of the recording <input type="checkbox"/> Able to encode/enter notation or comments during the recording <input type="checkbox"/> Montage, filter, sensitivity setting is available during the recording <input type="checkbox"/> Has an ability to check electrode impedance <input type="checkbox"/> End-user is able to create montage and photic stimulation according to their needs <input type="checkbox"/> Availability of high and low cut frequency and notch filter <input type="checkbox"/> Has an EKG display <input type="checkbox"/> Paper speed display must be 10 seconds per page</p>	1	unit	₱ 3,038,333.33	₱ 3,038,333.33

"There is no Health without Mental Health"

	<input type="checkbox"/> Has measurement tool for frequency, amplitude and duration				
	<input type="checkbox"/> Hyperventilation time can be set or update during the recording <input type="checkbox"/> Can create result/report templates <input type="checkbox"/> Automatically archives/save EEG recording <input type="checkbox"/> Can easily review the previously done recording <input type="checkbox"/> Operating System must be a minimum of Windows 10, 64 Bit <input type="checkbox"/> CPU must be a minimum of Intel Core 7, 1 TBB HDD, 8GB Ram"				
	<p>B. Hardware</p> <input type="checkbox"/> Amplifier or Jockbox must have at least 32 channels <input type="checkbox"/> Has a continuity check of electrodes <input type="checkbox"/> Camera must be pan, tilt, zoom IR IP with video program and real time recording <input type="checkbox"/> Amplifier or Jockbox must have at least 32 channels <input type="checkbox"/> Has a continuity check of electrodes <input type="checkbox"/> Camera must be pan, tilt, zoom IR IP with video program and real time recording <input type="checkbox"/> Has microphone and speakers <input type="checkbox"/> With Photic Stimulator and Amplifier stand <input type="checkbox"/> Has 1KV Uninterrupted Power Supply (UPS)				

	<input type="checkbox"/> Has trolley or cart with wheels that can cater the PC, printer and UPS"				
	<p>STANDARD REQUIREMENTS:</p> <ol style="list-style-type: none"> 1) Notarized / Certified True Copy (CTC) current and valid certificate of Authorized of Exclusive or Distributorship 2) Notarized / Certified True Copy (CTC) current and valid certificate of Manufacturer's compliance with ISO 13485 3) Notarized certificate from the manufacturer stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall. 4) Notarized bidder's certificate guaranteeing the availability on the supply of spare parts 10 (ten) years from end of production. A certification must be issued by the manufacturer for the bidder. 5) Notarized Certification that the bidder must be in the business in the local market for a minimum of 10 (ten) years. 6) Notarized certificate of manufacturer's approved US FDA Premarket Notification (PMN) or CE Mark Approval certificate. 7) The winning bidder shall provide current and valid calibration certificate for each equipment during delivery. 8) Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit. 9) Notarized certification from the manufacturer authenticated by the Philippine consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor. 				

	<p>10) Notarized warranty certification. Notarized Certification that the bidder/supplier shall provide a 3</p>				
	<p>(three) years warranty for parts and services that include corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end user.</p> <p>11) Bidder must provide preventive maintenance and/or calibration schedule within warranty period.</p> <p>12) Notarized certification that the supplier/bidder shall provide applications training and certification for end users and maintenance personnel of the hospital.</p> <p>Notarized certification to provide Manuals: Two (2) sets of Service Manual in English Language and Two (2) sets of User Manual in English Language upon delivery of the equipment.</p>				
	GRAND TOTAL:				₱ 3,038,333.33

Submitted by:

RAUL T. GARDAYA, MD, FPNA, DPSBP, FPPA
 Officer In Charge, Neurological Unit

Recommending Approval:

-SGD-

ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
 Chairperson, BAC for Equipment

Approved By:

-SGD-

NOEL V. REYES, MD, FPPA, MMHoA
 Medical Center Chief II