



REPUBLIC OF THE PHILIPPINES
 Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
 Nueve de Febrero Street, Mandaluyong City, Philippines



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: bacnemh@yahoo.com

Website: www.ncmh.gov.ph

LIST OF ITEMS

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW ELECTROCONVULSIVE THERAPY MACHINE FOR PAVILION 1 & 3 CY 2024
 IB No. E-010-2024-PB

ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
ECT - 2024	<p>Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Electroconvulsive Therapy Machine for Pavilion 1 & 3 CY 2024</p> <p>SPECIFICATION:</p> <p>1) 8 Stimulus Parameter Knobs 2) 200-300 Joules, 220V 60Hz 3) Standard Monitoring Channels: One (1) EEG & ECG per unit</p> <p>Power Requirement for Electrical/Electronic unit:</p> <p>A. Voltage: 220 – 240 volts B. Frequency: 60Hz</p> <p>Complete with the following features:</p> <p>* Optical Motion Sensor with OMS Cable and Sensor *EEG Data Analysis Feature (EEG/ECG) *Dual Hand-Held Electrodes with Remote Button</p> <p>Inclusions:</p> <p>* Textbook, APA Task Force Report – latest edition * Patient Stimulus Cable * Patient Safety Monitor Cable * Adjustable Headband * Flat Stimulus Electrodes * Concave Stimulus Electrodes</p>	2	unit	4,990,000.00	₱ 9,980,000.00

	<p>*Automatic Voltage Regulator for the Safety of the Unit</p> <p>*40 pieces Bite block</p> <p>*10 boxes Chart paper (10 rolls/box)</p>				
	<p>* 4 pieces Electrode gel</p> <p>* 4 pieces Electrode paste</p> <p>* EEG safety leads</p> <p>* ECG safety leads</p> <p>*200 pieces EEG disposable pads</p> <p>*200 pieces ECG disposable pads</p> <p>* Sensor strip box</p> <p>* Power Cord: Dual Insulated cable, at least 2 meters</p> <p>-Plug: 3 prong male/female plug, Hospital Grade</p> <p>STANDARD REQUIREMENTS</p> <p>A. CTC/Notarized certification of current and valid Authorized or Exclusive Distributorship.</p> <p>B. CTC/Notarized certificate of current and valid Manufacturer's compliance with ISO 13485.</p> <p>C. Notarized certification from the manufacturer stating that the equipment is brand new, unused and not a discontinued model or was listed in the market recall.</p> <p>D. Notarized certificate guaranteeing the availability on the supply of spare parts within six (6) years from end of production.</p> <p>E. Notarized certification that the bidder must be in the business in the local market for a minimum of five (5) years.</p> <p>F. Notarized certification that the system or machine is not a retrofit solution.</p> <p>G. Notarized Certificate of manufacturer's approved US FDA Premarket Notification (PMN) or CE Mark Approval.</p> <p>H. The winning bidder shall provide current and valid calibration certificate for the equipment during delivery.</p> <p>I. Notarized certification that the supplier has the capability or authority for corrective and preventive maintenance of the unit.</p> <p>J. Notarized certification from the manufacturer authenticated by the Philippine consulate from the country of origin of the unit that the</p>				



	<p>warranty should not be affected with a change of distributor.</p> <p>K. Notarized certificate that the bidder/supplier shall provide a three (3) years warranty for parts and services that include corrective maintenance, preventive maintenance, and/or calibration. The warranty shall commence upon the acceptance of the end user.</p> <p>L. Notarized certification the supplier/bidder shall provide applications training and certification for end users and maintenance personnel of the hospital.</p> <p>M. CTC/Notarized certification to provide Manuals: two (2) sets of Service Manual in English Language and two (2) sets of User Manual in English Language upon delivery of the equipment.</p> <p>1. Scope of Warranty:</p> <p>A. Warranty Replacement: In case of unit malfunction the bidder should replace the unit with a brand new unit within one (1) month from the start of the warranty period.</p> <p>B. Service and Parts Warranty: at least three (3) years for both service and parts.</p> <p>C. Warranty Certificates: Comprehensive Warranty Certificates, must be included and define in the contract.</p> <p>E. Provide Service Report per unit to End-Users.</p> <p>2. Acceptance and Maintenance</p> <p>A. Valid certificates of the Technicians/Engineers to conduct service/maintenance.</p> <p>B. Valid Certificates of calibration of the analyzer and testing equipment.</p> <p>C. List of Engineers/Technicians with their certificates to conduct service/maintenance</p> <p>D. List of Analyzers/Testing tools with their Brand/Model/Serial No. and its valid certificate of calibration.</p> <p>E. List of Analyzers/Testing tools with their Brand/Model/Serial No. and its valid certificate of calibration.</p> <p>3. Preventive Maintenance</p> <p>a. Conduct quarterly cleaning and testing all parameters including all accessories</p> <p>b. Conduct quarterly Qualitative and Quantitative Test</p> <p>c. Conduct quarterly Calibration</p> <p>d. Provide Service Report per unit. Calibration Certificates or equivalent</p>				
--	---	--	--	--	--

e. Report Findings, Suggestions and Recommendations.

4. Scope of Trainings:

a. At least 16 hours training for NCMH medical personnel.

b. Minimum of four (4) hours of basic trouble shooting and maintenance for the NCMH technical personnel.

GRAND TOTAL:

₱ 9,980,000.00

Submitted by:



KENNETH KRISTOFFERSON D. EUROPA, MD, FPPA
Medical Specialist III

Recommending Approval:



ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
Chairperson, BAC for Equipment

Approved By:



NOEL V. REYES, MD, FPPA, MMHoA
Medical Center Chief II