



REPUBLIC OF THE PHILIPPINES
 Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
 Nueve de Febrero Street, Mandaluyong City, Philippines



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: baencmh@yahoo.com

Website: www.ncmh.gov.ph

SECTION VII

Technical Specifications

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW ELECTROCONVULSIVE THERAPY MACHING FOR PAVILION 1 & 3 CY 2024
 IB No. E-010-2024-PB

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

Item	Specification	Statement of Compliance	Supporting Documents for the Statement of Compliance (if applicable)
	1. 8 Stimulus Parameter Knobs		
	2. 200-300 Joules, 220V 60Hz		
	3). Standard Monitoring Channels: One (1) EEG & ECG per unit		
	Power Requirement for Electrical/Electronic Unit:		
	1. Voltage: 220 – 240 volts		
	2. Frequency: 60Hz		
	Complete with the following features:		
	1. Optical Motion Sensor with OMS Cable and Sensor		

	2. EEG Data Analysis Feature (EEG/ECG)		
	3. Dual Hand-Held Electrodes with Remote Button		
	Inclusions:		
	1. Textbook, APA Task Force Report – latest edition		
	2. Patient Stimulus Cable		
	3. Patient Safety Monitor Cable		
	4. Adjustable Headband		
	5. Flat Stimulus Electrodes		
	6. Concave Stimulus Electrodes		
	7. Automatic Voltage Regulator for the Safety of the Unit		
	8. 40 pieces Bite block		
	9. 10 boxes Chart paper (10 rolls/box)		
	10. 4 pieces Electrode gel		
	11. 4 pieces Electrode paste		
	12. EEG safety leads		
	13. ECG safety leads		
	14. 200 pieces EEG disposable pads		
	15. 200 pieces ECG disposable pads		
	16. Sensor strip box		
	17. Power Cord: Dual Insulated cable, at least 2 meters -Plug: 3 prong male/female plug, Hospital Grade		

CONFORME:

(Company Name)

(Name and Signature of Authorized Representative)

