

## REPUBLIC OF THE PHILIPPINES Department Of Health



## National Center for Mental Health

LIST OF ITEMS

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW BIO REFRIGERATOR FOR THE USE OF PHARMACY SECTION CY 2024 (LOT)

IB No. E-012-2024-PB

ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
PS-01-	BIO REFRIGERATOR SINGLE	1	unit	₱ 300,000.00	₱ 300,000.00
2024	DOOR				.,
	SPECIFICATIONS:				
	- WITH UPS				
	- 509 LITERS - TEMP RANGE 2-8 C				
	- HGT 78 INCH				
	- WIDTH 27.3INCH				
	- DEPTH 32 INCH				
	- WEIGHT 215KG				
	- RFID TECHNOLOGY				
	- GLASS DOOR				
	- Delivery and installation to			1	
	Pharmacy Pavilion 7.				
	STANDARD REQUIREMENTS				
	1) Notarized/CTC of current and valid				
	certificate of manufacturer's				
1	compliance with ISO 13485 and/or	1			
	CE certificate or its equivalent.				
	2) Notarized certification that the unit				
	being offered shall be brand new.	1			
1	3) Notarized/CTC of certificate of			1	
1	exclusive or authorized				
1	distributorship from the			1	
	manufacturer.				
1	4) Notarized certification that the				
	supplier/bidder shall provide	1			
	applications training for users and	1			
l l	maintenance personnel of the				
	hospital.  5) Notarized certification that the				
48	,				
1	supplier/bidder shall be responsible for the transportation to site,				
	delivery, installation, and testing				
	on the site and expenses for such				
	will on the account of the supplier.			1	
1	6) Notarized certification that the				
1,	supplier/bidder has the capability				

32

"There is no Health without Mental Health"







1					
	for the corrective and preventive				
	maintenance of the unit, to include				
	service engineers and technicians.				
	7) Notarized/CTC of bidder's	,			
	certification that the parts shall be				
	available at the authorized				
	Philippine service centers for a				
	period of five (5) years after the				
	warranty period.				
	8) Notarized warranty certificate: one				
	(1) year for parts and two (2) years				
	for service from date of delivery,				
	inspection and acceptance.				
	9) Notarized certification that the				
	supplier/bidder shall provide free				
	installation of the equipment.				
	10) Notarized certification that the				
	·				
	brand has been in the local market				
	for at least two (2) years with at				
	least one (1) installation of the				
	equipment in the local market.				
	11) Notarized certification to provide				
	user's Manual in English language	- 0			
	A. Operations Manual	I			
	B. Service Manual				
	BIO REFRIGERATOR TWO-	1	unit	₱ 450,000.00	₱ 450,000.00
	DOOR	1		,	1 15 0,0 0 0 10 0
	SPECIFICATIONS:				
	- WITH UPS				
	- 1440 x 830 x 1950 mm				
	- 1364 liters	1	1		
	- 248 kg				
	+2 to +14 °C (at fridge part center,				
	AT -5 to 0°C)				
	+2 to +23 °C (at fridge part center,				1.
			1		
	AT 0 to 35°C)				
	AT 0 to 35°C)  Microprocessor with non-volatile				
	- Microprocessor with non-volatile				
	/				
	- Microprocessor with non-volatile memory (Safety Lock with keypad)				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed insulation (CFC-free)</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed insulation (CFC-free)</li> <li>2 doors</li> <li>double pane glass</li> <li>self-closing</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed insulation (CFC-free)</li> <li>2 doors</li> <li>double pane glass</li> <li>self-closing</li> <li>8 wire shelves (polyethylene-</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed insulation (CFC-free)</li> <li>2 doors</li> <li>double pane glass</li> <li>self-closing</li> <li>8 wire shelves (polyethylene-coated, 530 x 604 mm, adjustable)</li> </ul>				
	<ul> <li>Microprocessor with non-volatile memory (Safety Lock with keypad)</li> <li>LED Thermistor</li> <li>Forced cool air circulation</li> <li>Forced type (cycle defrost)</li> <li>Fully automatic</li> <li>HFC and HFO combined (CFC-Free)</li> <li>Rigid polyurethane foamed insulation (CFC-free)</li> <li>2 doors</li> <li>double pane glass</li> <li>self-closing</li> <li>8 wire shelves (polyethylene-</li> </ul>				

<i>p</i>	
	- Delivery and installation to
	Pharmacy Pavilion 2.
	STANDARD REQUIREMENTS
	1) Notarized/CTC of current and valid
	certificate of manufacturer's
	compliance with ISO 13485 and/or
	CE certificate or its equivalent.
	2) Notarized certification that the unit
	being offered shall be brand new.
	3) Notarized/CTC of certificate of
	exclusive or authorized
	distributorship from the
	manufacturer.
	4) Notarized certification that the
	supplier/bidder shall provide
	applications training for users and
	maintenance personnel of the
	hospital.
	5) Notarized certification that the
	supplier/bidder shall be responsible
	for the transportation to site,
	delivery, installation, and testing
	on the site and expenses for such
	will on the account of the supplier.
	6) Notarized certification that the
	supplier/bidder has the capability
	for the corrective and preventive
	maintenance of the unit, to include
	service engineers and technicians.
	7) Notarized/CTC of bidder's
	certification that the parts shall be
	available at the authorized
	Philippine service centers for a
	period of five (5) years after the
	warranty period.
	8) Notarized warranty certificate: one
	(1) year for parts and two (2) years
	for service from date of delivery,
	inspection and acceptance.
	9) Notarized certification that the
Ï	supplier/bidder shall provide free
	installation of the equipment.
	10) Notarized certification that the
	brand has been in the local market
	for at least two (2) years with at
	least one (1) installation of the
	equipment in the local market.
	11) Notarized certification to provide user's Manual in English language
	A. Operations Manual
	AA V PARMAVIM ITAMIAMI

B. Service Manual		
	GRAND TOTAL:	P 750,000,00

Submitted by:

CLARITA LACERNA, RPh

Chief, Pharmacy Section

Recommending App

ALDEN C. CUVOS, MD, FPPA, IFAPA, MMHoA Chairperson, BAC for Equipment

Approved By:

-SGD-

NOEL V. REYES, MD, FPPA, MMHoA

Medical Center Chief II