



REPUBLIC OF THE PHILIPPINES  
 Department of Health  
**NATIONAL CENTER FOR MENTAL HEALTH**  
 Nueve de Febrero Street, Mandaluyong City, Philippines



**BIDS AND AWARDS COMMITTEE**

Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

**SECTION VII**

**Technical Specifications**

**PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW BIO REFRIGERATOR FOR THE USE OF PHARMACY SECTION CY 2024 (LOT)  
 IB No. E-012-2024-PB**

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

<b>Item</b>	<b>Specification</b>	<b>Statement of Compliance</b>	<b>Supporting Documents for the Statement of Compliance (if applicable)</b>
1	<b>BIO REFRIGERATOR SINGLE DOOR</b>		
	<b>SPECIFICATIONS:</b>		
	WITH UPS		
	509 LITERS		
	TEMP RANGE 2-8 C		
	HGT 78 INCH		
	WIDTH 27.3INCH		
	DEPTH 32 INCH		
	WEIGHT 215KG		
	RFID TECHNOLOGY		
	GLASS DOOR		

2	<b>BIO REFRIGERATOR TWO-DOOR</b>		
	<b>SPECIFICATIONS:</b>		
	WITH UPS		
	1440 x 830 x 1950 mm		
	1364 liters		
	248 kg		
	+2 to +14 °C (at fridge part center, AT -5 to 0°C)		
	+2 to +23 °C (at fridge part center, AT 0 to 35°C)		
	Microprocessor with non-volatile memory (Safety Lock with keypad)		
	LED Thermistor		
	Forced cool air circulation		
	Forced type (cycle defrost)		
	Fully automatic		
	HFC and HFO combined (CFC-Free)		
	Rigid polyurethane foamed insulation (CFC-free)		
	2 doors		
	double pane glass		
	Self-closing		
	8 wire shelves (polyethylene-coated, 530 x 604 mm, adjustable)		
	max load 50 kg		
<b>CONFORME:</b>			
<i>(Company Name)</i>			
<i>(Name and Signature of Authorized Representative)</i>			