



LIST OF ITEMS

**Public Bidding for the Supply, Delivery, Installation, Testing and Commissioning of Various Brand-New Equipment for Anatomic Laboratory Section (Additional) CY 2024 (Line Item)
IB No. E-019-2024-PB**

ITEM CODE	ITEM DESCRIPTION	QTY	UOM	UNIT PRICE	TOTAL PRICE
ALA-01-2024	<p>Supply, Delivery, Installation, Testing and Commissioning of BENCHTOP CLINICAL CENTRIFUGE</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> Twelve (12) Placer Centrifuge with stainless steel centrifuging chamber. Housing: Metal, with metal lid, with viewing port in the lid. Lid locking and holding during rotor run, with one hand lid lock. Accommodate 7ml, 10ml, and 15ml tubes without changing rotor ABS outer shell can resist collision. Speed and time of table centrifuge are adjustable With safety switch, when cover is opened the motor can be stopped or emergency lid- lock Speed linear and stable Auto-balancing system With induction motor Brushless motor, free maintenance With auditory message after conduction of the centrifugation Dimension: at least (H*W*D) 250x 350x430mm) Speed range (RPM) 6,000 rpm With Rotor adjustable Power supply; 220V- 240V, 50-60Hz <p>Accessories:</p> <ol style="list-style-type: none"> Automatic Voltage Regulator (AVR) compatible with the equipment. <p>Standard Requirements:</p> <ol style="list-style-type: none"> Notarized/CTC current and valid certificate of manufacturer's compliance with ISO 13485: 2016 and/or CE certificate or its equivalent. Notarized certification that the unit being offered shall be brand new. Notarized/CTC certificate of authorized or exclusive distributorship from the manufacturer. Notarized certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital. Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder. 	1	unit	₱ 375,200.00	₱ 375,200.00



	<ol style="list-style-type: none"> 6. Notarized certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit, to include service engineers and technicians. 7. CTC certification to provide users' Manual in English language. <ol style="list-style-type: none"> a) Operations Manual - 2 copies b) Service Manual - 2 copies 8. Notarized bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 9. Notarized certification that the supplier/ bidder has at least one (1) installation of the Same equipment in the local market with attached list of installation/s (hospitals/institutions). 10. Notarized warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance. 11. Notarized certification that the brand has been in the local market for at least two (2) years. 12. Notarized certification that the bidder/supplier has available trained engineer/technician direct from the manufacturer for corrective and preventive maintenance. 13. Delivery period - 60 calendar days 14. To be Installed at NCMH New Laboratory Building (Histopathology Unit) 				
ALA-02-2024	<p>Supply, Delivery, Installation, Testing and Commissioning of BIOLOGICAL MICROSCOPE</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> 1. Infinity optical system 2. Right-handle low position mechanical stage with single specimen holder 3. Low positioned focusing knob 4. Built-in LED Illumination 5. Fixed Quadruple revolving nosepiece with inward tilt 6. Focusing: Coarse /Fine Adjustment 7. Observation Method: Brightfield, darkfield 8. Plan achromat, anti-fungal; Objectives; 4X,10X,40,100X 9. Abbe condenser with built-in aperture iris diaphragm 10. Trinocular tube (anti-fungal) 11. Eyepiece pair with 10x magnification, Field Number 20mm.and anti-fungal coating. 12. With built -in power cord holder at the back of microscope 13. Power supply Input: AC 100-240V,50/60Hz 0.4 A <p>Accessories:</p> <ol style="list-style-type: none"> a) Microscope Dust Cover b) Power cord c) One (1) Automatic Voltage Regulator (AVR) compatible with the equipment 	2	unit	₱ 181,440.00	₱ 362,880.00



	<p>Standard Requirements:</p> <ol style="list-style-type: none"> 1. Notarized/CTC current and valid certificate of manufacturer's compliance with ISO 13485: 2016 and/or CE certificate or its equivalent. 2. Notarized certification that the unit being offered shall be brand new. 3. Notarized/CTC certificate of authorized or exclusive distributorship from the manufacturer. 4. Notarized certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital. 5. Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder. 6. Notarized certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit, to include service engineers and technicians. 7. CTC certification to provide users' Manual in English language. <ol style="list-style-type: none"> a) Operations Manual - 2 copies b) Service Manual - 2 copies 8. Notarized bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 9. Notarized certification that the supplier/ bidder has at least one (1) installation of the Same equipment in the local market with attached list of installation/s (hospitals/institutions). 10. Notarized warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance. 11. Notarized certification that the brand has been in the local market for at least ten (10) years. 12. Notarized certification that the bidder/supplier has available trained engineer/technician direct from the manufacturer for corrective and preventive maintenance. 13. Certification of good standing record from six (6) hospital institutions 14. Delivery period - 60 calendar days 15. To be Installed at NCMH New Laboratory Building (Histopathology Unit) 				
<p>ALA-03-2024</p>	<p>Supply, Delivery, Installation, Testing and Commissioning of CADAVER REFRIGERATOR</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> 1. Two (2) Body cadaver compartment, with two (2) doors 2. End-opening configuration 3. Doors equipped with perimeter heaters 4. Built-in Electronic temperature control with digital temperature display and high/low audio visual alarm 	<p>1</p>	<p>unit</p>	<p>₱ 3,080,000.00</p>	<p>₱ 3,080,000.00</p>



<p>5. Operating Temperature range:2-4 degrees Centi-grade/ 35-30 degrees Fahrenheit</p> <p>6. Exterior finishing: stainless steel and/or aluminum</p> <p>7. Interior chamber finishing: Stainless steel and/or aluminum</p> <p>8. With two (2) removable stainless steel body trays</p> <p>9. With three (3)-piece telescoping rack system for each body compartment that can be extended and capable of supporting the body tray and cadaver at full extension</p> <p>10. Vapor-proof light with switch</p> <p>11. With four (4) heavy-duty casters, two of which are swivel type with wheel brakes</p> <p>12. Power Supply Input:220V,50/60Hz</p> <p>Accessories:</p> <p>1. One (1) Automatic Voltage Regulator (AVR) compatible with the equipment</p> <p>Standard Requirements:</p> <p>1. Notarized/CTC current and valid certificate of manufacturer's compliance with ISO 13485: 2016 and/or CE certificate or its equivalent.</p> <p>2. Notarized certification that the unit being offered shall be brand new.</p> <p>3. Notarized/CTC certificate of authorized or exclusive distributorship from the manufacturer.</p> <p>4. Notarized certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital.</p> <p>5. Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder.</p> <p>6. Notarized certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit, to include service engineers and technicians.</p> <p>7. CTC certification to provide users' Manual in English language.</p> <p>a) Operations Manual - 2 copies</p> <p>b) Service Manual - 2 copies</p> <p>8. Notarized bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period.</p> <p>9. Notarized certification that the supplier/ bidder has at least one (1) installation of the Same equipment in the local market with attached list of installation/s (hospitals/institutions).</p> <p>10. Notarized warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance.</p> <p>11. Notarized certification that the brand has been in the local market for at least two (2) years.</p> <p>12. Notarized certification that the bidder/supplier has available trained engineer/technician direct from the manufacturer for corrective and preventive maintenance.</p>				
---	--	--	--	--



	13. Delivery period - 60 calendar days 14. To be Installed in NCMH Morgue Building				
ALA-04-2024	<p>Supply, Delivery, Installation, Testing and Commissioning of PARAFFIN BLOCK METAL CABINET</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> Three (3) Layers, 18 Drawers ,6 grids, plus a base Loading capacity at least 12,000 pcs paraffin blocks Material: Metal steel plate Cabinet With specialized slide drawer, ABS slide Drawer Slot made by special mold With attached pocket of label per drawer Dimension: with at least (W*D*H*) 450*475*1295mm <p>Standard Requirements:</p> <ol style="list-style-type: none"> Notarized certification that the unit being offered shall be brand new. Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder. Notarized warranty certificate: one (1) year from date of delivery, inspection and acceptance. Delivery period - 60 calendar days To be Installed in NCMH Morgue Building 	1	unit	₱ 224,000.00	₱ 224,000.00
ALA-05-2024	<p>Supply, Delivery, Installation, Testing and Commissioning of SLIDES METAL CABINET</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> Twelve (12) Layers, 72 Drawers ,6 grids, plus a base Loading capacity at least 50,000 pcs microscope glass slides Material: Metal steel plate Cabinet With specialized slide drawer, ABS slide Drawer With Upward/vertical position of slides to be stored. With removable divider serve as stopper to hold slides per drawer. Dimension: with at least (W*D*H*) 403*478*1625mm <p>Standard Requirements:</p> <ol style="list-style-type: none"> Notarized certification that the unit being offered shall be brand new. Notarized certification that the supplier/bidder shall be responsible for the transportation to site, delivery, installation and testing on the site (hospital) and expenses for such will be on the account of the supplier/bidder. Notarized warranty certificate: one (1) year from date of delivery, inspection and acceptance. Delivery period - 60 calendar days 	1	unit	₱ 224,000.00	₱ 224,000.00



	5. To be Installed in NCMH Morgue Building				
ALA-06-2024	<p>Supply, Delivery, Installation, Testing and Commissioning of VENTILATED PATHOLOGY GROSS STATION</p> <p>SPECIFICATIONS:</p> <ol style="list-style-type: none"> 1. Gross Pathology workstation, floor-type 2. Ventilated hood made of stainless steel containing set of replaceable filters consisting of primary pre-filter carbon filter(formalin) and HEPA filter. 3. Equipment Dimensions: Width:1200mm, Depth:750 to 850 mm, Height: 2400 to 2500mm 4. Stainless steel workspaces with sink, faucet, spray hose and stainless steel shelf 5. LED lighting system directed to the work surface 6. Sliding/movable shield made of tempered glass, motorized 7. Control panel with LCD display 8. Magnetic instrument holder 9. Polypropylene dissection board 10. Automatic formalin loading tank 11. Foot pedal control for formalin dispensing 12. Waste formalin collection tank with level sensor 13. Foot pedal control for water dispensing 14. Trolley for formalin tanks 15. With external electrical sockets 16. With pulverizer 17. Power Supply Input:220-240 V, 50/60Hz <p>Accessories:</p> <ol style="list-style-type: none"> 1. One (1) Voltage Regulator compatible with the equipment 2. Three (3) sets of filter (each set contains prefilter, activated carbon filter for formalin, and HEPA filter) 3. One (1) Digital top Loading Balance 4. One (1) Extra formalin loading tank 5. One (1) Extra waste formalin collection tank 6. One (1) Set of tissue marking dyes 7. One (1) set of stainless steel surgical instruments including the following (two pieces for each instrument): a. Mayo scissors, b. Metzenbaum scissors, c. Tissue forceps, non-toothed, d. Tissue forceps, toothed, e. hemostat <p>Standard Requirements:</p> <ol style="list-style-type: none"> 1. Notarized/CTC current and valid certificate of manufacturer's compliance with ISO 13485: 2016 and/or CE certificate or its equivalent. 2. Notarized certification that the unit being offered shall be brand new. 3. Notarized/CTC certificate of authorized or exclusive distributorship from the manufacturer. 4. Notarized certification that the supplier/bidder shall provide applications training for users and maintenance personnel of the hospital. 	1	unit	₱ 3,024,000.00	₱ 3,024,000.00



	<p>and expenses for such will be on the account of the supplier/bidder.</p> <ol style="list-style-type: none"> 6. Notarized certification that the supplier/ bidder has the capability for the corrective and preventive maintenance of the unit, to include service engineers and technicians. 7. CTC certification to provide users' Manual in English language. <ol style="list-style-type: none"> a) Operations Manual - 2 copies b) Service Manual - 2 copies 8. Notarized bidder's certification that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 9. Notarized certification that the supplier/ bidder has at least one (1) installation of a Histopathology equipment in the local market with attached list of installation/s (hospitals/institutions). 10. Notarized warranty certificate: one (1) year for parts and two (2) years for service from date of delivery, inspection and acceptance. 11. Notarized certification that the brand has been in the local market for at least two (2) years. 12. Notarized certification that the bidder/supplier has available trained engineer/technician direct from the manufacturer for corrective and preventive maintenance. 13. Delivery period - 60 calendar days 14. To be Installed at NCMH New Laboratory Building (Histopathology Unit) 				
GRAND TOTAL					₱ 7,290,080.00

Submitted by:

ROMIL M. AGUIRRE, MD, FPSP, MMHoA
End-user, Anatomic Laboratory Section

Recommending Approval:

ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
Chairperson, BAC for Equipment

Approved by:

-SGD-

NOEL V. REYES, MD, FPPA, MMHoA
Medical Center Chief II