



REPUBLIC OF THE PHILIPPINES  
 Department of Health  
**NATIONAL CENTER FOR MENTAL HEALTH**  
 Nueve de Febrero Street, Mandaluyong City, Philippines  
**BIDS AND AWARDS COMMITTEE**



Telephone No. 531-9001 loc. 239

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

**SECTION VII**

**Technical Specifications**

**PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW EQUIPMENT FOR PHYSICAL REHABILITATION UNIT CY 2024 (LINE ITEM)**

IB No. E-005-2024-PB

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

Item	Specification	Statement of Compliance	Supporting Documents for the Statement of Compliance (if applicable)
1	<b>Supply, Delivery, Installation, Testing and Commissioning of a Brand-New Shockwave Therapy Machine</b>		
	<b>SPECIFICATIONS:</b>		
	1. Radial shockwave		
	2. Electromagnetic generator w/o air compressor		
	3. With Intermittent and Progressive modes		
	4. 10-200mJ		
	5. 1-25Hz		
	6. Min 6 Burst modes		
	7. Min. 7 preset programs		
	8. Min. 25 preset treatment protocols		

	9. Touch screen console		
	10. Min 3 applicators with 6mm, 15mm, and 25mm diameter		
	11. Not more than 3kgs weight		
	12. Min dimension 250 x 220 x 100mm (LxWxH)		
	13. At least 2 million shots per applicator		
	14. With carrying case		
2	<b>Supply, Delivery, Installation, Testing and Commissioning of a Brand-New High-Intensity Laser Machine</b>		
	1. 100-240VAC		
	2. Class IV Laser		
	3. 15W continuous mode		
	4. 20W Pulsed mode		
	5. 660-970nm wavelength		
	6. High-definition color LCD touch screen monitor		
	7. With variable treatment preset protocol		
	8. Variable zoom distal 1-5cm diameter		
	9. Includes operator, adult, and child goggles		
	10. Lithium-ion battery		
	11. Two years warranty		
	12. Semi-annual PMS		
<b>CONFORME:</b>			
<i>(Company Name)</i>			
<i>(Name and Signature of Authorized Representative)</i>			