



REPUBLIC OF THE PHILIPPINES
Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
Nueve de Febrero Street, Mandaluyong City, Philippines
BIDS AND AWARDS COMMITTEE



Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bac@ncmh.gov.ph

Website: www.ncmh.gov.ph

SECTION VII TECHNICAL SPECIFICATIONS

ITB No. 019-2023

INSTRUCTION: Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
USS23-01	ULTRASONIC SCALER	
	1. Stylish, compact all unit for periodontic, endodontic, scaling, MI applications and implant maintenance.	
	2. Convenient, large capacity bottles of 400ml of the same solutions with independent pumps	
	3. The engine assures stable output by automatically adjusting to the optimum frequency depending on the loading of the tip	
	4. All operating parameters are clearly displayed on the large LED display controlled by easily recognizable buttons.	
	5. Has auto-cleaning mode that automatically flushes the tubings, handpieces, and tips at the push of a button when required, The two channels can be flushed independently	
	6. The handpiece, tips and tip wrenches can be sterilized together for your convenience	
	7. Has sterilization	
	SPECIFICATION:	
	1. Power Supply: AC 230v 50/60 Hz	
	2. Frequency: 28-32 KHz	
	3. Frequency: 28-32 KHz	
	4. Dimension: W 160 x D 270 x H 190(mm), including bottle	
	5. Bottle Volume: 400ml (2 bottles)	

	6. Bottle Volume: 400ml (2 bottles)	
	CONTENTS:	
	1. Control Unit	
	2. Optic Handpiece	
	3. 3 scaling tips (G4, G6, G8)	
	4. Foot control (FC-70)	
	5. 2 x irrigation bottle	
	6. 3 x hip wrench with torque limiter	
	7. AC cord	
	8. Sterilization case	
	WARRANTY:	
	1. At least 1 year on parts	
	2. 3 years free services	
	3. Company can provide the availability of parts for the next 10 years	
	4. Company has been in the dental business for the last 15 years.	

Conformed by:

**Authorized Representative's
Signature over printed name**

Date: _____