



REPUBLIC OF THE PHILIPPINES  
Department of Health  
**NATIONAL CENTER FOR MENTAL HEALTH**  
Nueve de Febrero Street, Mandaluyong City, Philippines  
**BIDS AND AWARDS COMMITTEE**



Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bac@ncmh.gov.ph

Website: www.ncmh.gov.ph

**SECTION VII**  
**Technical Specifications**

**ITB No. 023-2022**

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
1.	Alcohol, Isopropyl 70% (at least 3.785L/gallon) . (with at least 24 months expiration).	
2.	Aneroid Sphygmomanometer with Stethoscope, ADULT - Latex Free with single tube. Range: 0-300mmHg. Precision valve with slow deflation rate. Extra large inflation bulb. Gauge with internal diameter of 65mm. Velcro cuff, for adults, washable made in nylon with ring. With Certificate of Calibration (PORTABLE)	
3.	Blood Glucose Test Strips with Strip Ejector, for in vitro diagnostic use, no coding, individually foil wrapped strips, small blood sample required at least (0.5µl), 5 second test time, 5 second blood sample reapplication time, top or end fill with visual confirmation, fill trigger ensures test starts only when sufficient blood sample has been applied. Must be compatible to existing Glucometer. (with at least 24 months expiration).	
4.	Blood Transfusion Set, sterilized by EO gas, non-toxic, no heavy metal & pyrogen free w/ vein needle G.18, double chamber w/ large mesh filter minimizing platelet loss & hemolysis, flexible table, individually packed into plastic pouch. (with at least 24 months expiration).	
5.	Boufant Cap, Single Elastic, Strip type, disposable	
6.	Disposable ECG Electrodes, ADULT, 36x45mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
7.	Foley Catheter 2-way, Fr.16, latex pyrogen-free, disposable, sterilized by EO gas, needle free (with at least 24 months expiration).	
8.	Gloves, Nitrile examination, disposable, powder free, non-sterile, ambidextrous, single use, size SMALL (with at least 24 months expiration)	
9.	Gloves, Nitrile examination, disposable, powder free, non-sterile, ambidextrous, single use, size MEDIUM (with at least 24 months expiration)	
10.	Gloves, Nitrile examination, disposable, powder free, non-sterile, ambidextrous, single use, size LARGE (with at least 24 months expiration)	
11.	Gloves, 7.0 size, sterile, powder free, thick and elastic, not easily tore. (with at least 24 months expiration).	
12.	KN95, 5 layer Respirator Mask, single, individually packed, sterile, earloop type, with high filtration efficiency of > 95%.	
13.	Mask, surgical disposable, 3 PLY, Elastic without glass fibers, hypoallergenic, very low resistance to breathing, nose bar adaptable, high filtration capacity, perfect fitting, EAR-LOOP	
14.	MEDICAL OXYGEN REFILL Standard size, 5.66 cu.m, Pressure gauge of 1,880 psi at least 50lbs. (No delivery charge and service for the use of equipment and gas containers, and for other services.	
15.	Sharp Disposal Container, double lid, with hermetic lock seal, 30 Liters	
16.	Skin Stapler wire, taller and sharper, pre-cock function, 35 staple count, sterile with free remover (with at least 24 months expiration).	
17.	Surgical Gown (FREE SIZE), disposable, non-woven fabric, anti-water, anti-blood, anti-alcohol, anti-static, light color only.	
18.	Surgical Shoe Cover, disposable, non-woven fabric, anti-water, anti-blood, anti-alcohol, anti-static.	
19.	Syringes, disposable, 1cc, with needle G.26 x1/2 inch, slip tip, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
	atleast 24 months expiration).	
20.	THERMAL PAPER, 3 CHANNELS 80mm x 20mm	
21.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 6 (9cm x 15cm) P: 1 3/4 x 4 (4.5cm x 10cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
22.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 8 (10cm x 20cm) P: 1 3/4 x 6 (4.5cm x 15cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)	
23.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 10 (10cm x 25cm) P: 1 3/4 x 8 (4.5cm x 20cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)	
24.	Underpads Super Dry, 60x90cm, disposable. (with at least 24 months expiration).	
25.	Suction Poole Drain, non-toxic, non-pyrogenic, sterilized by EO gas, with 210cm universal connecting tube (with at least 24 months expiration).	
26.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 18 to 24 months expiration).	
27.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
28.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).	
29.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm / 40mm (with at least 24 months expiration).	
30.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with at least 24 months expiration).	
31.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm / 40mm (with at least 24 months expiration).	
32.	Scalpel Blade #10 with plastic handle, single use (with at least 24 months expiration).	
33.	Scalpel Blade #20 with plastic handle, single use (with at least 24 months expiration).	
34.	Silk braided, non-absorbable, size 0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).	
35.	Silk braided, non-absorbable, size 1, 60cm-75cm, 12-15 threads, no needles (with at least 24 months expiration).	
36.	Silk braided, non-absorbable, size 2/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).	
37.	Silk braided, non-absorbable, size 3/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).	
38.	Silk braided, non-absorbable, size 4/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).	
39.	Silk braided, non-absorbable, size 3/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with at least 24 months expiration).	
40.	Silk braided, non-absorbable, size 2/0, 75cm, round needle, 1/2 circle, 26mm / 37mm (with at least 24 months expiration).	
41.	Silk braided, non-absorbable, size 3/0, 75cm, round needle, 1/2 circle, 26mm / 37mm (with at least 24 months expiration).	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
42.	Wound Drainage System - Silicone 100-150mL reservoir capacity, with clear calibration and built-in clip hanger, NRV (Non return valve) system, 720mm Silicone connecting tube with radio opaque line, 110mmHg Negative pressure, with or without Trocar needle. (with at least 24 months expiration).	
43.	Diaper, disposable, Large, Adult (for the use of Pavilion 8)	
44.	Arm Sling - adjustable shoulder strap allows for a custome comfortable fit. Supports left or right arm. 70% polyester, latex free.	
45.	Abdominal Binder	
46.	Adhesive Bandage, flexible fabric and strong adhesive, all one size, 1x3" (100pcs/box)	
47.	Colostomy bag with wafer 57mm	
48.	Colostomy bag with wafer 70mm	
49.	Duoderm Patch Extra Thin dressing 8x8 inches / 20x20cm	
50.	Endotracheal tube 2.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, radio-opaque (with at least 24 months expiration)	
51.	Endotracheal tube 2.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, radio-opaque (with at least 24 months expiration)	
52.	Endotracheal tube 3.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, radio-opaque (with at least 24 months expiration)	
53.	Endotracheal tube 4.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, radio-opaque (with at least 24 months expiration)	

**INSTRUCTION:** Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
54.	Endotracheal tube 6.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, radio-opaque (with at least 24 months expiration)	
55.	Foley Catheter, Fr.5, pyrogen free, disposable, sterilized by EO gas (with at least 24 months expiration)	
56.	Isopropanol, Benzalkonium Chloride, 1,000ml - Skin, wound and operative site disinfectant. Skin antiseptics before injections, punctures, etc (with at least 24 months expiration)	
57.	Nephrostomy Catheter kit, with safety string lock, Fr.14	
58.	Orthoplast Skin Traction with adhesive plaster, good quality and single packed (with at least 24 months expiration)	
59.	Straight Catheter (Nelaton Catheter) transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic, Fr.5 (with at least 24 months expiration)	
60.	Zinc Oxide Adhesive Plaster, 5 inches, with spool and snap ring, 7.5cm x 5m, 100% adhesive (with at least 24 months expiration)	
61.	Pre-heparinized arterial blood gas 3cc syringe	
62.	Adult full body convective warming blanket	
63.	Pediatric Full body convective warming blanket	
64.	Anesthesia Set Flex, latex free, adult, breathing circuit with fitting mask - disposable anesthesia set consisting of 2 hose, connector, fastening ring 2 latex free breathing bag	
65.	Anesthesia Set Flex, latex free - pedia, breathing circuit with fitting mask - disposable pediatric anesthesia set, consisting of 2 hoses (diameter 15mm), y-hose (diameter 15mm), connector, fastening ring, 1L latex free, breathing bag	
66.	Anesthesia Breathing circuit, adult (with at least 24 months expiration)	
67.	Anesthesia breathing circuit, pedia (with at least 24 months expiration)	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
68.	Anesthesia tubing - jackson-reese (with at least 24 months expiration)	
69.	Antibiotic Mesh, 10cm x 10cm (with at least 24 months expiration)	
70.	Antiseptic surgical scrub solution, non-ionic surfactants. Has a pleasant perfume, and tartrazine as colours. Easy to wash off. Can be used for neonates and thyroid patients. Not degraded due to bioburden. Effective in presence of blood, seriuos fluids, sweat etc. Non-staining and skin friendly. (5 liters / gallon) (with at least 12 months expiration)	
71.	Chest tube catheter / Thoracic Catheter, Fr.20, thermosensitive tube, minimizing trauma without tension and compression, clear and transparent catheter, smooth catheter eye transition, sterile (with at least 24 months expiration)	
72.	Chest tube thoracostomy bottle with glass tube and rubber cock (CTT Bottle) with 3ft rubber hose tubing	
73.	Compact Extendable (.5 inches), 59 inches breathing system with luer elbow pediatric circuit (with at least 24 months expiration)	
74.	Compact Extendable (.8 inches), 6ft, breathing system, 2L bag and 4ft extra arm, 35's/box (with at least 24 months expiration)	
75.	Epidural Set, Epidural Catheter, G.18, for continuous epidural anesthesia with Tuohy tape needle, catheter (closed tip, 3 lateral eyes), with connector, epidural flat filter and LOR syringe, filter 0.2 micron, Loss-of-Resistance syringe - Luer Slip (with at least 24 months expiration)	
76.	Extension tube with male luer lock adapter, with roller clamp (with at least 24 months expiration)	
77.	Liquid Handrub antiseptic, alcoholic skin disinfectant, bactericidal, fungicidal and virucidal, skin emollients, 1 liter / bottle (with at least 24 months expiration)	
78.	Silk braided, non-absorbable, size 4.0, 75cm cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with at least 24 months expiration)	
79.	Sterile Water, 100ml (with at least 24 months expiration)	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM	SPECIFICATION	STATEMENT OF COMPLIANCE
80.	Surgical Scrub with chlorhexidine and brushes (with at least 24 months expiration)	
81.	Extension tube with tri-lumen NF (needle free) with luer access valve, closed system, color coded slide clamp. Kink resistant tubing. (with at least 24 months expiration)	
82.	T-Tube, single use, individually packed, smooth surface, non-toxic latex, with radiopaque line, Fr.12 (with at least 24 months expiration)	
83.	Chlorine Test Strip, with shelf life of 18 months, with manufacturing year of 2020	
84.	Disinfecting Agent for hemodialysis based on hydrogen peroxide and peroxyatic acid mixture, stabilized 5.1 and 8 II, 5 kilogram, cold sterilant, at least 5 liters per container, with at least 24 months expiration	
85.	Subclavian kit inclusive of 2 pairs of Sterile Gloves 6.5, 2 packs of sterile gauze sponges 4x4x8 5's, 1 pc sterile drape, 8 pcs Chlorhexidine pad, 2pcs heplock, with at least 24 months expiration	
86.	Test for residual peroxide in rinse water, 100 strips per bottle, with shelf life of 18 months	
87.	Thermochemical Disinfectant for heat disinfection of hemodialysis machines, each 100g of disinfectant contains 21g citric acid 1-hydrate, 5 liters per container, with at least 24 months expiration	

**CONFORMED BY:** \_\_\_\_\_

Authorized Representative  
(Signature over printed name)

**DATE:** \_\_\_\_\_





# PhilGEPS

Philippine Government Electronic Procurement System

Central Portal for  
Philippine Government  
Procurement Opportunities

## Bid Notice Abstract

### Invitation to Bid (ITB)

**Reference Number** 8984971  
**Procuring Entity** NATIONAL CENTER FOR MENTAL HEALTH  
**Title** SUPPLY AND DELIVERY OF MEDICAL SUPPLIES (ADDITIONAL 2) CY 2022  
**Area of Delivery** Metro Manila

<b>Solicitation Number:</b>	NCMH-PB-MEDICAL SUPPLIES(MS2) CY 2022	<b>Status</b>	<b>Active</b>
<b>Trade Agreement:</b>	Implementing Rules and Regulations	<b>Associated Components</b>	5
<b>Procurement Mode:</b>	Public Bidding	<b>Bid Supplements</b>	0
<b>Classification:</b>	Goods	<b>Document Request List</b>	1
<b>Category:</b>	Medical Supplies and Laboratory Instrument	<b>Date Published</b>	06/09/2022
<b>Approved Budget for the Contract:</b>	PHP 16,510,256.39	<b>Last Updated / Time</b>	06/09/2022 00:00 AM
<b>Delivery Period:</b>	7 Day/s	<b>Closing Date / Time</b>	26/09/2022 17:00 PM
<b>Client Agency:</b>			
<b>Contact Person:</b>	elmer ilanan sale admin. aide III Nueve de Pebrero St., Mandaluyong City Metro Manila Philippines 1550 63-2-5318318 63-2-5318318 elsale73@yahoo.com		

#### Description

The National Center for Mental Health (NCMH) is inviting all Phil-GEPS registered to apply the Invitation to Bid for PUBLIC BIDDING FOR SUPPLY AND DELIVERY OF MEDICAL SUPPLIES (ADDITIONAL 2) CY 2022. (Complete Specifications are indicated in the List of items.

Complete set of Bid Documents shall be provided to all interested bidders on September 6 - September 26, 2022 (Monday to Friday 8:00AM to 5:00PM) at the National Center for Mental Health (NCMH), Procurement Section, Mandaluyong City upon payment of non-refundable fee in the form of cash and upon submission of LETTER OF INTENT.

The National Center for Mental Health assumes no responsibility whatsoever to compensate or indemnify bidders for any expenses incurred in the preparation of their bid.

FOR INQUIRIES, PLEASE VISIT THE BAC SECRETARIAT OFFICE – PROCUREMENT SECTION, National Center for Mental Health, Nueve De Pebrero St., Mandaluyong city.

Contact Persons:  
 JERRY C. RODRIGUEZ, MGM-ESP  
 Contact No. 531-9001(Local-242, 240,239; Telefax:531-8318

#### Line Items

Item No.	Product/Service Name	Description	Quantity	UOM	Budget (PHP)
1	Alcohol, Isopropyl	Please see attached complete specification at the list of items	2,710	Gallon	813,000.00
2	Aneroid Sphygmomanometer with	Please see attached complete specification at the list of	110	Piece	143,000.00

	Stethoscope, ADULT	items			
3	Blood Glucose Test Strips with Strip Ejector	Please see attached complete specification at the list of items	2,700	Piece	29,700.00
4	Blood Transfusion Set	Please see attached complete specification at the list of items	260	Piece	13,000.00
5	Bouffant Cap	Please see attached complete specification at the list of items	17,360	Piece	38,192.00
6	Disposable ECG Electrodes, ADULT,	Please see attached complete specification at the list of items	950	Piece	34,200.00
7	Foley Catheter 2-way, Fr.16, latex pyrogen free	Please see attached complete specification at the list of items	270	Piece	8,640.00
8	Gloves, Nitrile examination, disposable, Small	Please see attached complete specification at the list of items	116,000	Piece	580,000.00
9	Gloves, Nitrile examination, disposable, Medium	Please see attached complete specification at the list of items	159,000	Piece	795,000.00
10	Gloves, Nitrile examination, disposable, Large	Please see attached complete specification at the list of items	155,500	Piece	777,500.00
11	Gloves, 7.0 size, sterile	Please see attached complete specification at the list of items	2,700	Piece	52,650.00
12	KN95, 5 layer Respirator Mask	Please see attached complete specification at the list of items	39,300	Piece	2,043,600.00
13	Mask, surgical disposable	Please see attached complete specification at the list of items	231,900	Piece	2,087,100.00
14	MEDICAL OXYGEN REFILL Standard size	Please see attached complete specification at the list of items	1,615	Piece	726,750.00
15	Sharp Disposal Container	Please see attached complete specification at the list of items	30	Piece	27,000.00
16	Skin Stapler wire	Please see attached complete specification at the list of items	35	Piece	11,235.00
17	Surgical Gown (FREE SIZE)	Please see attached complete specification at the list of items	49,750	Piece	3,482,500.00
18	Surgical Shoe Cover	Please see attached complete specification at the list of items	57,465	Piece	229,860.00
19	Syringes, disposable, 1cc,	Please see attached complete specification at the list of items	3,000	Piece	10,500.00
20	THERMAL PAPER	Please see attached complete specification at the list of items	215	Piece	21,500.00
21	Transparent Dressing with absorbent pad	Please see attached complete specification at the list of items	35	Piece	7,000.00
22	Transparent Dressing with absorbent pad	Please see attached complete specification at the list of items	30	Piece	7,410.00
23	Transparent Dressing with absorbent pad	Please see attached complete specification at the list of items	60	Piece	17,580.00
24	Underpads Super Dry,	Please see attached complete	2,500	Piece	72,500.00



	60x90cm	specification at the list of items			
25	Suction Poole Drain, non-toxic	Please see attached complete specification at the list of items	80	Piece	14,800.00
26	PGA (Polyglycolic Acid) braided	Please see attached complete specification at the list of items	36	Piece	16,200.00
27	PGA (Polyglycolic Acid) braided	Please see attached complete specification at the list of items	36	Piece	16,200.00
28	PGA (Polyglycolic Acid)	Please see attached complete specification at the list of items	36	Piece	16,200.00
29	PGA (Polyglycolic Acid) braided	Please see attached complete specification at the list of items	48	Piece	23,040.00
30	PGA (Polyglycolic Acid) braided	Please see attached complete specification at the list of items	48	Piece	21,600.00
31	PGA (Polyglycolic Acid) braided	Please see attached complete specification at the list of items	48	Piece	21,600.00
32	Scalpel Blade #10	Please see attached complete specification at the list of items	100	Piece	7,500.00
33	Scalpel Blade #20	Please see attached complete specification at the list of items	150	Piece	11,250.00
34	Silk braided, non asorbable, size 0, 60cm	Please see attached complete specification at the list of items	36	Piece	8,136.00
35	Silk braided, non asorbable, size 1, 60cm-75cm	Please see attached complete specification at the list of items	36	Piece	8,064.00
36	Silk braided, non asorbable, size 2/0, 60cm	Please see attached complete specification at the list of items	48	Piece	10,920.00
37	Silk braided, non asorbable, size 3/0	Please see attached complete specification at the list of items	48	Piece	10,560.00
38	Silk braided, non asorbable, size 4/0	Please see attached complete specification at the list of items	24	Piece	4,920.00
39	Silk braided, non asorbable, size 3/0,	Please see attached complete specification at the list of items	48	Piece	9,120.00
40	Silk braided, non asorbable, size 2/0	Please see attached complete specification at the list of items	48	Piece	11,040.00
41	Silk braided, non asorbable, size 3/0	Please see attached complete specification at the list of items	48	Piece	9,840.00
42	Wound Drainage System	Please see attached complete specification at the list of items	40	Piece	48,000.00
43	Diaper, disposable, Large, Adult	Please see attached complete specification at the list of items	111,600	Piece	2,790,000.00
44	Arm Sling	Please see attached complete specification at the list of items	50	Piece	3,600.00
45	Abdominal Binder	-	75	Piece	9,600.00
46	Adhesive Bandage	Please see attached complete specification at the list of	58	Box	20,880.00

		items			
47	Colostomy bag with wafer 57mm	-	200	Piece	96,000.00
48	Colostomy bag with wafer 70mm	-	386	Piece	258,620.00
49	Duoderm Patch Extra Thin dressing 8x8 inches / 20x	-	12	Piece	20,400.00
50	Endotracheal tube 2.0 high volume	Please see attached complete specification at the list of items	10	Piece	900.00
51	Endotracheal tube 2.5 high volume	Please see attached complete specification at the list of items	10	Piece	900.00
52	Endotracheal tube 3.0 high volume	Please see attached complete specification at the list of items	10	Piece	900.00
53	Endotracheal tube 4.5 high volume	Please see attached complete specification at the list of items	10	Piece	900.00
54	Endotracheal tube 6.0 high volume	Please see attached complete specification at the list of items	10	Piece	900.00
55	Foley Catheter, Fr.5, pyrogen free	Please see attached complete specification at the list of items	100	Piece	16,000.00
56	Isopropanol, Benzalkonium Chloride	Please see attached complete specification at the list of items	50	Bottle	130,000.00
57	Nephrostomy Catheter kit	Please see attached complete specification at the list of items	8	Piece	79,200.00
58	Orthoplast Skin Traction	Please see attached complete specification at the list of items	50	Pack	37,500.00
59	Straight Catheter	Please see attached complete specification at the list of items	41	Piece	1,517.00
60	Zinc Oxide Adhesive	Please see attached complete specification at the list of items	20	Roll	14,000.00
61	Pre-heparinized arterial blood gas 3cc syringe	-	960	Piece	96,000.00
62	Adult full body convective warming blanket	-	16	Piece	32,000.00
63	Pediatric Full body convective warming blanket	-	15	Piece	30,000.00
64	Anesthesia Set Flex	Please see attached complete specification at the list of items	11	Roll	15,400.00
65	Anesthesia Set Flex	Please see attached complete specification at the list of items	2	Roll	2,800.00
66	Anesthesia Breathing circuit, adul	Please see attached complete specification at the list of items	6	Piece	840.00
67	Anesthesia breathing circuit	Please see attached complete specification at the list of items	4	Piece	560.00
68	Anesthesia tubing	Please see attached complete specification at the list of items	2	Piece	280.00
69	Antibiotic Mesh	Please see attached complete specification at the list of items	12	Piece	3,540.00



70	Antiseptic surgical scrub solution	Please see attached complete specification at the list of items	8	Gallon	10,000.00
71	Chest tube catheter	Please see attached complete specification at the list of items	10	Piece	6,000.00
72	Chest tube thoracostomy bottle	Please see attached complete specification at the list of items	40	Piece	50,000.00
73	Compact Extendable	Please see attached complete specification at the list of items	4	Piece	3,240.00
74	Compact Extendable	Please see attached complete specification at the list of items	5	Piece	9,750.00
75	Epidural Set, Epidural Catheter	Please see attached complete specification at the list of items	30	Pack	20,100.00
76	Extension tube	Please see attached complete specification at the list of items	30	Piece	1,890.00
77	Liquid Handrub antiseptic,	Please see attached complete specification at the list of items	50	Bottle	72,500.00
78	Silk braided, non-absorbable	Please see attached complete specification at the list of items	48	Piece	13,920.00
79	Sterile Water, 100ml	Please see attached complete specification at the list of items	102	Piece	15,300.00
80	Surgical Scrub with chlorhexidine	Please see attached complete specification at the list of items	150	Piece	15,000.00
81	Extension tube with tri-lumen NF	Please see attached complete specification at the list of items	198	Piece	19,800.00
82	T-Tube, single use	Please see attached complete specification at the list of items	10	Piece	1,612.39
83	Chlorine Test Strip	Please see attached complete specification at the list of items	700	Piece	24,500.00
84	Disinfecting Agent	Please see attached complete specification at the list of items	9	Gallon	54,000.00
85	Subclavian kit	Please see attached complete specification at the list of items	600	Piece	99,000.00
86	Test for residual peroxide in rinse water	Please see attached complete specification at the list of items	1,000	Piece	12,000.00
87	Thermochemical Disinfectant	Please see attached complete specification at the list of items	20	Gallon	120,000.00

**Pre-bid Conference**

Date	Time	Venue
13/09/2022	9:00:00 AM	PAG-ASA HALL, NATIONAL CENTER FOR MENTAL HEALTH NUEVE DE PEBRERO STREET, BRGY., MAUWAY, MANDALUYONG CITY