



National Center for Mental Health

LIST OF ITEMS FOR PUBLIC BIDDING CY2022 SUPPLY, DELIVERY, COMMISSIONING AND TESTING OF VARIOUS DENTAL EQUIPMENT

NO	ITEM CODE	ITEM DESCRIPTION	QTY	UNIT OF MEASURE	UNIT PRICE	TOTAL PRICE
1	VDE22-01	EXTERNAL ORAL SUCTION MACHINE Specifications: 1. Voltage: AC110V – AC220V, 50HZ / 60HZ 2. Rated Power: 1000W – 1200W 3. Control Mode: touch screen / two gears adjustable 4. Suction flow: 3000 to 4000 liters/min 5. Filtration efficiency: (0.3 um) 99.9% 6. UV intensity (Distance = 1M): 1210uw/ cm2 7. UV light working life: 10000 hours 8. UV wave length: 254mm 9. Plasma sterilization working life: 6000 hours 10. Noise: 52 – 60dB 11. Net Weight: 20 – 25kg 12. Pressure: 25kPa 13. Warranty: 1 year (except filter and plasma lifetime service) Standard Requirements 1. A certification that the unit being offered must be brand new and not a discontinued model. 2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent 3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 4. Bidder's certificate that the parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance	2	Unit	45,833.33	91,666.66

"There is no Health without Mental Health"

2	VDE22-02	SURGICAL HANDPIECE 45	7	Unit	37,400.00	261,800.00
		Specifications: <ol style="list-style-type: none"> 1. Fiber optic 2. Standard head 3. Titanium body 4. 45 angle head 5. Ceramic bearing 6. Clean head system 7. Push button chuck 8. Triple water jet 9. Replacement parts available in the market 10. Speed: 380,000 – 450,000 rev./min. 11. Head size: 11.2 x H 13.5 mm 12. For FG extra- long burs (25mm) only 13. 2 Holes Standard Requirements <ol style="list-style-type: none"> 1. A certification that the unit being offered must be brand new and not a discontinued model. 2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent 3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 4. Bidder's certificate that the parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: <ol style="list-style-type: none"> a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance 				
3	VDE22-03	MICROMOTOR PLUS LOW SPEED HANDPIECE	7	Unit	23,595.00	165,165.00
		<ol style="list-style-type: none"> 1. FEATURES: Forward/Reverse, Auto Cruise, Speed Control Pedal for On/Off 2. Specification: 4,000 – 35,000 rpm 3. Weight: 1.3kg – 2.2kg 4. Size: W 168 x D 168 x H 75 (mm) 5. Input: AC 116V OR 230V, 50/60HZ 6. Contra Angle: Speed 25,000 – 30,000rpm 7. Length: 85 – 90 mm 8. Lightweight: 0.1 kg – 0.8kg 9. Autoclave up to 35C 				

	10. Straight Handpiece: High Speed up to 40,000rpm 11. Length: 105 – 120 mm 12. Net Weight: 0.5 – 0.9 kg 13. Replacement parts available in the market Standard Requirements <ol style="list-style-type: none"> 1. A certification that the unit being offered must be brand new and not a discontinued model. 2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent 3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 4. Bidder's certificate that the parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: <ol style="list-style-type: none"> a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance 				
TOTAL					Php 518,631.66


Submitted by:


JERVIS D. LORETO, DMD, MMHoA
(End-User)

Recommending Approval:


ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
Chairperson, BAC for Equipment CY2022

Approved by:


NOEL V. REYES, MD, FPPA, MMHoA
Medical Center Chief II