



REPUBLIC OF THE PHILIPPINES
Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
Nueve de Febrero Street, Mandaluyong City, Philippines



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

Section VI
Schedule of Requirements

ITB No. 025-2022

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

ITEM NO.	DESCRIPTION	QTY	TOTAL	Delivered, Weeks/Months
VDE22-01	<p>EXTERNAL ORAL SUCTION MACHINE</p> <p>Specifications:</p> <ol style="list-style-type: none">1. Voltage: AC110V – AC220V, 50HZ / 60HZ2. Rated Power: 1000W – 1200W3. Control Mode: touch screen / two gears adjustable4. Suction flow: 3000 to 4000 liters/min5. Filtration efficiency: (0.3 um) 99.9%6. UV intensity (Distance = 1M): 1210uw/ cm27. UV light working life: 10000 hours8. UV wave length: 254mm9. Plasma sterilization working life: 6000 hours10. Noise: 52 – 60dB11. Net Weight: 20 – 25kg12. Pressure: 25kPa13. Warranty: 1 year (except filter and plasma lifetime service) <p>Standard Requirements</p> <ol style="list-style-type: none">1. A certification that the unit being offered must be brand new and not a discontinued model.2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit.4. Bidder's certificate that the parts shall be available at the authorized Philippine	2 units	91,666.66	60 to 90 calendar days

	<p>service center/s for a period of five (5) years after the warranty period.</p> <ol style="list-style-type: none"> 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: <ol style="list-style-type: none"> a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance 			
VDE22-02	<p>Surgical Handpiece 45</p> <p>Specifications:</p> <ol style="list-style-type: none"> 1. Fiber optic 2. Standard head 3. Titanium body 4. 45 angle head 5. Ceramic bearing 6. Clean head system 7. Push button chuck 8. Triple water jet 9. Replacement parts available in the market 10. Speed: 380,000 – 450,000 rev./min. 11. Head size: 11.2 x H 13.5 mm 12. For FG extra- long burs (25mm) only 13. 2 Holes <p>Standard Requirements</p> <ol style="list-style-type: none"> 1. A certification that the unit being offered must be brand new and not a discontinued model. 2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent 	7 units	261,800.00	60 to 90 calendar days

	<ol style="list-style-type: none"> 3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 4. Bidder's certificate that the parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: <ol style="list-style-type: none"> a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance 			
VDE22-03	<p>Micromotor Plus Low Speed Handpiece Specifications:</p> <ol style="list-style-type: none"> 1. FEATURES: Forward/Reverse, Auto Cruise, Speed Control Pedal for On/Off 2. Specification: 4,000 – 35,000 rpm 3. Weight: 1.3kg – 2.2kg 4. Size: W 168 x D 168 x H 75 (mm) 5. Input: AC 116V OR 230V, 50/60HZ 6. Contra Angle: Speed 25,000 – 30,000rpm 7. Length: 85 – 90 mm 8. Lightweight: 0.1 kg – 0.8kg 9. Autoclave up to 35C 10. Straight Handpiece: High Speed up to 40,000rpm 11. Length: 105 – 120 mm 12. Net Weight: 0.5 – 0.9 kg 13. Replacement parts available in the market <p>Standard Requirements</p> <ol style="list-style-type: none"> 1. A certification that the unit being offered must be brand new and not a discontinued model. 			

	<ol style="list-style-type: none"> 2. Current and valid Certificate of Manufacturer's compliance with ISO and or CE certificate or its equivalent 3. Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 4. Bidder's certificate that the parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 5. Certification that the brand has been in the local market for at least five (5) years with at least 2 to 3 installations. 6. Certification that the supplier will be responsible for the notification, transportation to the site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier 7. Certification to provide users' Manual in English language: <ol style="list-style-type: none"> a. Operational Manual – 2 copies b. Service Manual – 2 copies 8. Certification that the supplier / bidder shall provide free installation of the equipment 9. Warranty Certificate: one (1) year for parts and two (2) years for service from the date of delivery, inspection and acceptance 			
--	--	--	--	--

CONFORMED BY:
(Signature over printed name)

DATE: