



**BIDS AND AWARDS COMMITTEE**

Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

**SECTION VII**  
**Technical Specifications**

**ITB No. 034-2023**

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
1.	Hypochlorous Acid + Sodium Hypochlorite, 1000mL (with at least 24 months expiration).	
2.	Antimicrobial Dressing 1" Disk, 4mm, Sterile, Latex free. Chlorhexidine Topical - (for the skin) is used to clean the skin to prevent infection that may be caused by surgery, injection, or skin injury. Disk Applications: Central Venous Catheters, Dialysis Catheters, Arterial Catheters, PICC Lines, Peripheral IVs, Epidural Catheters, Implanted Venous Ports, Drains. (with at least 24 months expiration).	
3.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 6.0 (with at least 24 months expiration).	
4.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 6.5 (with at least 24 months expiration).	
5.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 7.0 (with at least 24 months expiration).	
6.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 7.5 (with at least 24 months expiration).	
7.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 8.0 (with at least 24 months expiration).	
8.	Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High Volume, Low Pressure Cuff, sterile, size 8.5 (with at least 24 months expiration).	
9.	Laryngeal Mask Airway (LMA), sterile, single use, size 4 (with at least 24 months expiration).	
10.	Laryngeal Mask Airway (LMA), sterile, single use, size 5 (with at least 24 months expiration).	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
11.	Tracheostomy tube cuffed fenestrated, size 6.0 (with at least 24 months expiration).	
12.	Tracheostomy Mask, ADULT (with at least 24 months expiration).	
13.	Spinal Anesthesia Needle, sterile, G. 20 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
14.	Spinal Anesthesia Needle, sterile, G. 22 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
15.	Enema Set - Enema can aluminum stainless steel with douche tube and douche fitting set (stop cork, vaginal pipe and rectal pipe).	
16.	Droplight / Medical Gooseneck Lamp - Adjustable height, Goose neck examination lamp is a multipurpose lamp attached to a flexible adjustable shaft that can be positioned in various ways. Stainless steel body, high quality	
17.	Anti Embolic Stockings Thigh High (Small) - For prevention of venous disease. Used for treatment after vein operation. Comfortable open-toe stockings	
18.	Anti Embolic Stockings Thigh High (Medium) - For prevention of venous disease. Used for treatment after vein operation. Comfortable open-toe stockings	
19.	Propan 2-ol, propan-1-ol, benzyl - C12-16- alkydimethylammonium chloride 1,000ml - Surgical and Hygienic Hand Disinfection Solution of medical grade with spray (with at least 24 months expiration	
20.	Polypropylene monofilament, round 5-0, 17mm, 1/2c, 90cm (with at least 24 months expiration).	
21.	Balance salt solution 500ml	
22.	Carbachol Ophthalmic Solution, 1mL	
23.	Corneal knife 2.75mm (6 pc/box)	
24.	Corneal Stab knife 15 degree (6 pc/box)	
25.	Crescent bevel up knife 2.3mm	
26.	Eye Pad Sterile, Constructed of absorbent cotton filler. Covered with soft non-woven fabric. Provides protection to the eye when required. Oval shape. Sterile packaging. 4 x 6 cm	
27.	Eye drape, sterile, halfbody with pouch	
28.	Silk Suture, Reverse Cutting, Size 6-0, Black, 18", Needle HE-7, ½ Circle (with atleast 24 months expiration).	
29.	Sodium Hyaluronate 3% Chondroitin sulfate 4% viscoelastic eye gel, 0.5mL	
30.	Sodium Hyaluronate 1.5% viscoelastic eye gel, 1mL	
31.	Tryphan blue ophthalmic solution, 1mL	
32.	Phacoemulsification basic pack (must be compatible with Stellaris Phaco machine)	
33.	Alcohol, Isopropyl 70% (3.785L/gallon) (with at least 24 months expiration).	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
34.	Gloves, Nitrile Examination, Disposable powder free, non sterile, ambidextrous, single use, size MEDIUM, with at least 24 months expiration.	
35.	Blood Glucose Test Strips with Lancet, with Strip Ejector, for in vitro diagnostic use, no coding, individually foil wrapped strips, small blood sample required at least (0.5µl), 5 second test time, 5 second blood sample reapplication time, top or end fill with visual confirmation, fill trigger ensures test starts only when sufficient blood sample has been applied. Must be compatible to existing Glucometer. (with at least 24 months expiration).	
36.	Blood Transfusion Set, sterilized by EO gas, non-toxic, no heavy metal & pyrogen free w/ vein needle G.18, double chamber w/ large mesh filter minimizing platelet loss & hemolysis, flexible table, individually packed into plastic pouch. (with at least 24 months expiration).	
37.	Disposable ECG Electrodes, PEDIATRIC, 32-36x36-44mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).	
38.	Endotracheal tube 7.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
39.	Gloves, 7.0 size, sterile, powder free, thick and elastic, not easily tore. (with at least 24 months expiration).	
40.	Infusion set, PEDIA, vented type (air cap) / 60 microdrops/mL/y-injection port/rubber/ 195cm length, with luer lock (with at least 24 months expiration).	
41.	KN95, 5 layer Respirator Mask, single, Individually pack, sterile, earloop type, with High Filtration Efficiency of ≥95%	
42.	Manual Resuscitator, ADULT with oxygen Reservoir bag, 2 Face Masks, 3 Guedel Airway and Oxygen Tubing, (Portable)	
43.	Mayo Table Painted. Minimizing Wasted Space in the. Operating Room. Mobility and Convenience. Customizable Organization. Increased Operating Room Efficiency. Surgical. Ergonomics Cleanliness. Stability. Durability and Reliability	
44.	Nasal cannula for oxygen, ADULT, made from nontoxic PVC compound, 1.8 meters of non-kink star lumen PVC tubing, nasal prongs made of soft PVC (with at least 24 months expiration)	
45.	Nebulizing kit with mask, inhalation adult set, includes 1 mask, 1 medication chamber cup, 1 oxygen tube (with at least 24 months expiration).	
46.	Sharp Disposal Container, double lid, with hermetic lock seal, 30 Liters	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
47.	Straight Catheter, Fr.12, transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic (with at least 24 months expiration).	
48.	Straight Catheter, Fr.14, transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic (with at least 24 months expiration).	
49.	Suction catheter, Fr. 14 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
50.	Suction catheter, Fr. 16 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
51.	Suction catheter, Fr. 18 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
52.	Syringe, Asepto, Irrigation Bulb 60cc, sterile, catheter tip, with plastic tip protector, allows one hand control, strong uniformly shaped barrel. (with atleast 24 months expiration).	
53.	Syringes, disposable, 30cc, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).	
54.	ECG Paper (Electrocardiograph Thermal Paper), 80mm x 20m	
55.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 6 (9cm x 15cm) P: 1 3/4 x 4 (4.5cm x 10cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
56.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 8 (10cm x 20cm) P: 1 3/4 x 6 (4.5cm x 15cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
57.	Transparent Dressing with absorbent pad, size: D: 3 1/2 x 10 (10cm x 25cm) P: 1 3/4 x 8 (4.5cm x 20cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
58.	Underpads Super Dry, 60x90cm, disposable. (with atleast 24 months expiration).	
59.	Urinal, plastic, for male, reusable, translucent, with lid, permits measuring ease and visualization contents	
60.	Zinc Oxide Adhesive Plaster, 3 inches, tan-coloured zinc oxide tape with spool and snap ring, 100% adhesive, Viscose tape with a high tensile strength. Strong, long-lasting adhesive with excellent initial grab. Water and dirt repellent for longer wear. The tape is is easy to tear and is permeable to both air and water vapour to prevent maceration. (with at least 24 months expiration).	
61.	Zinc Oxide Adhesive Plaster, 5 inches, tan-coloured zinc oxide tape with spool and snap ring, 100% adhesive, Viscose tape with a high tensile strength. Strong, long-lasting adhesive with excellent initial grab. Water and dirt repellent for longer wear. The tape is is easy to tear and is permeable to both air and water vapour to prevent maceration. (with at least 24 months expiration).	
62.	Anesthesia Set Flex, latex free, adult, breathing circuit with fitting mask - disposable anesthesia set consisting of 2 hose, connector, fastening ring 2 latex free breathing bag	
63.	Antiseptic surgical scrub solution. Non-ionic Surfactants, Has a pleasant perfume, and Tartrazine as colours. Easy to wash off. Can be used for neonates and thyroid patients. Not degraded due to bioburden. Effective in presence of blood, serious fluid, sweat etc. Non-staining and skin friendly. (5liters/gallon) (with at least 24 months expiration).	
64.	Cigarette drain (Penrose drainage tubing) size 1/2" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
65.	Cigarette drain (Penrose drainage tubing) size 1/4" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
66.	Cigarette drain (Penrose drainage tubing) size 1" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
67.	CO2 Absorber Lime Soda; Intersorb Plus Sodalime 5L Gallon, white to violet change. (with atleast 24 months expiration).	
68.	'High Level Disinfectant Solution Ortho-Phthalaldehyde, Ready-to-use for Medical Instruments, 6mins contact time, atleast 28days use-life. Atleast 5L/Gallon Contains: 0.55% Ortho-Phthalaldehyde active ingredients. Effective againts wide array of microorganisms.	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	It kills vegetative bacteria, mycobacteria viruses and fungi. (with atleast 24 months expiration).	
69.	PGA (Polyglycolic Acid/Polyglactin), braided, coated, braided synthetic absorbable, size 1, 70-90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 18 to 24 months expiration).	
70.	PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 36mm-40mm (with atleast 24 months expiration).	
71.	PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 3/0, 75-90cm, round needle, 1/2 circle, 36mm-40mm (with atleast 24 months expiration).	
72.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
73.	Polypropylene knitted non-absorbable Mesh, undyed, 7.6cm x 15cm (with atleast 24 months expiration).	
74.	Silk braided, non absorbable, size 2/0, 60cm, 12-15 threads/strands, no needles (with atleast 24 months expiration).	
75.	Silk braided, non absorbable, size 3/0, 75cm, 12-15 threads/strands, no needles (with atleast 24 months expiration).	
76.	Silk braided, non absorbable, size 0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm (with atleast 24 months expiration).	
77.	Silk braided, non absorbable, size 1, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm (with atleast 24 months expiration).	
78.	Silk braided, non absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with atleast 24 months expiration).	
79.	Silk braided, non absorbable, size 3/0, 75cm, round needle, 1/2 circle, 25mm- 37mm (with at least 24 months expiration).	
80.	Spinal Anesthesia Needle, sterile, G. 25 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
81.	Surgical Scrub with Chlorhexidine and brushes (with atleast 24 months expiration).	
82.	Arterio-Venous Fistula (AVF) kit, inclusive of 4 sterile cherry balls, drape, 2 sterile gauze, antiseptic pads, and 3 10cc syringes with needle, with at least 24 months expiration	
83.	AED PADS - CPR-D-padz – CPR Electrodes for Automated External Defibrillators (Compatible with AED Plus ZOLL Resuscitation Systems) - Pre-connected electrode with Real CPR Help feedback and See-Thru CPR. The one-piece design simplifies proper positioning on the chest. One size fits all, with the ability to expand for a larger chest. Application: Public Access – Non-clinical, EMS – Resuscitation. Five (5) year shelf-life from manufacturing date. To be used with Adults or Childrens 8 years and older or 25 kilos and heavier.	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
84.	BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER for PEDIA	
85.	BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER for ADULT	
86.	Catgut Chromic, Chromic natural absorbable, size 2/0, 75 cm, round taper point, 1/2 circle, 37mm (with atleast 24 months expiration).	
87.	Central Venous Pressure (CVP) Manometer Crystal clear manometer tube calibrated for accurate pressure readings and attached to a 3-way connector. Red floatation ball permits easy determination of meniscus level. With 3-way stopcock, 120cm long extension tubing, With 1" and 3/4" pole holder.	
88.	Clear Sterile Specimen bottle with rubber stopper, 10mL.	
89.	Extra Thin Patch Dressing 4x4inches/ 10x10cm, Thin flexible and versatile. Designed to reduce the risk of further skin breakdown due to friction. Can be used as a primary hydrocolloid dressing for dry to lightly exuding wounds.(with at least 24 months expiration).	
90.	Blood Pressure Latex Bulb With Air Release Valve Sphygmomanometer. Blood Pressure Monitor Manual Inflation Pump	
91.	Mask, surgical disposable, 3 PLY, Elastic without glass fibres, hypoallergenic, very low resistance to breathing, nose baradaptable, high filtration capacity, perfect fitting, EAR-LOOP	
92.	Foam Egg Mattress Size: 36x75x2.5" - has a convoluted shape which prevents and helps heal skin ulcers, eliminates bed sores, reduces pressure pains, allows air to flow and prevents your body from becoming too hot during sleep. Its peaks provide softness and comfort while its hollows provide passageway for air to pass through.	
93.	Air Purifier with Mosquito Catcher, 40sq.m, Emits 7,000 ions/cm <sup>3</sup> , 5-step Mosquito Catching Mechanism, Glue Sheet, UV Light, On/Off Plasmacluster Mode, Built-in Dust & Odor Sensor, Inverter Operation, 3-steps Filtration (New Pre-Filter, Deodorizing Filter, Electrostatic HEPA Filter)	
94.	Digital BP with Stand, Extra Large LCD Display. Ambient Temperature Display (C/F adjustable). Built-in buttin battery for keeping the date/time uninterruptedly during changing of batteries. IHB (irregular heart beat) Detector. Date/Time display. 2 users with 60 sets of memories each. Alarm clock reminding function. WHO blood pressure classification indicator. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.. Perform installation, demonstration for end-user and maintenance personnel before handover of the unit.	
95.	ECG Machine, digital and portable 3 channel; 12 lead EKG + electrode pad + lead wires orange; USB PC Software; with thermal printing system. For 80mm x 20m ECG Paper size.	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	Provide User's Manual and Service Manual. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.	
96.	Emergency Medical Trolley / Cart - size: 700x490x920mm (L*W*H*), 46kg. With Trolley body, Plastic-steel Columns, IV pole, Dust Basket, Utility Container, File Bag, Defibrillator Shelf, power outlet and hooks, Sliding Side Shelf, Stainless steel Guard Rail, Concave Top Covered Transparent Soft Plastic BAGs, Centralized Lock, Two (2) Small and Three (3) Big Drawer, Inner with partitions can be well organized, Drawers repositioned Easily, Luxurious Noiseless Castersinner with Partitions Can be well Organized, ABS Basket and Drawer Frame can be repositioned easily, Luxurious Whish Caster, two with brakes, CPR board, Oxygen Tank Holder, Needle Disposable Bottle.	
97.	Examining Bed – Hamilton Type, Stainless, Examination and treatment table, Hamilton Type, with drawer and compartment door, upholstered three section top, provided with stirrup and sliding foot stool. Size: 58 inc. x 20 inc. x 34 inc.	
98.	Fetal Doppler, Pocket type, high sensitive doppler probe, accurate FHR detection. Low ultrasound power, safe to the fetal. Optional display screen (value/diagram). Auto shut-off in 1 minute. Display 1.3 inches LCD. Ultrasound Frequency: 2.125MHZ-2.875MHZ. Comprehensive sensitivity: >90dB. FHR Range: 50/min~210/min. Display Resolution: 1bpm. FHR detection Dviation: -2/min~+2/min. Probe and doppler are seperated, more easier to operate. Rechargeable Battery, Charger included. Included Manual. Provide Calibration Certificate each unit. Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.. Perform installation, demonstration for end-user and maintenance personnel before handover of the unit.	
99.	IV Stand 4 Hooks Stainless - Heavy duty. Chrome plated steel tubing offers a round surface – no corners to collect dirt and grime, Well-balanced, stable no-tip cast aluminum base. Easy rolling wheel ball bearing nylon casters, Height range 52.5” – 94”, Twist lock for easy height adjustment, 4 Hooks, 4 Wheels	
100.	Fiber-Optic McIntosh (Curved 5 blades) for ADULT Laryngoscope ◆ LED Light Source ◆ Complete matt type-w/5 S.S Blades Size (0,00,1,2,3) or (0,1,2,3,4) -w/ carrying case	





**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	<p>Spacious basket on the reverse side for tubes and cuffs. Latex bulb to inflate the cuff. BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER. Non-wearing air-release valve with fine adjustment. Micro-filter to protect the release valve and measurement system. Specially hardened copper-beryllium membrane, which is resistant to aging, can withstand pressures of up to maximum 600 mm Hg Precision mechanism mounted on a special support, in order to ensure complete stability of the measuring system and the zero point. No zero-point fixation required. Maximum error tolerance: <math>\pm 3</math> mm Hg Provide Calibration Certificate each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration for end-user and maintenance personnel before handover of the unit.</p>	
105.	<p>Patient Transport Bed, Ambulance Collapsible Stretcher Automatic Loading 181kg Aluminum alloy, Adjustable height With IV Stand. Made of light weight aluminum alloy. This Stretcher considering narrow spaces such as elevators, moving the stretcher after folding it. The loading height of the stretcher is 64cm. It is made of high-strength aluminum alloy materials, use-safety and easy for sterilization and clean. High Position: 195x57x87.5cm Low Position: 198x57x22cm Maximum Angle of the Back: 58 degrees Self-Weight: 35.13kg Loading Bearing: 181kg</p>	
106.	<p>Stretcher, Foldable, high strength AL-alloy material, Surface made of nylon mesh, Light weight, portable, easy to carry out and safety to use, Folded size: 110 x 18 x10 cm, Laid flat size: 220 x53 x14cm, Load capacity: 160 kg.</p>	
107.	<p>Suction Machine Heavy Duty- multiple power supply, oil free lubrication pump; low noise; square negative pressure meter and plastic cover; negative pressure regulating system; power supply; Heavy duty monitor. Large Reservoir Bottle at 1000mL. Easy to use. Adjustable pressure at least negative 400mmHg. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.</p>	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
108.	Thermal Scanner with Automatic Alcohol Dispenser, For Automatic Alcohol Dispenser: Infrared Detection, Motion Activated, Content: 700ml., Low battery Indication, Easy battery replacement (AA size), No warring. For Infrared Thermal Scanner: Infrared Measurement, Distance Sensor, Speaker / Buzzer, Alarm Light, ON /OFF Button, Mode Button, Indicator, USB Port, Display, Response time : 0.5s, Infrared Measuring Range : 0°C – 40°C, Weight : 227g., Dimensions : 155 x 104 x 84mm, Stand, Infrared Thermal Scanner	
109.	Walker, Foldable. Anodized aluminum structure ensures maximum safety and strength while remaining lightweight, Two button push release is easy to extend/fold, Slip resistant rubber tips in the back with 5 inch wheels in the front provide maximum comfort while in use, Width 44cm Height adjusts from 32 inch to 39 inch in 1 inch increment to accommodate users with different needs.	
110.	Weighing Scale with Height and Weight Measurement, Heavy-duty solid stable platform. Superior Accuracy. Heavy duty understructure. Die cast beam, and aluminum on black insert. Easily read from either side. Factory assembled. All scales are tested with weights traceable to the National Institute of Standards and Technology (NIST). The die cast beam is dual reading and measures in BOTH pounds and kilograms. Capacity: 400 lb x 4 oz (180 kg x 100 g). Platform Size: 10-1/2 x 14-1/2 inches (267 x 368 mm). Height Rod: Measures 30-78 inches (76-198 cm). Scale height: 59 inches (1.5 m) Carton size: Height: 57" x Width: 19 1/2". Length: 5 1/2", 20kg Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.	
111.	BATHROOM WEIGHING SCALE HEAVY DUTY - made with non-skid vinyl mat features white, baked enamel finish battery-free mechanised operation analogue. Capability: 150kg maximum 500g graduation, sturdy, big metallic non-slip podium, easy-to-read dial, used to monitor body weight. Dimension: 32cm x 17cm x 40cm Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.	
112.	Wheelchair, 4 wheels, heavy duty (2 rear wheels and 2 front swivel castors), Seat and backrest providing sitting support for the wheelchair user, Foot supports. Armrests, Rear wheel locks (to park the wheelchair), Folding frame wheelchairs and buggies allow	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	folding through cross-brace or other folding mechanism, Rigid frame wheelchairs and buggies must offer mechanism to fold and/or dismantle the device into smaller, separate components	
113.	ECG Trunk Cable with 3-lead wire. (must be compatible to <b>GE B125</b> Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.	
114.	SPO2 Cable Trusignal with SPO2 Sensor (must be compatible to GE B125 Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.	
115.	NIBP Cuff and hose for ADULT (must be compatible to GE B125 Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.	
116.	NIBP Cuff and hose for PEDIA (must be compatible to GE B125 Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.	
117.	Negatoscope (Double) - The Fluorescent Illuminator provides a soft diffuse bluish white light found to be most satisfactory for film illumination. Made of ordinary metal with white paint or stainless steel. Negatoscope Double. High color temperature. Low power consumption. Size: L:30 x W: 3" x H: 20" A device for reading x-rays – a negatoscope – greatly simplifies the work of a diagnostic physician and improves the accuracy of interpretation of an x-ray. A negatoscope is a parallelepiped made of a metal profile or durable plastic with a frontal plane (screen) of opaque white acrylic glass or plexiglass.	
118.	Needle Holder 14cm	
119.	Needle Holder 16cm	
120.	Needle Holder 18cm	
121.	Mixer Fine, Mixer Forceps, curved, 14cm	
122.	Mixer Fine, Mixer Forceps, delicate, curved, 18cm	
123.	Skin Hook, Kilner Skin Hook, 1 prong, 15cm	
124.	Suction Metal Tip, AS Pool Suction tube, straight, with outer tube to be screwed off, 10,0,mm, TL 22cm	
125.	Long De Bakey 24cm, AS de Bakey Atrauma forceps, jaw 20 mm straight, 24cm	
126.	Long De Bakey, AS de Bakey Atrauma forceps, jaw 15 mm straight, 20cm	
127.	Big Kelly, AS Kelly-Rankin Forceps, straight, 16cm	
128.	Big Kelly, AS Kelly-Rankin Forceps, curved, 16cm	
129.	Big Kelly, AS Kelly Forceps, Fig. 1, curved, 19cm	
130.	Kidney Basin, Kidney Bowls 170x100x35mm	
131.	Kidney Basin, Kidney Bowls 250x140x40mm	
132.	Kidney Basin, Kidney Bowls 275x150x45mm	

**INSTRUCTION:** Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
133.	Bowls, AS Round Bowls, 150 x 75mm	
134.	Bowls, AS Round Bowls, 160 x 75mm	
135.	Mayo scissors, straight 19cm	
136.	Mayo scissors, straight 23cm	
137.	Mayo Scissors, curved 14cm	
138.	Mayo Scissors, curved 17cm	
139.	Mayo Scissors, curved 19cm	
140.	Mayo Scissors, curved 23cm	
141.	Mayo-Stille Scissors, straight, 15cm	
142.	Metzenbaum Scissors, Fig. 1, straight 13cm	
143.	Metzenbaum Scissors, Fig. 1, straight 14cm	
144.	Metzenbaum Scissors, Fig. 1, straight 18cm, SUPERCUT	
145.	Metzenbaum Scissors, Fig. 1, curved, 13cm	
146.	Metzenbaum Scissors, Fig. 1, curved, 14cm	
147.	Metzenbaum Scissors, Fig. 1, curved, 18cm	
148.	Metzenbaum Scissors, Fig. 1, curved, 23cm	
149.	Metzenbaum Scissors, Fig. 1, curved, 25cm	
150.	Metzenbaum Scissors, Fig. 1, curved, 28cm	
151.	Metzenbaum Scissors, Fig. 1, curved, delicate pattern, 14cm	
152.	Metzenbaum Scissors, Fig. 1, curved, delicate pattern, 20cm	
153.	Ochsner-Kocher Forceps 1:2, straight, 14cm	
154.	Ochsner-Kocher Forceps 1:2, straight, 16cm	
155.	Ochsner-Kocher Forceps 1:2, straight, 18cm	
156.	Ochsner-Kocher Forceps 1:2, curved, 16cm	
157.	Ochsner-Kocher Forceps 1:2, curved, 18cm	
158.	Ochsner-Kocher Forceps 1:2, curved, 22cm	
159.	Halsted-Mosquito Forceps, straight, 12.5cm	
160.	Halsted-Mosquito Forceps, curved, 12.5cm	
161.	Micro-Mosquito Forceps, straight, 12.5cm	
162.	Micro-Mosquito Forceps, curved, 10cm	
163.	Micro-Mosquito Forceps, curved, 12.5cm	
164.	Blade Handle No.3, Scalpel Handle, Standard No. 3, 12cm	
165.	Blade Handle No.3, Scalpel Handle with scale	
166.	Blade Handle Long, Scalpel Handle, Standard No.4	
167.	Vaginal Retractor, Cusco Vaginal Specula, Fig.1, Standard, 75x32mm	
168.	Vaginal Retractor, Cusco Vaginal Specula, Swiss Model, Fig.2 100x25mm	
169.	Vaginal Retractor, Sims Vaginal Specula, Fig.2, 70x30mm / 70x35mm	
170.	Ronguer Pennybacker Bone ronguer, straight, 20.5cm	
171.	Bone Curette Volkmann Bone curette, 14cm, double	
172.	Gigli Saw, Gigli Wire saw, fine 500mm	

**INSTRUCTION:** Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
173.	Gigli Saw, Gigli Handle	
174.	Vein Retractor Cushing Nerve Hook, Fig.1, 19cm	
175.	Army Navy Retractor (1 pair) Size:8"	
176.	Adson Dressing Forcep with Teeth 4 3/4"	
177.	Allis Tissue Forceps 7.0"	

**CONFORMED BY:** \_\_\_\_\_  
 Authorized Representative  
 (Signature over printed name)