



REPUBLIC OF THE PHILIPPINES
Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
Nueve de Febrero Street, Mandaluyong City, Philippines



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

SECTION VII
Technical Specifications

ITB No. 046-2023

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
1.	Distilled Water, 10 liters / bottle (for Clinical Lab, Dental Clinic and OR use)	
2.	Zinc Oxide Adhesive Plaster, 2 inches, with spool and snap ring, 100% adhesive. Tape contains natural rubber latex. Viscose tape with a high tensile strength. Strong, long-lasting adhesive with excellent initial grab. Water and dirt repellent for longer wear. The tape is breathable, reducing the risk of skin maceration (with atleast 24 months expiration).	
3.	Zinc Oxide Adhesive Plaster, 3 inches, with spool and snap ring, 100% adhesive. Tape contains natural rubber latex. Viscose tape with a high tensile strength. Strong, long-lasting adhesive with excellent initial grab. Water and dirt repellent for longer wear. The tape is breathable, reducing the risk of skin maceration (with at least 24 months expiration).	
4.	Zinc Oxide Adhesive Plaster, 5 inches, with spool and snap ring, 7.5cm x 5m, 100% adhesive. Tape contains natural rubber latex. Viscose tape with a high tensile strength. Strong, long-lasting adhesive with excellent initial grab. Water and dirt repellent for longer wear. The tape is breathable, reducing the risk of skin maceration (with at least 24 months expiration).	
5.	Epidural set, Epidural Catheter, G. 18, for continuous epidural anesthesia with Tuohy type needle, catheter (closed tip, 3 lateral eyes) with connector, epidural flat filter and LOR syringe, filter 0.2 micron, Loss-of-Resistance syringe - Luer Slip (with atleast 24 months expiration).	
6.	PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 0, 75cm-90cm, cutting needle, 1/2 circle / 3/8 circle, 37mm / 40mm(with atleast 24 months expiration).	
7.	PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 6/0, 75cm-90cm, cutting needle, 0.7 metric, 18", 45cm, 7-6mm, 1/4c, spatulated (with atleast 24 months expiration).	
8.	PGA (Polyglycolic Acid/Polyglactin) 7-0 cutting, 5.5mm, 1/2c, spatulated, 18", 45cm, 0.5metric	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
9.	PGA (Polyglycolic Acid/Polyglactin) 8-0 cutting, 6.5mm, 3/8C, spatulated, 12", 30cm,	
10.	Polypropylene monofilament, non absorbable, size 3/0, 75 cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 37mm (with atleast 24 months expiration).	
11.	Polypropylene Suture with round needle 6-0, 24" (60cm), 9mm, 3/8 (with atleast 24 months expiration).	
12.	Rust Remover - Acidic cleaning concentrate and troubleshooting aid, 1 liter - can be used to remove silicate deposits in washer, removal of annealing colours, water spots, transferred/initial rust, and encrustations, based on acids and surfactants and effectively removes inorganic contamination, exclusively used with stainless steel instruments, and alkali-free.	
13.	Silk braided, non absorbable, size 4/0, 75-90cm, round needle, 1/2 circle, 20mm / 22mm / 26mm (with at least 24 months expiration).	
14.	Steri strip, 12mm x 100mm	
15.	Steri strip, 6mm x 100mm	
16.	Sterile water, 100ml (with atleast 24 months expiration).	
17.	Pre-heparinized Arterial Blood Gas Syringe 3cc (with at least 24 months expiration).	
18.	Total hardness paper test strip, 100 strips per bottle, with shelf life of at least 18 months	
19.	Bedside screen with 3-4 fold. Overall dimension 1680mm(H)x 2450mm (W)	
20.	UV light water sterilizer, 12gpm, 220v-60	
21.	"Double-J Ureteral Stent Set. All items must be sterile. Included in the set: (1 pc) Guidewire with 0.035 inch or 0.038 inch in diameter, length of 145-155cm that should fit inside the lumen of the included double-J stent and the included open ended stent; (1 pc) Double-J Stent Pusher that should be radioopaque or should have a portion detectable by fluoroscopy; (1 pc) Open End Stent. Diameter: 6Fr (1 pc) Double-J Ureteral Stent. Diameter: 6Fr. Length: 24 cm."	30
22.	Double-J Tumor Ureteral Stent Set. All items must be sterile. Included in the set: (1 pc) Guidewire with 0.035 inch or 0.038 inch in diameter, length of 145-155cm that should fit inside the lumen of the included double-J stent and the included open ended stent; (1 pc) Double-J Stent Pusher that should be radioopaque or should have a portion detectable by fluoroscopy; (1 pc) Open End Stent. Diameter: 6Fr (1 pc) Double-J Ureteral Stent. Diameter: 6Fr. Length: 24 cm.	
23.	Double-J Ureteral Stent Set. All items must be sterile.	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	<p>Included in the set: (1 pc) Guidewire with 0.035 inch or 0.038 inch in diameter, length of 145-155cm that should fit inside the lumen of the included double-J stent and the included open ended stent; (1 pc) Double-J Stent Pusher that should be radioopaque or should have a portion detectable by fluoroscopy; (1 pc) Open End Stent. Diameter: 6Fr (1 pc) Double-J Ureteral Stent. Diameter: 6Fr. Length: 26 cm.</p>	
24.	<p>Open ended Ureteral Catheter. Must be Sterile - Diameter: 6Fr - Length: 70cm - Tip: Soft/Open Tip</p>	
25.	<p>Sterile Percutaneous Nephrostomy Tube Set that includes the following: (1 pc) 8Fr nephrostomy tube, Length: at least 30cm; (1 pc) Metallic probe to straighten cannula; (1 pc) Flexible probe to straighten cannula; (1 pc) Gauge 18, 20 cm, introducer needle with hollow outer sheath;(1 pc) J-tip guide wire with a 0.035 inch diameter, length of 90cm-150cm; (4 pcs) tract dilators — 6Fr, 8Fr, 10Fr, 11Fr; (1 pc) Contrasting needle; (1 pc) Connecting Tube from Nephrostomy to Urine Bag; (1 pc) Blade 11 Scalpel; (1 pc) Stopcock</p>	
26.	<p>Sterile Percutaneous Nephrostomy Tube Set that includes the following: (1 pc) 10Fr nephrostomy tube, Length: at least 30cm; (1 pc) Metallic probe to straighten cannula; (1 pc) Flexible probe to straighten cannula; (1 pc) Gauge 18, 20 cm, introducer needle with hollow outer sheath; (1 pc) J-tip guide wire with a 0.035 inch diameter, length of 90cm-150cm; (4 pcs) tract dilators — 6Fr, 8Fr, 10Fr, 11Fr; (1 pc) Contrasting needle; (1 pc) Connecting Tube from Nephrostomy to Urine Bag; (1 pc) Blade 11 Scalpel; (1 pc) Stopcock</p>	
27.	<p>Sterile Percutaneous Nephrostomy Tube Set that includes the following: (1 pc) 12Fr nephrostomy tube, Length 30cm; (1 pc) Metallic probe to straighten cannula; (1 pc) Flexible probe to straighten cannula; (1 pc) Gauge 18, 20 cm, introducer needle with hollow outer sheath; (1 pc) J-tip guide wire with a 0.035 inch diameter, length of 90cm-150cm; (5 pcs) tract dilators — 6F, 8F, 10F, 12F, 13F (1 pc) Contrasting needle; (1 pc) Connecting Tube from Nephrostomy to Urine Bag; (1 pc) Blade 11 Scalpel; (1 pc) Stopcock</p>	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
28.	<p>Sterile Percutaneous Nephrostomy Tube Set that includes the following:</p> <ul style="list-style-type: none"> (1 pc) 14Fr nephrostomy tube, Length: at least 30cm; (1 pc) Metallic probe to straighten cannula; (1 pc) Flexible probe to straighten cannula; (1 pc) Gauge 18, 20 cm, introducer needle with hollow outer sheath; (1 pc) J-tip guide wire with a 0.035 inch diameter, length of 90cm-150cm; (6 pcs) tract dilators — 6F, 8F, 10F, 12F, 14F, 15F (1 pc) Contrasting needle; (1 pc) Connecting Tube from Nephrostomy to Urine Bag; (1 pc) Blade 11 Scalpel; (1 pc) Stopcock 	
29.	<p>Sterile Suprapubic Cystostomy Set- 15Fr or 16Fr or 18Fr Malecot type or Foley type suprapubic catheter- Catheter should have an open end where a guidewire may be passed thru- Balloon of catheter: 5-10mL- Introducer needle- Drainage tube</p>	
30.	<p>Sterile Urinary Drainage Bag for specific use for postoperative irrigation and drainage of bladder</p> <ul style="list-style-type: none"> - With universal connector compatible with most indwelling foley catheters - With hooks/cords to secure the bag at bedside - With anti-reflux valve to help prevent ascending infections - Capacity: 4000mL-5000mL - T-type outlet - Latex-free - Single-use 	
31.	<p>Sterile, disposable core needle biopsy gun</p> <ul style="list-style-type: none"> - Gauge:18 - Length: 25 cm disposable needle that is detectable by ultrasound and can get ample amount of prostatic tissue on transrectal prostate ultrasound-guided biopsies. <p>Length and girth of the needle should be compatible with the existing ultrasound machine of the Renal Lithotripsy Center.</p>	
32.	<p>Urological Guidewire</p> <ul style="list-style-type: none"> - Sterile - Polytetraflouroethylene coated shaft - Very Stiff-bodied - Diameter: 0.038in - For general urologic procedures - Tip: straight, atraumatic, 3-7cm flexible - Length: 140cm-200cm 	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
33.	Urological Guidewire - Sterile - Polytetrafluoroethylene coated shaft - Nitinol core - Hydrophilic, atraumatic, flexible tip (3-5cm) - Diameter: .035 inch - Length: 150-200 cm - Tip Shape: Straight	
34.	Urological Guidewire - Sterile - Bentson-Type - Diameter: 0.035 inch - Length: 140cm-200cm - Tip: Straight 5-7cm flexible.	
35.	Urological Guidewire - Sterile - Nitinol-core guidewire - Polytetrafluoroethylene coating - Striped pattern for better visualization - Flexible distal tip 3-7 cm. - Diameter: 0.035in. - Length: 150-200cm.	
36.	Urological Guidewire - Sterile - Nitinol-core guidewire - Polytetrafluoroethylene coating - Striped pattern for better visualization - Flexible distal tip 3-7 cm. - Diameter: 0.032in. - Length: 150-200cm.	
37.	Urological Guidewire- Sterile- Stainless Steel body- J-tip - The tip is atraumatic and flexible by 3-5cm- The tip is ball type- Body should be stiff. - Length: 90cm-150cm. - Diameter: 0.038 inch	
38.	Pre-filled syringe 11mL containing sterile lubricating jelly, 2% lidocaine and 0.255 chlorhexidine gluconate. Syringe with long nozzle and blunt edge for easy insertion to urethra and to avoid mucosal injury	
39.	Air Purification Filter - Adopt HEPA+ activated carbon composite filter. Improved purification effect. Volume reduction. Must be compatible to existing unit.	
40.	Elastic Bandage Velcro Closure, 4 inches x 5yards, sterile, cotton, latex free, permanent, duralastic, (for pressurized dressing)	
41.	Elastic Bandage Velcro Closure, 6 inches x 5yards, sterile, cotton, latex free, permanent, duralastic, (for pressurized dressing)	
42.	Automatic Labeler System Specification for Steam: Upper line must have batch number, sterilizer number 2. Center line must have 12 or more alphanumeric characters, initial of the person in charge and chemical	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	indicator for steam color pink then turn dark brown or black after sterilization 3. Lower line MUST have expiration date. 750pcs or more sticker/roll (with one (1) free gun labeler for Operating Room,) with 3 FREE INK REFILL	
43.	Spray foam - Ready-to-use enzymatic pre-cleaning with disinfection foam based with active ingredients such as protease and surfactants for keeping surgical instruments moist, active against enveloped viruses, bactericidal, levurocidal, has N,N-Didecyl-N-methyl-poly (oxyethyl) ammonium propionate and N-(3-Aminopropyl)-N-dodecylpropane-1, 3-diamine, 0.5L	
44.	Surgical Scrub with Chlorhexidine and brushes(with atleast 24 months expiration).	
45.	Dialysis Needle for Fistula and Graft, G16, 1.6 x 25mm, with metal stainless needle, with back eye, and sharp cutting edge, providing smooth insertion, fixed wing, with at least 24 months expiration	
46.	Surface Disinfectant, Towelletes / Wipes. Cleaner and disinfectant in one. Kill clinically relevant organisms such as bacteria, viruses, fungi and TB in just 3 minutes. Ready-to-use with no dilution or mixing required. Clean, disinfect and decontaminate the hard, non-porous surfaces with speed and confidence. Durable, nonwoven towellettes will not bunch up during use.	

CONFORMED BY: _____
 Authorized Representative
 (Signature over printed name)