



REPUBLIC OF THE PHILIPPINES
 Department of Health
NATIONAL CENTER FOR MENTAL HEALTH
 Nueve de Febrero Street, Mandaluyong City, Philippines



BIDS AND AWARDS COMMITTEE

Telephone No. 531-9001 loc. 242

Telefax No. 5318318

E-mail: bacncmh@yahoo.com

Website: www.ncmh.gov.ph

SECTION VII
Technical Specifications

ITB No. G-005-2024-PB

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
1.	Abdominal Binder - made of cotton pile fabric and polyurethane foam, lined with tricot for additional comfort elastic. Velcro or hook and loop closures comes in 8" x 52". Dimensions: 11cm x 8cm x 20cm, recommended for wear post-abdominal surgery, Caesarean section, bariatric surgery, exploratory laparotomy, hysterectomy, tummy tuck spinal surgery Dimension: 31.5cm x 3cm x 8cm	
2.	AED PADS - CPR-D-padz – CPR Electrodes for Automated External Defibrillators (Compatible with AED Plus ZOLL Resuscitation Systems) - Pre-connected electrode with Real CPR Help feedback and See-Thru CPR. The one-piece design simplifies proper positioning on the chest. One size fits all, with the ability to expand for a larger chest. Application: Public Access – Non-clinical, EMS – Resuscitation. Five (5) year shelf-life from manufacturing date. To be used with Adults or Childrens 8 years and older or 25 kilos and heavier.	
3.	Alcohol Breath Analyzer	
4.	Alcohol, Isopropyl 70% (with 3.785L/gallon) (with at least 24 months expiration).	
5.	Alcohol Swab/ Alcohol Prep Pads 2x2 (with at least 24 months expiration).	
6.	Aneroid Sphygmomanometer with Stethoscope, ADULT - Latex Free with single tube. Range: 0-300mmHg. Precision valve with slow deflation rate. Extra large inflation bulb. Gauge with internal diameter of 65mm. Velcro cuff, for adults, washable made in nylon with ring. Sphygmomanometer, Long lasting, thicker rubber bladder, Authentic, luminous dial face (Durable Adult, Inflation System, Genuine brass gauge. With Handy zipper case, Durable adult inflation system, Authentic luminous dial face, Long lasting thicker rubber bladder). With Certificate of Calibration (PORTABLE)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
7.	Aneroid Sphygmomanometer with Stethoscope, PEDIATRIC, Manometer, range 0-300 mm. accuracy +/-3 mm. Newborn and child velcro cuff. Standard model in vinyl zipper case. (PORTABLE)	
8.	Arm Sling - adjustable shoulder strap allows for a custom comfortable fit. Supports left or right arm. 70% polyester, Latex free. MEDIUM	
9.	Arm Sling - adjustable shoulder strap allows for a custom comfortable fit. Supports left or right arm. 70% polyester, Latex free. LARGE	
10.	Bedpan, Plastic	
11.	Biohazard Needle Disposal Container, 3L capacity; puncture-proof hard plastic; with locks for final disposal and safe transport.	
12.	BP App, Digital - Fully automatic Personalized inflation for maximum comfort. Quick deflation releases valve for speedy measurement. With heartbeat indicator and velcro cuff.	
13.	BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER for PEDIA	
14.	BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER for ADULT	
15.	Blood Glucose Test Strips and Lancet with free 100pcs Glucometer with Strip Ejector, for in vitro diagnostic use, no coding, individually foil wrapped strips, small blood sample required atleast (0.5µl), 5 second test time, 5 second blood sample reapplication time, top or end fill with visual confirmation, fill trigger ensures test starts only when sufficient blood sample has been applied. Must be compatible to existing Glucometer. (with at least 24 months expiration).	
16.	Blood Pressure Latex Bulb With Air Release Valve Sphygmomanometer. Blood Pressure Monitor Manual Inflation Pump	
17.	Blood Transfusion Set, sterilized by EO gas, non-toxic, no heavy metal & pyrogen free w/ vein needle G.18, double chamber w/ large mesh filter minimizing platelet loss & hemolysis, flexible table, individually packed into plastic pouch. With standard length of tube of 180cm and dropping of 20 drops/mL. (with at least 24 months expiration).	
18.	Bouffant Cap, Single Elastic, Strip type, disposable	
19.	Cervical Collar, small, 27.9-34.3cm, 17-mm foam, with cotton stockinette, contoured, with hook and lock closure system	
20.	Cervical Collar, medium, 34.3-39.4 cm, 17-mm foam, with cotton stockinette, contoured, with hook and lock closure system	
21.	Cervical Collar, large, 39.4-44.5 cm, 17-mm foam, with cotton stockinette, contoured, with hook and lock closure system	
22.	Cervical Collar, extra large, 44.5-48.3 cm, 17-mm foam, with cotton stockinette, contoured, with hook and lock closure system	
23.	Colostomy bag with wafer 60-70mm (with at least 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
24.	Cotton Tip Applicator, sterile, individually packed x 2 pcs. (with at least 24 months expiration).	
25.	Cotton Absorbent Balls, 400 grams - made from 100% pure natural cotton fibers, highly absorbent, super soft texture, clean and safe, economical, approximately 1,000 pieces per pack. (with at least 24 months expiration).	
26.	Disposable ECG Electrodes, ADULT, 36-37x40-45mm (oval) / circular snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).	
27.	Disposable ECG Electrodes, PEDIATRIC, 32x36mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).	
28.	Effervescent Chlorine Tablets. 2.5 g Sodium dichloro-s-triazinetriene, US EPA approved with Certificate, described as highly effective against both envelope and unenvelope virus on master label. 100's tablets per jar with at least 3 years of expiration	
29.	Egg Mattress Size: 36x75x2.5" - has a convoluted shape which prevents and helps heal skin ulcers, eliminates bed sores, reduces pressure pains, allows air to flow and prevents your body from becoming too hot during sleep. Its peaks provide softness and comfort while its hollows provide passageway for air to pass through.	
30.	Emergency Kit - 1 pack Cotton buds (72's), 1 roll Gauze Bandage 3", 4 pcs. Gauze Pad 2x2x8ply sterile, 1 pc. Triangular Bandage 96x96x136cm, 1 pc. CPR Mask Disposable, 1 pc. Emergency Blanket 130x210cm, 1 pair Vinyl Gloves, 1 pc. Ice Pack Instant, 1 pc. Cooling Fever Patch, 1 roll Surgical tape 1/2", 1 pc. Tourniquet Rubber Flat, 10 pcs. Safety Pins small, 1 pc. Whistle, 1 pc. Scissor small, 1 pc. Flashlight, 1 pc. Multi Tool kit, 1 tube Antibiotic, 10 pcs. Alcohol Swab, 10 pcs. Povidone Swab 1 pc. First Aid Bag	
31.	Endotracheal tube 2.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
32.	Endotracheal tube 3.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
33.	Endotracheal tube 3.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	

CPA

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
34.	Endotracheal tube 4.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
35.	Endotracheal tube 4.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
36.	Endotracheal tube 5.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
37.	Endotracheal tube 5.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
38.	Endotracheal tube 6.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
39.	Endotracheal tube 6.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
40.	Endotracheal tube 7.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
41.	Endotracheal tube 7.5 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
42.	Endotracheal tube 8.0 high volume, low pressure, blue line design, very smooth murphy eye, very smooth edge to avoid trauma, with cuff, size markings of ET Tube in pilot balloon, non pyrogenic, Radio-Opaque (with at least 24 months expiration).	
43.	Enema Can Aluminum Stainless Steel with douche tube and Douche fitting set (stop cork, vaginal pipe & rectal pipe). Dimension: 15cm x 15cm x 17cm	
44.	Foley Catheter, Sterile, Three-way, Fr. 24 - Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 50mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
45.	Foley Catheter, Sterile, Three-way, Fr. 22 - Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 100mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
46.	Foley Catheter 3-way, Fr.20, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 50mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
47.	Foley Catheter 2-way, Fr.20, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 50mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
48.	Foley Catheter 2-way, Fr.18, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 100mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
49.	Foley Catheter 3-way, Fr.18, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 50mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
50.	Foley Catheter 2-way, Fr.16, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 50mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
51.	Sterile Foley Catheter: All-silicon, Disposable, Two-way, Fr. 16, Needleless, Balloon should have high tensile strength, Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should with stand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet.	
52.	Sterile Foley Catheter: All-silicon, Disposable, Two-way, Fr. 14, Needleless, Balloon should have high tensile strength, Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should with stand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet.	
53.	Foley Catheter 2-way, Fr.14, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 30mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
54.	Foley Catheter 2-way, Fr.12,Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 30mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
55.	Foley Catheter 2-way, Fr.10, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 30mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
56.	Foley Catheter 2-way, Fr.8, Silicon elastomere coated, latex, disposable, Needleless, Balloon should have high tensile strength and can withstand up to 30mL without bursting. Inflow and outflow eyelets should be atraumatic to the urethral and bladder mucosa. The tip should not easily kink on manipulation and insertion. The lumen should withstand manual and repetitive suctioning without collapsing. The body should be easy to handle and should not be slippery to manipulate. The item should have only one round atraumatic inlet and one round atraumatic outlet at the tip. (with at least 24 months expiration).	
57.	Gauze non-woven sponges, 5cm x 5cm (100's/pack) (for Dental Clinic use)	
58.	Gauze, surgical bolt, white, 36 inches x 50 yards 2ply, 28 x 24 mesh, with x-ray detectable line.	
59.	Gloves, surgical, sterile, latex, powder free, size 6, with readable ink size marking in the packaging and gloves, at least 10" long, with AQL of at least 0.65, as indicated in the packaging or proof of laboratory test, durable, thick, with beaded cuff, comfortable to use, fits and follow the natural finger shape, can hold surgical sleeves, can withstand to chemotherapy drugs, can provide effective insulation (protection from electrosurgery burn), low endotoxin and non-pyrogenic. (with at least 24 months expiration).	
60.	Gloves, surgical, sterile, latex, powder free, size 6.5, with readable ink size marking in the packaging and gloves, at least 10" long, with AQL of at least 0.65, as indicated in the packaging or proof of laboratory test, durable, thick, with beaded cuff, comfortable to use, fits and follow the natural finger shape, can hold surgical sleeves, can withstand to chemotherapy drugs, can provide effective insulation (protection from electrosurgery burn), low endotoxin and non-pyrogenic. (with at least 24 months expiration).	
61.	Gloves, surgical, sterile, latex, powder free, size 7.0, with readable ink size marking in the packaging and gloves, at least 10" long, with AQL of at least 0.65, as indicated in the packaging or proof of laboratory test, durable, thick, with beaded cuff, comfortable to use, fits and follow the natural finger shape, can hold surgical sleeves, can withstand to chemotherapy drugs, can provide	

QA

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	effective insulation (protection from electrosurgery burn), low endotoxin and non-pyrogenic. (with at least 24 months expiration).	
62.	Gloves, surgical, sterile, latex, powder free, size 7.5, with readable ink size marking in the packaging and gloves, at least 10" long, with AQL of at least 0.65, as indicated in the packaging or proof of laboratory test, durable, thick, with beaded cuff, comfortable to use, fits and follow the natural finger shape, can hold surgical sleeves, can withstand to chemotherapy drugs, can provide effective insulation (protection from electrosurgery burn), low endotoxin and non-pyrogenic. (with at least 24 months expiration).	
63.	Gloves, surgical, sterile, latex, powder free, size 8.0, with readable ink size marking in the packaging and gloves, at least 10" long, with AQL of at least 0.65, as indicated in the packaging or proof of laboratory test, durable, thick, with beaded cuff, comfortable to use, fits and follow the natural finger shape, can hold surgical sleeves, can withstand to chemotherapy drugs, can provide effective insulation (protection from electrosurgery burn), low endotoxin and non-pyrogenic. (with at least 24 months expiration).	
64.	Gloves, Latex, examination, disposable, powder free, non-sterile, ambidextrous, single use, size LARGE (with at least 24 months expiration)	
65.	Gloves, Latex examination, disposable, powder free, non-sterile, ambidextrous, single use, size MEDIUM (with at least 24 months expiration)	
66.	Gloves, Nitrile, examination, disposable, powder free, non-sterile, ambidextrous, single use, size LARGE (with at least 24 months expiration)	
67.	Gloves, Nitrile, examination, disposable, powder free, non-sterile, ambidextrous, single use, size MEDIUM (with at least 24 months expiration)	
68.	Gloves, Nitrile, examination, disposable, powder free, non-sterile, ambidextrous, single use, size SMALL (with at least 24 months expiration)	
69.	Heparin cap luer lock in-stopper ABS or PC material, <u>needle-free</u> disposable, made from fine material (medical grade PVC) and easy to puncture, latex free, sterile blister packaging (with at least 24 months expiration)	
70.	Household Water Purification Tablets - 67mg tablet, 1 tablet purifies 20 liters. Use of drinking water disinfection. (with at least 24 months expiration).	
71.	Hydrogen Peroxide 3% Antiseptic Solution, 10 Volumes, 500mL per bottle (with at least 24 months expiration).	
72.	Hypodermic / InjectionTray, stainless steel, 8" x 3" x 2", with Strap Handle Lid	
73.	Infusion set, ADULT, 20 macrodrops/mL/y-injection port/rubber/ 215 cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
74.	Infusion set, PEDIA - 60 microdrops/mL/y-injection port/rubber/ 195cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration).	
75.	IV CANNULA G18 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 45mm length (with at least 24 months expiration)	
76.	IV CANNULA G20 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 32mm length (with at least 24 months expiration)	
77.	IV CANNULA G22 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 25 mm length (with at least 24 months expiration)	
78.	IV CANNULA G24 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length (with at least 24 months expiration)	
79.	IV CANNULA G26 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length (with at least 24 months expiration)	
80.	IV Splints Brace With straps for CHILD - Splint application to avoid unwanted movement of limbs and kinking of intravenous catheters, allowing free flow of intravenous fluids.	
81.	IV Splints Brace With straps for ADULT- Splint application to avoid unwanted movement of limbs and kinking of intravenous catheters, allowing free flow of intravenous fluids.	
82.	IV Stand 4 Hooks Stainless - Heavy duty. Chrome plated steel tubing offers a round surface – no corners to collect dirt and grime, Well-balanced, stable no-tip cast aluminum base. Easy rolling wheel ball bearing nylon casters, Height range 52.5” – 94”, Twist lock for easy height adjustment, 4 Hooks, 4 Wheels	
83.	IV Starter Kit, includes latex free IV constricting band (18 in. x 1in.), clear adhesive hypoallergenic tape (4 in. x 4.75in.), 2x alcohol pads; 2x 2x gauze pads (2in. X 2in.); Tegaderm Transparent Film (Pav 2 ER use)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
84.	KN95, 5 layer Respirator Mask, single, Individually pack, sterile, earloop type, with High Filtration Efficiency of ≥95%	
85.	Lubricating jelly, sterile, non-toxic, water-based, 5 grams per sachet (with atleast 24 months expiration).	
86.	Lubricating jelly, sterile, non-toxic, water-based, 80 gram/tube (with atleast 24 months expiration).	
87.	Manual Resuscitator, ADULT with oxygen Reservoir bag, 2 Face Masks, 3 Guedel Airway and Oxygen Tubing, (Portable)	
88.	Manual Resuscitator, PEDIA, with oxygen Reservoir bag, 2 Face Masks, 3 Guedel Airway and Oxygen Tubing (Portable)	
89.	Mask, surgical disposable, 3 PLY, Elastic without glass fibres, hypoallergenic, very low resistance to breathing, nose baradaptable, high filtration capacity, perfect fitting, EAR-LOOP	
90.	Mayo Stand with Trays (4 wheels)	
91.	Medical intubation stylet guide wire for endotracheal tube, made of malleable aluminium, smooth PVC sleeve, smooth formed tip, size Fr.12 (with atleast 24 months expiration).	
92.	Medical intubation stylet guide wire for endotracheal tube, made of malleable aluminium, smooth PVC sleeve, smooth formed tip, size Fr.10 (with atleast 24 months expiration).	
93.	Medicine Cups, 30mL - Medicine Cups Disposable 1 oz. Graduated. 1 oz. = 30m, Plastic.	
94.	Medicine Tablet Cutter / Pill Cutter Stainless - ACCURATE BLADES: Our blades are angled so that they will never dull, yet they are not sharp to the touch. TWIN BLADES: The double blades create a split point that gives you perfect halves every time!. COMPREHENSIVE DESIGN: This perfect pill cutter splits any size and shape tablet.. QUALITY CRAFTSMANSHIP: Made of solid aircraft aluminum. Crafted here in the USA.. GUARANTEE: We offer a Lifetime guarantee on the Equadose pill splitter!	
95.	Medicine Tray (Stainless) - Size:9×4.6" Stainless steel medicine trays are often used in conjunction with hospital bedside tables or overbed tables to hold medicines, and other items. The ease of sterilization for these steel surgical trays makes them easy to reuse over and over for patients for years. Used to store or sterilize medical, dental, surgical instruments and medicines.	
96.	Mucus Specimen Traps - use to collect specimen by using any suction pump device. The container has clearly marked measurement gradations that show collective amounts trapped within the 40cc to 80cc plastic container. It features a disposable adapter that prevents contamination and the screw on cap helps maintain a sterile pouring surface. It comes with a 5 inch, non-latex, heavy wall tubing. The tubing will accomodate all suction catheters to become a universal trap. (with at least 24 months expiration).	
97.	N95 mask, particulate respirator contains no components made from natural rubber latex, Collapse resistant cup shape design. Two-strap design with welded dual point attachment, braided headbands, cushioning nose foam, and lightweight construction	

QA

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	for comfortable wear, easy breathing/ speaking. Made from synthetic plastic fibers, usually polypropylene (PP). They also contain rubber and metal for the straps, staples, and nose bar. Teal color, with NIOSH Certificate (with at least 24 months expiration)	
98.	Nasal Oxygen Cannula, ADULT, made from nontoxic PVC compound, 1.8 meters of non-kink star lumen PVC tubing, nasal prongs made of soft PVC (with at least 24 months expiration)	
99.	Nasal Oxygen Cannula, PEDIA - Material: DEHP free, Latex-free. Connections: 6mm tubing for gas supply. Dimension: 18cm x 3cm x 23cm. Allergy-free silicone prong minimizes skin irritation. Ultra soft and lightweight nasal prong for better comfort. Curved nasal prong for better stabilities and flow path. Skin-friendly ear padding prevents pressure sores around ears. Magic fitting nasal cannula for easy cleaning. Flow indicator detects and indicates whether air is passing through or occlusions on tubing. (with at least 24 months expiration)	
100.	Nasogastric tube (NGT) Duodenal tube, Fr. 16, 100% Silicone, individually pack, sterilized (with atleast 24 months expiration).	
101.	Nasogastric tube (NGT) Duodenal tube, Fr. 14, 100% Silicone, individually pack, sterilized (with atleast 24 months expiration).	
102.	Nasogastric tube (NGT) Duodenal tube, Fr. 5, individually pack, sterilized by EO gas, pyrogen free (with atleast 24 months expiration).	
103.	Nasogastric tube (NGT) Duodenal tube, Fr. 8, individually pack, sterilized by EO gas, pyrogen free (with atleast 24 months expiration).	
104.	Nebulizing kit with mask and mouthpiece, inhalation adult set, includes 1 mask, 1 medication chamber cup, 1 oxygen tube (with at least 24 months expiration).	
105.	Nebulizing kit with Mask (PEDIA) Includes Neb kit, Mask and tubing with medicine chamber (with atleast 24 months expiration).	
106.	Needle, Butterfly Scalp Vein Set, gauge 21	
107.	Needle, Butterfly Scalp Vein Set, gauge 23	
108.	Needle, Butterfly Scalp Vein Set, gauge 25	
109.	NEEDLE G.18 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra thin wall, Tamper Evident peel packs, sterilized by electron beam. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).	
110.	NEEDLE G.20 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra thin wall, Tamper Evident peel packs, sterilized by electron beam. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).	
111.	NEEDLE G.22 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra thin wall, Tamper Evident peel packs, sterilized by electron beam. High	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).	
112.	Non-Rebreathing Oxygen Mask with safety Valve, ADULT (with at least 24 months expiration).	
113.	Oxygen Mask ADULT , with 7' crush-resistant tubing. Barbed fittings help prevent patient from disconnecting the mask. Comes with adjustable aluminum nose clips secure it for a better fit. The turn-up rim can ensure comfortable fit with good seal. Offered with head strap and adjustable nose clip. The standard length of tube is 2.1m (with at least 24 months expiration).	
114.	Oxygen Mask, under the chin style, medium concentration mask with 7ft. (84") tubing for single patient. (PEDIATRIC) (with atleast 24 months expiration).	
115.	Peak Flow Meter	
116.	Penlight - with Pupil Gauge, Chrome Plated finish, LED lamp. Chrome Plated Diagnostic Penlight with Pupil Gauge. An innovative extra bright cool light diagnostic penlight with chrome plated brass casing, pocket clip press activator switch, pupil gauge and slide rule.. Comes complete with 2 AAA batteries. Size: 13x2 inches. the cartridge is made of aluminum alloy with anodized treatment, light and brightness. – Pocket clip press activator switch – LED lamp is durable and power saving – 2 "AAA" batteries	
117.	Povidone Iodine, 10% solution antiseptic and disinfectant (with at least 24 months expiration).	
118.	Pulse Oximeter, fingertip, SPO2 and pulse rate (PR) display, PR waveform and bargraph display, built in alarm can be set to high and low, adjustable auto alarm range, simple operation, low power consumption, four directions and eight display modes, color OLED, automatic power off, with USB rechargeable cord	
119.	Sanitary Napkin	
120.	Sharp Disposal Container, double lid, with hermetic lock seal, 30 Liters	
121.	Skin Stapler wire with free pieces skin remover each, taller and sharper, pre-cock function, 35 staple count, sterile (with atleast 24 months expiration).	
122.	Sterile Gauze Pads, 4x4 individually packed	
123.	Straight Catheter, Fr.8, transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic (with at least 24 months expiration).	
124.	Straight Catheter, Fr.12, transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic, Fr.12 (with at least 24 months expiration).	
125.	Straight Catheter, Fr.14, transparent, individually pack, sterilized by EO gas, non-toxic, non-pyrogenic (with at least 24 months expiration).	
126.	Suction catheter, Fr. 5 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
127.	Suction catheter, Fr. 8 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
128.	Suction catheter, Fr. 12 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
129.	Suction catheter, Fr. 14 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
130.	Suction catheter, Fr. 16 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
131.	Suction catheter, Fr. 18 with control & cap, transparent, individually pack, sterilized by EO gas, non- toxic, non-heavy metal, non-pyrogenic (with at least 24 months expiration).	
132.	Surgical Micropore tape, 1 inch, reliably secures dressings and devices to skin, hypoallergenic, latex-free. Reliably secures dressings and devices to skin at least 2 hours, to prevent leakage of blood from AV fistula, IV Cannula, catheters, etc. with at least 24 months expiration	
133.	Surgical Micropore tape, 2 inches, reliably secures dressings and devices to skin, hypoallergenic, latex free, Reliably secures dressings and devices to skin at least 2 hours, to prevent leakage of blood from AV fistula, IV Cannula, catheters, etc. with at least 24 months expiration	
134.	Surgical Micropore tape, 3 inches, reliably secures dressings and devices to skin, hypoallergenic, latex-free, Reliably secures dressings and devices to skin at least 2 hours, to prevent leakage of blood from AV fistula, IV Cannula, catheters, etc. with at least 24 months expiration	
135.	Surgical Shoe Cover, disposable, non-woven fabric, anti-water, anti-blood, anti-alcohol, anti-static.	
136.	Syringe, Asepto, Irrigation Bulb 60cc, sterile, catheter tip, with plastic tip protector, allows one hand control, strong uniformly shaped barrel. (with atleast 24 months expiration).	
137.	Syringes, disposable, 10cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).	
138.	Syringes, disposable, 1cc, with needle G.26 x1/2 inch, slip tip, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with atleast 24 months expiration).	
139.	Syringes, disposable, 30cc, with G.18 needle, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
140.	Syringes, disposable, 3cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy.(with atleast 24 months expiration).	
141.	Syringes, disposable, 5cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy.(with atleast 24 months expiration).	
142.	Syringes, disposable, 5cc, with needle G.23x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).	
143.	Syringes, disposable, INSULIN (1cc), with needle G.30, slip tip, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy.(with atleast 24 months expiration).	
144.	Thermal Paper 80mm - Electrocardiograph (ECG) Thermal Paper, 80x20mm – 3.5 inches x 65.6 feet	
145.	Tongue depressor, individually packed, wooden (with at least 24 months expiration).	
146.	Tourniquet, disposable, elastic, textured finish/smooth, retractable, highly flexible, latex free cotton ribbon, snap-on safety closure, secure opening, quick and safe releasing, convenient roll clip	
147.	Underpads Super Dry, 60x90cm, disposable. (with atleast 24 months expiration).	
148.	Urinal, plastic, for male, reusable, translucent, with lid, permits measuring ease and visualization contents	
149.	Urine bag, with dual hanger, 2L capacity, 120 cm tube length. (with atleast 24 months expiration).	
150.	Volumetric Solution Infusion Set (Soluset) attached injection set on 100 mL burette made of PVC, non-toxic resin, sterilized by EO gas, non-heavy metal & pyrogen free w/ built in float valve interrupts the fluid path, with y-injection port, 60 microdrops/ml (with atleast 24 months expiration).	
151.	Wound Dressing Tray, 8x6 inches thin, stainless steel with instrument tray for cylinder tube and medicine cups	
152.	Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 1 inch x 10 yards (with at least 24 months expiration).	
153.	Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 2 inches x 10 yards (with at least 24 months expiration).	
154.	Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 3 inches x 10 yards (with at least 24 months expiration).	
155.	Arterial Blood Gas Calibrant (EDAN 15) (100pcs/pack)	
156.	Arterial Blood Gas Test Cartridge BG3 (EDAN i15) (25pcs/pack)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
157.	Transpore Plaster 1 inch	
158.	Pre-heparinized Arterial Blood Gas 3cc syringe	
159.	Ventilator Breathing Circuit - 2 limbs (with extra 150cc corrugated tube for the ventilator and humidifier connection), 2 H2O traps, non-heated, with nebulizer kit, disposable (single patient only) with at least 24 months expiration	
160.	BGE LEVEL 1 (5 AMPULES / SET)	
161.	BGE LEVEL 1 (5 AMPULES / SET)	
162.	BGE LEVEL 3 (5 AMPULES / SET)	
163.	Armored / Reinforced Endotracheal Tube (ETT), size 6.0 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
164.	Armored / Reinforced Endotracheal Tube (ETT), size 6.5 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
165.	Armored / Reinforced Endotracheal Tube (ETT), size 7.0 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
166.	Armored / Reinforced Endotracheal Tube (ETT), size 7.5 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
167.	Armored / Reinforced Endotracheal Tube (ETT), size 8.0 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
168.	Armored / Reinforced Endotracheal Tube (ETT), size 8.5 - has reinforced metal coils which make it less likely to get obstructed. (with at least 24 months expiration).	
169.	Anesthesia Set Flex, latex free Adult, Breathing Circuit with Fitting Mask - disposable anesthesia set, consisting of 2 hoses, Y-piece, Luer lock elbow and extendable bag hose, connector, fastening ring, 3L latex free breathing bag.	
170.	Anesthesia Set Flex, latex free - Pedia, Breathing Circuit with Fitting Mask - disposable pediatric anesthesia set, consisting of 2 hoses (diameter 15mm), Y-piece, Luer lock elbow and extendable bag hose (diameter 15mm), connector, fastening ring, 1L latex free breathing bag.	
171.	Anesthesia tubing - Jackson -Reese (with atleast 24 months expiration).	
172.	Antibiotic Mesh 10cm x 10cm (with atleast 24 months expiration).	
173.	Antiseptic surgical scrub solution. Non-ionic Surfactants, Has a pleasant perfume, and Tartrazine as colours. Easy to wash off. Can be used for neonates and thyroid patients. Not degraded due to bioburden. Effective in presence of blood, serious fluid, sweat etc. Non-staining and skin friendly. (5liters/gallon) (with atleast 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
174.	Catgut Chromic, Chromic natural absorbable, size 0, 75 cm, reverse cutting, 1/2 circle, 37mm (with atleast 24 months expiration).	
175.	Catgut Chromic, Chromic natural absorbable, size 0, 75 cm, reverse cutting, 3/8 circle, 37mm (with atleast 24 months expiration).	
176.	Catgut Chromic, Chromic natural absorbable, size 1-0, 75 cm, reverse cutting, 1/2 circle, 37mm (with atleast 24 months expiration).	
177.	Catgut Chromic, Chromic natural absorbable, size 1-0, 75 cm, reverse cutting, 3/8 circle, 37mm (with atleast 24 months expiration).	
178.	Catgut Chromic, Chromic natural absorbable, size 2/0, 75 cm, reverse cutting, 1/2 circle, 37mm (with atleast 24 months expiration).	
179.	Catgut Chromic, Chromic natural absorbable, size 2/0, 75 cm, reverse cutting, 3/8 circle, 37mm (with atleast 24 months expiration).	
180.	Catgut Chromic, Chromic natural absorbable, size 3/0, 75 cm, reverse cutting, 1/2 circle, 19mm-37mm (with at least 24 months expiration).	
181.	Catgut Chromic, Chromic natural absorbable, size 3/0, 75 cm, reverse cutting, 3/8 circle, 19mm-37mm (with at least 24 months expiration).	
182.	Catgut Chromic, Chromic natural absorbable, size 4/0, 75 cm, reverse cutting, 1/2 circle, 37mm (with atleast 24 months expiration)	
183.	Catgut Chromic, Chromic natural absorbable, size 4/0, 75 cm, reverse cutting, 3/8 circle, 37mm (with atleast 24 months expiration)	
184.	Catgut Chromic, Chromic natural absorbable, size 0, 75 cm, round taper point, 1/2 circle, 37mm (with atleast 24 months expiration).	
185.	Catgut Chromic, Chromic natural absorbable, size 1, 75 cm, round taper point, 1/2 circle, 37mm (with atleast 24 months expiration)	
186.	Catgut Chromic, Chromic natural absorbable, size 2/0, 75 cm, round taper point, 1/2 circle, 37mm (with atleast 24 months expiration).	
187.	Catgut Chromic, Chromic natural absorbable, size 3/0, 75 cm, round taper point, 1/2 circle, 26mm-37mm (with at least 24 months expiration).	
188.	Catgut Chromic, Chromic natural absorbable, size 4/0, 75 cm, round taper point, 1/2 circle, 22mm (with atleast 24 months expiration).	
189.	Cautery Pad, adult disposable, solid gel, foam support 202x101mm, bipolar (REM Control), 300m cable and valleylab connector (with at least 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
190.	Cautery Pencil - Hand controlled with ultra clean blade electrode, safety holster, Teflon coated tip. (with at least 24 months expiration).	
191.	Cautery tip Long, electrodes 15cm, lancet type (spatula)	
192.	Cautery tip Short, lancet type (spatula)	
193.	Chest tube thoracostomy 1 liter glass bottle with glass tube and rubber cork (CTT Bottle) with 3ft rubber hose tubing, sterile	
194.	Cigarette drain (Penrose drainage tubing) size 1/4" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
195.	Cigarette drain (Penrose drainage tubing) size 1/2" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
196.	Cigarette drain (Penrose drainage tubing) size 1" - Pure rubber latex, X-ray opaque, 18" length (with atleast 24 months expiration).	
197.	CO2 Absorber Lime Soda; Intersorb Plus Sodalime 5L Gallon, white to violet change. (with atleast 24 months expiration).	
198.	Decontamination Mat - Blue Color, 30 sheets per mat, adhesive with indicator - 90 x 115cm, Special antimicrobial and antifungal chemical agent (Benzoisothiazoline)	
199.	Distilled Water, 10 liters / bottle (for Clinical Lab, Dental Clinic and OR use)	
200.	Droplight / Medical Gooseneck Lamp - Adjustable height, Goose neck examination lamp is a multipurpose lamp attached to a flexible adjustable sgaff that can be positioned in various ways. Stainless steel body, high quality	
201.	Elastic Bandage, VELCRO TYPE, permanent, duralastic, (for pressurized dressing) 4" x 10 yards	
202.	Epidural set, Epidural Catheter, G. 18, for continuous epidural anesthesia with Tuohy type needle, catheter (closed tip, 3 lateral eyes) with connector, epidural flat filter and LOR syringe, filter 0.2 micron, Loss-of-Resistance syringe - Luer Slip (with at least 24 months expiration).	
203.	Full Body Convective Warming Blanket for smiths medical warmer, Adult	
204.	Full Body Convective Warming Blanket for smiths medical warmer, Pediatric	
205.	Guedel oral airway, size 1 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
206.	Guedel oral airway, size 2 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
207.	Guedel oral airway, size 3 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
208.	Guedel oral airway, size 4 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
209.	Guedel oral airway, size 5 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
210.	Guedel oral airway, size 6 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
211.	Guedel oral airway, size 7 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
212.	Guedel oral airway, size 8 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
213.	Guedel oral airway, size 9 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
214.	Guedel oral airway, size 10 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
215.	Guedel oral airway, size 11 - Color-coded for easy identification. Smooth finish and rounded edges. Reinforced bite block helps prevent occlusion. Full range of sizes and convenient packaging options. Emergency kit available with a clear plastic case.	
216.	High Level Disinfectant Solution Ortho-Phthalaldehyde, Ready-to-use for Medical Instruments, 6mins contact time, atleast 28days use-life. Atleast 5L/Gallon Contains: 0.55% Ortho-Phthalaldehyde active ingredients. Effective against wide array of microorganisms. It kills vegetative bacteria, mycobacteria viruses and fungi. (with atleast 24 months expiration).	
217.	Isopropanol, Benzalkonium Chloride 1,000mL - Skin, wound & operative site disinfectant. Skin antisepsis before injections, punctures etc. bactericidal (incl. MRSA), fungicidal, tuberculocidal, mycobactericidal, and virucidal effects against coated viruses (incl. HBV, HCV, HIV), and is effective against Herpes simplex virus type 1 and rotavirus. (with at least 24 months expiration)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
218.	Laparotomy Sponge 8x36x12, 40'sx40's, 28x24 mesh,with blue radio opaque and x-ray detectable strip. (with atleast 24 months expiration).	
219.	Laryngeal Mask Airway, Size 2, double lumen, Silicon, Disposable, LMA	
220.	Laryngeal Mask Airway, Size 3, double lumen, Silicon, Disposable, LMA	
221.	Laryngeal Mask Airway, Size 4, double lumen, Silicon, Disposable, LMA	
222.	Laryngeal Mask Airway, Size 5, double lumen, Silicon, Disposable, LMA	
223.	Liquid Hand Rub Antiseptic with 70% Ethanol and 0.5% Chlorhexidine Gluconate skin disinfectant, Bactericidal, Fungicidal and Virucidal, 5Liters/containers.(with atleast 24 months expiration).	
224.	Nylon cutting with needle size 4-0, 26mm, 1/2c, 75cm.	
225.	Nylon cutting with needle size 5-0, 26mm, 1/2c, 75cm.	
226.	Nylon Cutting 5-0, P3, 26mm, 1/2c, 75cm.	
227.	Nylon Cutting 8-0, 6.5mm, 1/4c, 30cm, spatula	
228.	Orthoplast Skin Traction Kit - Content: Fabric Strapping coated with hypoallergenic hot melt adhesive. Foam padding attached to the middle of the strap for cushioning effect at the malleolar region. Plastic spacer block (suitably drilled for cord attachment). Elastic non adhesive bandage.	
229.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, 1/2 circle , 37mm / 40mm(with atleast 24 months expiration).	
230.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, / 3/8 circle, 37mm / 40mm(with atleast 24 months expiration).	
231.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
232.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with atleast 24 months expiration).	
233.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
234.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with atleast 24 months expiration).	
235.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with atleast 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
236.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm (with atleast 24 months expiration).	
237.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm / 40mm (with atleast 24 months expiration).	
238.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm / 40mm (with atleast 24 months expiration).	
239.	PGA (Polyglycolic Acid) 5-0 P3 Needle cutting (with atleast 24 months expiration).	
240.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
241.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 18 to 24 months expiration).	
242.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
243.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
244.	PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with atleast 24 months expiration).	
245.	PGA (Polyglycolic Acid) 8-0 Round (with atleast 24 months expiration).	
246.	Pneumatic Tourniquet (with cuff for upper & lower extremities	
247.	Polypropylene knitted non-absorbable Mesh, undyed, 7.6cm x 15cm (with atleast 24 months expiration).	
248.	Polypropylene monofilament, non absorbable, size 0, 75 cm, cutting needle, 1/2 circle, 37mm (with atleast 24 months expiration).	
249.	Polypropylene monofilament, non-absorbable, size 1, 75 cm, cutting needle, 1/2 circle, 37mm – 40mm (with at least 24 months expiration).	
250.	Polypropylene monofilament, non absorbable, size 2/0, 75 cm, cutting needle, 1/2 circle / 3/8 circle, 26mm - 37mm (with at least 24 months expiration).	
251.	Polypropylene monofilament, non absorbable, size 3/0, 75 cm, cutting needle, 1/2 circle, 24mm / 37mm (with atleast 24 months expiration).	
252.	Polypropylene monofilament, non absorbable, size 3/0, 75 cm, cutting needle, 3/8 circle, 24mm / 37mm (with atleast 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
253.	Polypropylene monofilament, non absorbable, size 4/0, 45cm-75 cm, cutting needle, 1/2 circle, 19mm - 24mm (with at least 24 months expiration).	
254.	Polypropylene monofilament, non absorbable, size 4/0, 45cm-75 cm, cutting needle, 3/8 circle, 19mm - 24mm (with at least 24 months expiration).	
255.	Polypropylene monofilament, non absorbable, size 0, 75 cm, round needle, 1/2 circle, 37mm (with atleast 24 months expiration).	
256.	Polypropylene monofilament, non absorbable, size 1, 75 cm, round needle, 1/2 circle, 37mm (with atleast 24 months expiration).	
257.	Polypropylene monofilament, non absorbable, size 2/0, 75 cm, round needle / taper needle, 1/2 circle, 26mm-37mm (with at least 24 months expiration).	
258.	Polypropylene monofilament, non absorbable, size 3/0, 75cm-90cm, round needle, 1/2 circle, 37mm / 26mm (with atleast 24 months expiration).	
259.	Polypropylene monofilament, non absorbable, size 4/0, 75cm-90cm, round needle, 1/2 circle, 17mm(with atleast 24 months expiration).	
260.	Polypropylene monofilament, round 5-0, 17mm, 1/2c, 90cm	
261.	Polypropylene Suture with round needle 6-0, 24" (60cm), 9mm – 10mm, 3/8c (with at least 24 months expiration).	
262.	Propan 2-ol, propan-1-ol, benzyl - C12-16- alkydimethylammonium chloride 1,000ml - Surgical and Hygienic Hand Disinfection Solution of medical grade with spray (with at least 24 months expiration)	
263.	Rust Remover - Acidic cleaning concentrate and troubleshooting aid, 1 liter - can be used to remove silicate deposits in washer, removal of annealing colours, water spots, transferred/initial rust, and encrustations, based on acids and surfactants and effectively removes inorganic contamination, exclusively used with stainless steel instruments, and alkali-free. (with at least 24 months of expiration)	
264.	Scalpel Blade #10 Disposable Disposable Dermaplaning Blades with Plastic Handle, Scalpel Blades, High Carbon Steel Dermablade Blades Individually Wrapped Sterile (with at least 24 months of expiration)	
265.	Scalpel Blade #11 Disposable Disposable Dermaplaning Blades with Plastic Handle, Scalpel Blades, High Carbon Steel Dermablade Blades Individually Wrapped Sterile (with at least 24 months of expiration)	
266.	Scalpel Blade #15 Disposable Disposable Dermaplaning Blades with Plastic Handle, Scalpel Blades, High Carbon Steel Dermablade Blades Individually Wrapped Sterile (with at least 24 months of expiration)	
267.	Scalpel Blade #20 Disposable Disposable Dermaplaning Blades with Plastic Handle, Scalpel Blades, High Carbon Steel	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	Dermablade Blades Individually Wrapped Sterile (with at least 24 months of expiration)	
268.	Silk braided, non asorbable, size 0, 60cm, 12-15 threads, no needles (with atleast 24 months expiration).	
269.	Silk braided, non asorbable, size 1, 60cm-75cm, 12-15 threads, no needles (with atleast 24 months expiration).	
270.	Silk braided, non asorbable, size 2/0, 60cm, 12-15 threads, no needles (with atleast 24 months expiration).	
271.	Silk braided, non asorbable, size 3/0, 60cm, 12-15 threads, no needles (with atleast 24 months expiration).	
272.	Silk braided, non asorbable, size 4/0, 60cm, 12-15 threads, no needles (with atleast 24 months expiration).	
273.	Silk braided, non asorbable, size 0, 75cm, cutting needle, 1/2 circle, 37mm (with atleast 24 months expiration).	
274.	Silk braided, non asorbable, size 0, 75cm, cutting needle, 3/8 circle, 37mm (with atleast 24 months expiration).	
275.	Silk braided, non asorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm (with atleast 24 months expiration).	
276.	Silk braided, non asorbable, size 1, 75cm, cutting needle, 3/8 circle, 37mm (with atleast 24 months expiration).	
277.	Silk braided, non asorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with atleast 24 months expiration).	
278.	Silk braided, non asorbable, size 2/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm (with atleast 24 months expiration).	
279.	Silk braided, non absorbable, size 3/0, 45cm - 75cm, cutting needle, 1/2 circle, 26mm (with at least 24 months expiration). (For OR and Dental Clinic use)	
280.	Silk braided, non absorbable, size 3/0, 45cm - 75cm, cutting needle, 3/8 circle, 26mm (with at least 24 months expiration). (For OR and Dental Clinic use)	
281.	Silk braided, non absorbable, size 4/0, 45cm - 75cm, cutting needle, 1/2 circle, 19mm - 26mm (with at least 24 months expiration).	
282.	Silk braided, non absorbable, size 4/0, 45cm - 75cm, cutting needle, 3/8 circle, 19mm - 26mm (with at least 24 months expiration).	
283.	Silk braided, non asorbable, size 0, 75cm, round needle, 1/2 circle, 37mm(with atleast 24 months expiration).	
284.	Silk braided, non asorbable, size 1, 75cm, round needle, 1/2 circle, 37mm(with atleast 24 months expiration).	
285.	Silk braided, non asorbable, size 2/0, 75cm, round needle, 1/2 circle, 26mm / 37mm(with atleast 24 months expiration).	
286.	Silk braided, non absorbable, size 3/0, 75cm, round needle, 1/2 circle, 25mm- 37mm (with at least 24 months expiration).	
287.	Silk braided, non asorbable, size 4/0, 75cm, round needle, 1/2 circle, 22mm / 26mm (with atleast 24 months expiration).	
288.	Spinal Anesthesia Needle, sterile, G. 20 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
289.	Spinal Anesthesia Needle, sterile, G. 21 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
290.	Spinal Anesthesia Needle, sterile, G. 22 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
291.	Spinal Anesthesia Needle, sterile, G. 23 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
292.	Spinal Anesthesia Needle, sterile, G. 25 x 3 1/2", must be sharp and sturdy (with atleast 24 months expiration).	
293.	Spray foam - Ready-to-use enzymatic pre-cleaning with disinfection foam based with active ingredients such as protease and surfactants for keeping surgical instruments moist, active against enveloped viruses, bactericidal, levurocidal, has N,N-Didecyl-N-methyl-poly (oxyethyl) ammonium propionate and N-(3-Aminopropyl)-N-dodecylpropane-1, 3-diamine, 0.5L	
294.	Steri strip, 12mm x 100mm	
295.	Steri strip, 6mm x 100mm	
296.	Sterile water, 100ml (with atleast 24 months expiration).	
297.	Suction Poole Drain, non-toxic, non-pyrogenic, sterilized by EO gas, with 210cm universal connecting tube, with removable suction tip cover (with atleast 24 months expiration).	
298.	Surgical Blade #10. Superior cutting quality, strength and corrosion resistant characterize these stainless steel, disposable blades. Individually wrapped and sealed in aluminum foil to maintain sterility (gamma irradiation), ensuring long shelf-life. (with at least 24 months expiration).	
299.	Surgical Blade #11. Superior cutting quality, strength and corrosion resistant characterize these stainless steel, disposable blades. Individually wrapped and sealed in aluminum foil to maintain sterility (gamma irradiation), ensuring long shelf-life. (with at least 24 months expiration).	
300.	Surgical Blade #15. Superior cutting quality, strength and corrosion resistant characterize these stainless steel, disposable blades. Individually wrapped and sealed in aluminum foil to maintain sterility (gamma irradiation), ensuring long shelf-life. (with at least 24 months expiration).	
301.	Surgical Blade #20. Superior cutting quality, strength and corrosion resistant characterize these stainless steel, disposable blades. Individually wrapped and sealed in aluminum foil to maintain sterility (gamma irradiation), ensuring long shelf-life. (with at least 24 months expiration).	
302.	Surgical Marking pen with ruler, sterilized, individually packed. (with atleast 24 months expiration).	
303.	Surgical Scrub with Chlorhexidine and brushes(with atleast 24 months expiration).	
304.	T-tube, Fr.12, single use, individually packed, smooth surface, non-toxic latex, with radiopaque line. (with at least 24 months expiration)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
305.	T-tube, Fr.14, single use, individually packed, smooth surface, non-toxic latex, with radiopaque line. (with at least 24 months expiration)	
306.	T-tube, Fr.16, single use, individually packed, smooth surface, non-toxic latex, with radiopaque line. (with at least 24 months expiration)	
307.	T-tube, Fr.18, single use, individually packed, smooth surface, non-toxic latex, with radiopaque line. (with at least 24 months expiration)	
308.	Transparent Dressing with absorbent pad, size: D: 4 x 10 (10cm x 25cm) P: 2 x 8 (5cm x 20cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
309.	Transparent Dressing with absorbent pad, size: D: 4 x 6 (10cm x 15cm) P: 2 x 4 3/8 (5cm x 11cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
310.	Transparent Dressing with absorbent pad, size: D: 4 x 8 (10cm x 20cm) P: 2 x 6 5/16 (5cm x 16cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adheretn and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with atleast 18 to 24 months expiration)	
311.	Transparent film Dressing frame style, without gauze size: 4 in x 4 3/4 in (10cm x 12cm) waterproof and impermeable to bacteria and virus(with atleast 24 months expiration)	
312.	Multifix Tape - Universal securement device for effective fixation to the skin of range of medical tubing, with autraumatic pure soft silicone -re-adhesive gel strip. Gentle on skin, no skin prep required before application, leaves no sticky residue on skin upon removal, can be easily repositioned without the risk of skin and hair pulling, showerproof, breathable, hypoallergenic. LARGE	
313.	Wound Drainage System - Silicone 100-150mL reservoir capacity, with clear calibration and built-in clip hanger, NRV (Non return valve) system, 720mm Silicone connecting tube with radio opaque line, 110mmHg Negative pressure, with or without Trocar needle. (with at least24 months expiration).	
314.	Extension Tube with Tri-Lumen NF (Needle Free), with luer access valve, closed system, color coded slide clamp. Kink resistant tubing. (with at least 24 months expiration)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
315.	Steam Indicator Strips (200's/box) (with at least 24 months expiration)	
316.	Velocity Biological Indicator 30 CT for Plasma Sterilizer must be recommended by manufacturing brand (Sterrad)	
317.	Sealsure CI tape case, box of 6's for Plasma Sterilizer, must be recommended by manufacturing brand (Sterrad)	
318.	Chemical Indicator strip, box of 1000's, must be recommended by manufacturing brand (Sterrad)	
319.	Cassette 100 NX (2's), must be recommended by manufacturing brand (Sterrad)	
320.	Flexible adult Tracheostomy Tube size 6.0, Fenestrated (with at least 24 months expiration).	
321.	Flexible adult Tracheostomy Tube size 6.5, Fenestrated (with at least 24 months expiration).	
322.	Flexible adult Tracheostomy Tube size 7.0, Fenestrated (with at least 24 months expiration).	
323.	Flexible adult Tracheostomy Tube size 7.5, Fenestrated (with at least 24 months expiration).	
324.	Flexible adult Tracheostomy Tube size 8.0, Fenestrated (with at least 24 months expiration).	
325.	Flexible adult Tracheostomy Tube size 8.5, Fenestrated (with at least 24 months expiration).	
326.	Flexible adult Tracheostomy Tube size 6.0, with Disposable or Reusable Inner Cannula (with at least 24 months expiration).	
327.	Flexible adult Tracheostomy Tube size 6.5, with Disposable or Reusable Inner Cannula (with at least 24 months expiration).	
328.	Flexible adult Tracheostomy Tube size 7.0, with Disposable or Reusable Inner Cannula (with at least 24 months expiration).	
329.	Flexible adult Tracheostomy Tube size 8.5, with Disposable or Reusable Inner Cannula (with at least 24 months expiration).	
330.	Tracheostomy Mask, Adult - The neckband is made from a comfortable, non biting material, and the swivel tubing connector permits access from either side of the patient. The special strap clips allow the mask to be removed with minimal disturbance to the patient.	
331.	Absorbable Hemostat, SURGICEL - hemostatic agent (blood-clot-inducing material) made of an oxidized cellulose polymer (the unit is polyanhydroglucuronic acid) 3x4inches	
332.	Surgical Blade #12 (for Dental And ENT Service use) Superior cutting quality, strength and corrosion resistant characterize these stainless steel, disposable blades. Individually wrapped and sealed in aluminum foil to maintain sterility (gamma irradiation), ensuring long shelf-life. (with at least 24 months expiration).	
333.	Surgical Nasal Sponge Packing - Low density, Large Pore merocel microdispersed, oxidized cellulose standard nasal dressing with string, length 8cm, Thickness 1.5cm, Height 2cm	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
334.	Surgical Nasal Sponge Packing - Low density, Large Pore merocel microdispersed, oxidized cellulose standard nasal dressing with string, length 4.5cm, Thickness 1.5cm, Height 2cm	
335.	Nasal Splint (External) -Soft-Form Nasal Splint, Regular, Kit	
336.	Anti Embolic Stockings Thigh High (Small) - For prevention of venous disease. Used for treatment after vein operation. Comfortable open-toe stockings	
337.	Anti Embolic Stockings Thigh High (Medium) - For prevention of venous disease. Used for treatment after vein operation. Comfortable open-toe stockings	
338.	Anti Embolic Stockings Thigh High (Large) - For prevention of venous disease. Used for treatment after vein operation. Comfortable open-toe stockings	
339.	Polyester Suture, 2-0 - 30"(75cm) 17mm, 1/2c, Tapercut,	
340.	Polyester Suture, 5-0 - 30"(75cm) 17mm, 1/2c, Tapercut,	
341.	Ortho Gloves size 7.0, sterile, powder free, thick and elastic, not easily tore. (with atleast 24 months expiration).	
342.	Ortho Gloves size 7.5, sterile, powder free, thick and elastic, not easily tore. (with atleast 24 months expiration).	
343.	Ortho Gloves size 8.0, sterile, powder free, thick and elastic, not easily tore. (with atleast 24 months expiration).	
344.	Balance salt solution 500ml	
345.	Carbachol Ophthalmic Solution, 1mL	
346.	Corneal knife 2.75mm	
347.	Corneal Stab knife 15 degree	
348.	Crescent bevel up knife 2.3mm - 2.5mm	
349.	Eye Pad Sterile, Constructed of absorbent cotton filler. Covered with soft non-woven fabric. Provides protection to the eye when required. Oval shape. Sterile packaging. 4 x 6 cm	
350.	Eye drape, sterile, halfbody with pouch	
351.	PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 6/0, 75cm, cutting needle, 0.7 metric, 18", 45cm-76cm, 1/4c, spatulated (with at least 24 months expiration).	
352.	Silk Suture, Reverse Cutting, Size 6-0, Black, 18", Needle HE-7, 3/8 - 1/2 Circle (with atleast 24 months expiration).	
353.	Sodium Hyaluronate at least 1.5% viscoelastic eye gel, 1mL Pre-filled syringe	
354.	Sodium Hyaluronate 1.5% viscoelastic eye gel, 1mL	
355.	Tryphan blue ophthalmic solution, 1mL vial	
356.	Phacoemulsification basic pack (must be compatible with Stellaris Phaco machine)	
357.	Arterio-Venous Fistula (AVF) kit, inclusive of 4pcs sterile cherry balls, 1pc drape, 2pcs sterile gauze, 2pcs antiseptic pads, 3pcs 10cc syringes with needle g.21, and 2pcs sterile gauze pillows, with at least 24 months expiration	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
358.	Sub clavian kit inclusive of 2pairs of Sterile Gloves 6.5, 2 packs of sterile gauze sponges 4x4x8 5's, 1 pc sterile drape, 8 pcs Chlorhexidine pad, 2pcs heplock, with at least 24 months expiration	
359.	Dialysis Needle for Fistula and Graft, G16, arterial, 1.6 x 25mm, with metal stainless needle, with back eye, and sharp cutting edge, providing smooth insertion, fixed wing, with at least 24 months expiration	
360.	Dialysis Needle for Fistula and Graft, G16, venous , 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration	
361.	Dialysis Needle for Fistula and Graft, G17, arterial, 1.6 x 25mm, with metal stainless needle, with back eye, and sharp cutting edge, providing smooth insertion, fixed wing, with at least 24 months expiration	
362.	Dialysis Needle for Fistula and Graft, G17, venous , 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration	
363.	Acid Concentrate Pre-Mix 5 liter per container, with the following content: each liter contains: sodium chloride 210.67 grams - 219.70 grams, potassium chloride 5.219 grams to 5.50 grams, calcium chloride 6.47 grams to 8.10 grams, magnesium chloride 3.55 grams to 7.54 grams, glacial acetic acid 4.70 grams to 8.46 grams, with or without glucose with at least 24 months expiration. must be compatible with the technical requirements of the existing hemodialysis machine (Fresenius) and will not require technical adjustments/ service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine)	
364.	Bicarbonate Concentrate granules, 8.4kg/100L concentrate (8.4%), each 100 ml of this concentrate contain 84 grams sodium hydrogen carbonate, with at least 24 months expiration, will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine)	
365.	Potency Strips, test for potency of peracetic acid in dialyzer reuse, 100 strips per bottle, with shelf life of 18 months,	
366.	Test for residual peroxide in rinse water, 100 strips per bottle, with shelf life of 18 months,	
367.	Bloodlines AV-Set with drain bag and macroset, blood pump segment diameter (ID) 8mm x (OD), 12mm, chamber size 22mm , with at least 24 months expiration, must be compatible with the technical requirements of the existing hemodialysis machine (Fresenius), will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine)	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
368.	Low Flux Dialyzer, surface area 1.8-1.9 m ² , UF coeff 12.3-18 ml/hr/mmHg ² , with at least 24 months expiration. (must be compatible with 4008s Hemodialysis Machine)	
369.	High Flux Dialyzer, effective surface area 1.5 - 1.8 m ² , UF coeff 52-67 ml/hr/mmHg ² . (must be compatible with 4008s Hemodialysis Machine) with at least 24 months expiration	
370.	Thermochemical Disinfectant for heat disinfection of hemodialysis machines, each 100g of disinfectant contains 21g citric acid 1-hydrate, 5 liters per container, with at least 24 months expiration. (must be compatible with 4008s Hemodialysis Machine)	
371.	Dialysis Fluid Filter for existing Fresenius Hemodialysis Machines, effective surface 2.2m ² , weight 170g, polypropylene housing material, polyurethane potting material, silicone sealings, filtration rate of 5ml/min x mmhg, operating time: max 12 weeks or 100 treatments, with at least 24 months expiration, will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine)	
372.	UV sterilizer, with voltage 220 V, applicable area at least 40 m, with remote control, easy to use.	
373.	Industrial Salt, high sodium chloride content, chunky crystals, low impurities, reliable supply, 12.5 kg per pack	
374.	Disinfecting Agent for hemodialysis based on hydrogen peroxide and peroxyacetic acid mixture, stabilized 5.1 and 8 II, 5 kilogram, cold sterilant, At least 5 liters per container, with at least 24 months expiration. (must be compatible with 4008s Hemodialysis Machine)	
375.	Sodium hypochlorite, bleaching agent for disinfection at least 5.2% per 1 gal	
376.	ECG Lead, with 12-lead wire. (must be compatible to Mindray Beneheart R3 ECG Machine) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.	
377.	STAND TYPE BP APPARATUS - ADULT. Height - 25 cm, Weight - 400 gm, Breadth - 25 cm, Length - 25 cm. Generously dimensioned scale (Ø 147.2 mm) and high-contrast scale layout for optimum readability; scale up to 300 mm Hg. Spacious basket on the reverse side for tubes and cuffs. Latex bulb to inflate the cuff. BP Velcro cuff, LATEX RUBBER TYPE OF BLADDER. Non-wearing air-release valve with fine adjustment. Micro-filter to protect the release valve and measurement system. Specially hardened copper-beryllium membrane, which is resistant to aging, can withstand pressures of up to maximum 600 mm Hg Precision mechanism mounted on a special support, in order to ensure complete stability of the measuring system and the zero point. No zero-point fixation required. Maximum error tolerance: ± 3 mm Hg. Provide Calibration Certificate each unit and Warranty Certificate 1 (ONE) year for parts and services	

9A

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration for end-user and maintenance personnel before handover of the unit.	
378.	CPR Manikin with AED indicator	
379.	ECG Machine, digital and portable 3 channel; 12 lead EKG + electrode pad + lead wires orange; USB PC Software; with thermal printing system. For 80mm x 20m ECG Paper size. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.	
380.	Fetal Doppler, compact design; interchangeable waterproof probe; large backlight LCD display; accurate FHR detection with high-fidelity sound effect; support waterproof probes for both obstetrical and vascular usage; auto, average and manual working modes; various power supply solutions; automatic power-off; support 2MHz, 3MHz, 4MHz, 5MHz, and 8MHz waterproof probes. Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.. Perform installation, demonstration for end-user and maintenance personnel before handover of the unit.	
381.	Laboratory Balance with Enclosure, at least 0.1mg readability; 120-gram capacity; 3-sec stabilization time; 0.1mg repeatability; 90mm pan size; full- color VGA graphic touchscreen with user controlled brightness; AC adpter; RS 232/USB Host and device included; GLP/GMP data output with real-time clock; metal base, ABS top hopusing, stainless steel pan, glass drafshield with two-piece top-mounted side doors and sliding too door; iluminated up-front level indicator; integral weigh below hook; security bracket; calibration lock; full housing in use cover; selectable environmental filters; auto tare; user selectable span calibration points; software lockout and reset menu; user selectable communication settings and data print options; user defineable project and user IDs.	
382.	Mannequin / Manikin adult XC-401	
383.	Medicine Cabinet, 2 doors - Stainless. Medicine storage cabinet. 2 Door Glass Slide with 3 Glass shelves. Metal side. 4 Mounted swivel rubber caster wheels. Door lock provided. W900 x D400 x H1850mm	
384.	Medicine Distribution Trolley, with shelves with 16 drug dispensary bins; with transparent dividers and identification card holder; with lock and key	
385.	Medicine/ Emergency Cart with Cardiac Board Stainless - 3 Drawers, 1 Cabinet, IV Stand, Rubber bumper, Oxygen holder, Cardiac Board holder, Stainless Steel.	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
386.	Nebulizing Machine, with compressor; mouthpiece; T-piece; corrugated tubing; baffle; medication cup; nebulizer air inlet connector; VixOne Reusable nebulizer; VixOne Disposable nebulizer; PediNeb TM pacifier; Adult/Pedia Aerosol mask; compressor filter; filter kit (6 each- caps and filters); tubing connector (pacjage of 6)	
387.	Needle Burner / Shredder and Syringe Destroyer, AC/DC power, 220V; low-pressure melting method	
388.	OB / Delivery Instruments Set with Sterelizing Tray: 4 pcs. Backhaus Towel Clamp 5 1/4" 1 pc. Operating Scissors S/B 5 1/2" straight 1 pc. Braun Episiotomy Scissors 8 1/2" Angled to side 1 pc. Umbilican Scissors 4" 2 pcs. Thumb Forceps 5 1/2" Serrated 2 pcs. Tissue Forceps (1x2T) 5 1/2" 6 pcs. Crile Forceps 5 1/2" straight 1 pc. Mayo-Hegar Needle Holder 6" 1 pc. Mayo- Hegar Needle Holder 7" 5 pcs. Towel Clip, 11cm 1 pc. Sterilizing tray (Perforated)	
389.	Oxygen Tank Cart, 50 Lbs. (Single) - Oxygen tank carriage double handle. Size: Fits 50 lbs Oxygen tank. Locally made. Steel Painted or Stainless steel with rubber handle 4 Wheels.	
390.	Stainless Steel Panel Screen (3 folded parts) - Made of tubular stainless mounted 4pieces rubber and amp; 2 pieces wheel caster. Provided with curtain. Size L:122" x H: 65"	
391.	Stretcher, Foldable, high strength AL-alloy material, Surface made of nylon mesh, Light weight, portable, easy to carry out and safety to use, Folded size: 110 x 18 x10 cm, Laid flat size: 220 x53 x14cm, Load capacity: 160 kg.	
392.	Stretcher with wheels, Made of light weight aluminum alloy, High Position: 195x57x87.5cm, Low Position: 198x57x22cm, Maximum Angle of the Back: 58 degrees, Loading Bearing: 181kg	
393.	Suction Machine Heavy Duty- multiple power supply, oil free lubrication pump; low noise; square negative pressure meter and plastic cover; negative pressure regulating system; power supply; Heavy duty monitor. Large Reservoir Bottle at 1000mL. Easy to use. Minimum Pressure: 400mmHg. Adjustable pressure at least negative 400mmHg. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.	
394.	Walker, Foldable, Anodized aluminum structure ensures maximum safety and strength while remaining lightweight, Two button push release is easy to extend/fold, Slip resistant rubber	

INSTRUCTION: Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a) (ii) and/or GCC Clause 2.1(a)(ii).

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	tips in the back with 5 inch wheels in the front provide maximum comfort while in use, Width 44cm Height adjusts from 32 inch to 39 inch in 1 inch increment to accommodate users with different needs.	
395.	BATHROOM WEIGHING SCALE HEAVY DUTY - made with non-skid vinyl mat features white, baked enamel finish battery-free mechanised operation analogue. Capability: 150kg maximum 500g graduation, sturdy, big metallic non-slip podium, easy-to-read dial, used to monitor body weight. Dimension: 32cm x 17cm x 40cm. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.	
396.	Weighing Scale with Height and Weight Measurement, Heavy-duty solid stable platform. Superior Accuracy. Heavy duty understructure. Die cast beam, and aluminum on black insert. Easily read from either side. Factory assembled. All scales are tested with weights traceable to the National Institute of Standards and Technology (NIST). The die cast beam is dual reading and measures in BOTH pounds and kilograms. Capacity: 400 lb x 4 oz (180 kg x 100 g). Platform Size: 10-1/2 x 14-1/2 inches (267 x 368 mm). Height Rod: Measures 30-78 inches (76-198 cm). Scale height: 59 inches (1.5 m) Cartoon size: Height: 57" x Width: 19 1/2". Length: 5 1/2", 20kg. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and 2 (TWO) years for services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.	
397.	Weighing Scale, Pediatric	
398.	Wheelchair, 4 wheels (2 rear wheels and 2 front swivel castors), Seat and backrest providing sitting support for the wheelchair user, Foot supports Armrests, Rear wheel locks (to park the wheelchair), Folding frame wheelchairs and buggies allow folding through cross-brace or other folding mechanism, Rigid frame wheelchairs and buggies must offer mechanism to fold and/or dismantle the device into smaller separate components	

CONFORMED BY: _____
 Authorized Representative
 (Signature over printed name)