

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

**PUBLIC BIDDING FOR THE SUPPLY AND DELIVERY OF MEDICAL SUPPLIES (ADDITIONAL 2)
CY 2022**

This Supplemental Bid Bulletin No. 1 dated **September 20, 2022**, is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

| ISSUES | CLARIFICATION |
|---|---|
| <p>LIST OF ITEMS:</p> <p>1) Blood Glucose Test Strips with Strip Ejector, for in vitro diagnostic use, no coding, individually foil wrapped strips, small blood sample required at least (0.5µl), 5-second test time, 5-second blood sample reapplication time, top or end fill with visual confirmation, fill trigger ensures test starts only when sufficient blood sample has been applied. Must be compatible to the existing Glucometer. (with at least 24 months expiration). - MS A2-003</p> <p>2) Disposable ECG Electrodes, ADULT, 36x45mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration). -MS A2-006</p> <p>3) Unit of measure PIECE -MS A2-008 -MS A2-009 -MS A2-010 -MS A2-011 - MS A2-018</p> <p>4) Surgical Gown (FREE SIZE), disposable, non-woven fabric, anti-water, anti-blood, anti-alcohol, anti-static, light color only. - MS A2-017</p> <p>5) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 90cm, round needle, 1/2</p> | <p>LIST OF ITEMS:</p> <p>1) Blood Glucose Test Strips with <u>FREE 50 units of GLUCOMETER</u> and with Strip Ejector, for in vitro diagnostic use, no coding, individually foil wrapped strips, small blood sample required at least (0.5µl), 5-second test time, 5-second blood sample reapplication time, top or end fill with visual confirmation, fill trigger ensures test starts only when sufficient blood sample has been applied. Must be compatible to the existing Glucometer. (with at least 24 months expiration). - MS A2-003</p> <p>2) Disposable ECG Electrodes, ADULT, <u>36-37x40-45mm</u> (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration). -MS A2-006</p> <p>3) <u>Unit of measure PAIR</u> - MS A2-008 - MS A2-009 -MS A2-010 -MS A2-011 - MS A2-018</p> <p>4) Surgical Gown (FREE SIZE), <u>60GSM</u>, disposable, non-woven fabric, anti-water, anti-blood, anti-alcohol, anti-static, light color only. - MS A2-017</p> <p>5) <u>PGA (Polyglycolic Acid/Polyglactin)</u> braided, coated, braided synthetic absorbable, size 1, <u>70-90cm</u>,</p> |

| | |
|--|--|
| <p>circle, 37mm - 40mm (with at least 18 to 24 months expiration). -MS A2-026</p> | <p>round needle, 1/2 circle, 37mm - 40mm (with at least 18 to 24 months expiration). -MS A2-026</p> |
| <p>6) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 37mm - 40mm (with at least 24 months expiration). -MS A2-027</p> | <p>6) PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 36mm-45mm (with at least 24 months expiration). -MS A2-027</p> |
| <p>7) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm - 40mm (with at least 24 months expiration). -MS A2-028</p> | <p>7) PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm - 40mm (with at least 24 months expiration). NOTE: STICK TO SPECS MS A2-028</p> |
| <p>8) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm - 40mm (with at least 24 months expiration). - MS A2-029</p> | <p>8) PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm - 40mm (with at least 24 months expiration). NOTE: STICK TO THE SPECIFICATIONS -MS A2-029</p> |
| <p>9) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm / 40mm (with at least 24 months expiration). - MS A2-031</p> | <p>9) PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 20mm-40mm (with at least 24 months expiration). - MS A2-031</p> |
| <p>10) Silk braided, non-absorbable, size 3/0, 75cm, round needle, 1/2 circle, 26mm / 37mm (with at least 24 months expiration). -MS A2-041</p> | <p>10) Silk braided, non-absorbable, size 3/0, 75cm, round needle, 1/2 circle, 25mm- 37mm (with at least 24 months expiration). -MS A2-041</p> |
| <p>11) -MS A2 – 050 -MS A2 – 051 -MS A2 – 052 -MS A2 – 053 -MS A2 – 054</p> | <p>11) -MS A2 – 050 -MS A2 – 051 -MS A2 – 052 -MS A2 – 053 -MS A2 – 054</p> |
| <p>12) Nephrostomy Catheter kit, with safety string lock, Fr.14 -MS A2 – 057</p> | <p>NOTE: STICK TO THE SPECIFICATIONS</p> <p>12) Nephrostomy Catheter kit, with safety string lock, Fr.14, 30cm -MS A2 – 057</p> |

SECTION III: BID DATA SHEET

Additional contract documents relevant to the Project that is required by the Procuring Entity:

- a. Notice of Award or Notice to Proceed or Contract issued by the owners, as attachment for the Statement of the prospective bidder of All its On Going Government and Private Contracts
- b. Notice of Award or Contract issued by the owners, as attachment for the Single Largest Completed Contract (SLCC)
- c. Current and Valid Certificate of Distributorship, if applicable
- d. Notarized Letter of Assurance from the Bidder on the Availability of Stocks
- e. Certificate of Performance [For current supplier, it shall be issued by the Head of the Procuring Entity for the current year. For non-current supplier, certificate issued from other hospitals or agencies are acceptable (at least SATISFACTORY RATING)]

Color code for folders and envelope:

The Eligibility Documents and Technical Proposal combined in one folder and the Financial Proposal in separate Folder (RED FOLDER). Each folder with fastener/shoelace on top, with Table of Contents and Tabs (properly labeled according to the checklist for bidders issued).

NOTE: Color of **Folder** is RED

Color of **Envelope** is RED

SECTION III: BID DATA SHEET

Additional contract documents relevant to the Project that is required by the Procuring Entity:

- a. Notice of Award or Notice to Proceed or Contract issued by the owners, as attachment for the Statement of the prospective bidder of All its On Going Government and Private Contracts
- b. Notice of Award or Contract issued by the owners, as attachment for the Single Largest Completed Contract (SLCC)
- c. Current and Valid Certificate of Distributorship, if applicable
- d. Current and Valid License to Operate from FDA, if applicable**
- e. Current and Valid Certificate of Product Registration (CPR) from FDA, if applicable**
- f. Notarized Letter of Assurance from the Bidder on the Availability of Stocks
- g. Certificate of Performance [For current supplier, it shall be issued by the Head of the Procuring Entity for the current year. For non-current supplier, certificate issued from other hospitals or agencies are acceptable (at least SATISFACTORY RATING)]

Color code for folders and envelope:

The Eligibility Documents and Technical Proposal combined in one folder and the Financial Proposal in separate Folder (**YELLOW FOLDER**). Each folder with fastener/shoelace on top, with Table of Contents and Tabs (properly labeled according to the checklist for bidders issued).

NOTE: Color of **Folder** is **YELLOW**


Color of **Envelope** is **YELLOW**

Other Matters:

- A. *Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top or ring bound instead of a fastener, table of contents, and index tabs in words, not numbers.*
- B. *Folder of Eligibility requirements and technical proposal should be placed in one envelope. And the folder of the Financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as “**Original Bid**”*
- C. Documents should be arranged chronologically according to the checklist issued.
- D. Color code for folders and envelope:
YELLOW – Public Bidding for Supply and Delivery of Medical Supplies (Additional 2) CY 2022

- E. All other provisions on the bidding documents which are not affected shall remain in effect.
- F. The deadline for **Submission and Opening of Bids** is scheduled on **September 27, 2022 (Tuesday), 9:00 AM**, at the **PAG-ASA HALL**, National Center for Mental Health Compound, Mandaluyong City.
- G. Any bid submitted after the deadline for submission shall be declared "**LATE**" and shall NOT be accepted.
- H. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.

For the information and guidance of all concerned.


ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
Chairperson, BAC for Goods