

REPUBLIC OF THE PHILIPPINES Department Of Health



National Center for Mental Health

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF VARIOUS MEDICAL AND NON-MEDICAL EQUIPMENTS CY 2022

This Supplemental Bid Bulletin No. 1 dated October 18, 2022 is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

ISSUES	CLARIFICATION
SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF DENTAL UNIT AND CHAIR FULLY MOTORIZED WITH COMPLETE ACCESSORIES CY 2022	SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF DENTAL UNIT AND CHAIR FULLY MOTORIZED WITH COMPLETE ACCESSORIES CY 2022
FROM	<u>TO</u>
SECTION VI Schedule of Requirements	SECTION VI Schedule of Requirements
Delivered, Weeks/Months 60 to 90 calendar days	Delivered, Weeks/Months 90 to 120 calendar days
SUPPLY, DELIVERY, COMMISSIONING AND TESTING OF ADDITIONAL PC DESKTOP AND PRINTERS CY 2022	SUPPLY, DELIVERY, COMMISSIONING AND TESTING OF ADDITIONAL PC DESKTOP AND PRINTERS CY 2022
3-in-1 Ink Tank Printer	3-in-1 Ink Tank Printer
LIST OF ITEMS: Supplier Requirements: 2. Delivery period of 90 days	LIST OF ITEMS: Supplier Requirements: 2. Delivery period of 90 days – 120 days
All-In-One Desktop Computer	All-In-One Desktop Computer
LIST OF ITEMS: Supplier Requirements: 2. Delivery period of 90 days 8. Wired Keyboard and Mouse	LIST OF ITEMS: Supplier Requirements: 2. Delivery period of 90 – 120 days 8. Item no. 8 and 16 are same.

"There is no Health without Mental Health"







ICT Network Infrastructure on NCMH Mega Modular Hospital and NCMH Triage

BID DATA SHEET ITB Clause No. 21.2

Additional contract documents relevant to the Project that is required by the Procuring Entity:

- a. Notice of Award or Notice to Proceed or Contract issued by the owners, as attachment for the Statement of the prospective bidder of All its On Going Government and Private Contracts
- b. Notice of Award or Contract issued by the owners, as attachment for the Single Largest Completed Contract (SLCC)
- e. Certificate Good Performance [For current service provider, it shall be issued by the Head of Procuring Entity / Medical Center Chief II of NCMH, for the current year. For non-current service provider, certificate issued from other Hospitals or agencies are acceptable (as least SATISFACTORY RATING).

SECTION VI

Schedule of Requirements

Delivered, Weeks/Months

60 to 90 calendar days

SECTION VII

Technical Specifications

VIII. Warranty Period and Services

- i. Period: Three (3) years warranty is required on all delivered goods and shall take effect immediately after final acceptance of the project with NCMH.
- ii. Period: Three (3) year of workmanship on support and cabling and shall take effect immediately after final acceptance of the project with NCMH.

ICT Network Infrastructure on NCMH Mega Modular Hospital and NCMH Triage

BID DATA SHEET

ITB Clause No. 21.2

Additional contract documents relevant to the Project that is required by the Procuring Entity:

- a. and b. Notice of Award or Notice to Proceed or Contract or **Purchase Order** issued by the owners, as attachment for the Statement of the prospective bidder of All its On Going Government and Private Contracts.
- e. Certificate Good Performance [For current service provider, it shall be issued by the Head of Procuring Entity / Medical Center Chief II of NCMH, for the current year. For non-current service provider, certificate issued from other Hospitals or agencies, government or private company and must be similar to the project to be bid are acceptable (as least SATISFACTORY RATING).

SECTION VI

Schedule of Requirements

Delivered, Weeks/Months 150 calendar days

SECTION VII

Technical Specifications

VIII. Warranty Period and Services

i. AS IS

ii. AS IS

SUPPLY, DELIVERY, INSTALLATION, COMMISIONING AND TESTING OF ULTRASOUND MACHINE WITH PORTABLE ULTRASOUND

SECTION VII

Technical Specifications

System Architecture Design

- 5. Height with Monitor at least 1600 mm
- 6. Width at least 500 mm
- 7. Depth at least 700 mm
- 8. Weight at least 50 kg (without peripherals)
- 11. Integrated image storage at least 500 GB

System Standard Features

30. The system shall provide post-processing functions in:

i. 4D

Portable Ultrasound:

- 1. Physical Specification
- a. Console: Laptop Style with Cart

Monitor

- 61. Display: at least 15 inch High Resolution LCD Monitor
- 62. Display Size: 1924 x 768

Standard Requirements

97. The manufacturer must have a local office and service center in the Philippines for at least 10 years. Provide a certification from manufacturer.

SUPPLY, DELIVERY, INSTALLATION, COMMISIONING AND TESTING OF ULTRASOUND MACHINE WITH PORTABLE ULTRASOUND

SECTION VII

Technical Specifications

System Architecture Design

- 5. AS IS
- 6. AS IS
- 7. AS IS
- 8. AS IS
- 11. AS IS

System Standard Features

30. AS IS

Portable Ultrasound:

1. AS IS

Monitor

- 61. AS IS
- 62. Display Size: 1024 x 768

Standard Requirements

97. The manufacturer must have a local office and health care service center in the Philippines for at least 10 years. Provide a certification from manufacturer.

SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF MEDICAL SURGICAL EQUIPMENT

General Technical Requirements:

- 1. Pump capabilities:
- flow range 0.1 to ≥ 999 mL/hr
- 10. "Dose error" reduction system (preferable).
- 12. Clearly visible optical alarms.
- 19. Continuous operation within specification in ambient temperature of at least 30-45°C
- 22. Monitored and displayed parameters (color and graphic preferable)
- 26. Power supply (voltage, frequency and plug vary across the countries)
- Total re-charging time not greater than 6 hours.

Supply, Delivery, Installation, Commissioning, and Testing of Various Laboratory Equipment CY 2022

REFRIGERATED CENTRIFUGE FOR BLOOD BAGS

Specifications:

- 3. Replaceable receptacle/bucket that can also hold conical tubes and test tubes
- 9. Maximum speed: at least 4500 rpm

SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF MEDICAL SURGICAL EQUIPMENT

General Technical Requirements:

- 1. Pump capabilities:
- flow range 0.1 to ≥ 1200 mL/hr
- 10. its equivalent
- 12. its equivalent
- 19. Continuous operation within specification in ambient temperature of at least **5 45°C**
- 22. Monitored and displayed parameters (monochromatic and/or color and graphic preferable)
- 26. Power supply (voltage, frequency and plug vary across the countries)
- Total re-charging time is 6 10 hours.

Supply, Delivery, Installation, Commissioning, and Testing of Various Laboratory Equipment CY 2022

REFRIGERATED CENTRIFUGE FOR BLOOD BAGS

Specifications:

- 3. AS IS. It is INCLUDED
- 9. Maximum speed: at least 4500 rpm or higher

Other Matters:

- A. Additional requirements will be required during the post qualification stage.
- B. Product demonstrations of small equipment will be conducted at National Center for Mental Health and for large equipment an on-site installation demo will be required.
- C. Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with <u>shoelace</u> on top or <u>ring binded</u> instead of fastener, <u>table of contents and index</u> tabs in words, not numbers.

- D. Folder of Eligibility requirements and technical proposal should be placed in one envelope. And folder of financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as "Original Bid".
- E. Documents should be arranged chronologically according to the checklist issued.
- F. Color code for folders and envelopes:
 - GREEN
- G. All other provisions on the bidding documents which are not affected shall remain in effect.
- H. The deadline for Submission and Opening of Bids is scheduled on <u>October 25, 2022, 9:00 AM</u>, <u>Tuesday</u> at the <u>ADMIN. CONFERENCE ROOM</u>, National Center for Mental Health Compound, Mandaluyong City.

For the information and guidance of all concerned.

ALDEN C. CUYOS, MD, MMHoA, FPPA, IFAPA (Life)

Chairperson, Bids & Awards Committee for Hospital Equipment