



# National Center for Mental Health

## SUPPLEMENTAL BID BULLETIN

### ADDENDUM NO. 1

#### PUBLIC BIDDING FOR SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF:

1. MANUAL ROTARY MICROTOME CY2023
2. ULTRASONIC SCALER CY2023
3. PHYSICAL THERAPY EQUIPMENT CY 2023
4. FLOOR MOUNTED FULLY DIGITAL X-RAY MACHINE CY 2023
5. GENERAL HOSPITAL EQUIPMENT CY 2023

This Supplemental Bid Bulletin No. 1 dated, **June 09, 2023** is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

ISSUES	CLARIFICATION
<u>FROM</u>	<u>TO</u>
<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF MANUAL ROTARY MICROTOME</b></p> <p>TECHNICAL SPECIFICATIONS</p> <p>- Manual Rotary Microtome</p>	<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF MANUAL ROTARY MICROTOME</b></p> <p>TECHNICAL SPECIFICATIONS</p> <p>- <u>AS IS</u></p>
<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF PHYSICAL THERAPY EQUIPMENT</b></p> <p>PTE23-01: THERAPEUTIC ULTRASOUND MACHINE</p> <p>TECHNICAL SPECIFICATIONS</p> <p>2. With 1, 5, and 10cm 2 applicator heads</p>	<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF PHYSICAL THERAPY EQUIPMENT</b></p> <p>PTE23-01: THERAPEUTIC ULTRASOUND MACHINE</p> <p>TECHNICAL SPECIFICATIONS</p> <p>2. With 2 applicator heads (<u>1 small head and 1 large head</u>)</p>
<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF GENERAL HOSPITAL EQUIPMENT</b></p> <p>GHE23-02: ENT TREATMENT UNIT WITH HYDRAULIC CHAIR</p>	<p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF GENERAL HOSPITAL EQUIPMENT</b></p> <p>GHE23-02: ENT TREATMENT UNIT WITH HYDRAULIC <u>OR PNEUMATIC CHAIR</u></p>

*"There is no Health without Mental Health"*



<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> <li>Suction Device, vacuum: 650mmHg – 700mmHg</li> <li>22" LED Monitor</li> <li>ENT Examination Chair             <ol style="list-style-type: none"> <li>Hydraulic Auto Controller System</li> </ol> </li> </ul> <p>STANDARD REQUIREMENT:</p> <ol style="list-style-type: none"> <li>Certification from the manufacturer authenticated by the Philippine Consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor</li> <li>Certification that the brand has been in the local market for at least five (5) years with at least 10 installations</li> </ol> <p><b>SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF FLOOR MOUNTED FULLY DIGITAL X-RAY MACHINE</b></p>	<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> <li>Suction Device, vacuum: <b>650mmHg – 750mmHg</b></li> <li><b>22" – 25"</b> LED Monitor</li> <li>ENT Examination Chair             <ol style="list-style-type: none"> <li>Hydraulic <b>or Pneumatic</b> Auto Controller System</li> </ol> </li> </ul> <p>STANDARD REQUIREMENT:</p> <ol style="list-style-type: none"> <li>Certification from the manufacturer authenticated by the Philippine Consulate <b>or Notarized from the country of origin</b> of the unit that the warranty should not be affected with a change of distributor</li> <li>Certification that the brand has been in the local market for at least five (5) years with <b>at least five (5) installations</b></li> </ol> <p><b>CANCELLED</b> (Reason: <u>For modification of Technical Specifications</u>)</p>
---	---

**Other Matters:**

- Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top or ring bound instead of fastener, table of contents and index tabs in words, not numbers.
- Folder of Eligibility requirements and technical proposal should be placed in one envelope. And folder of financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as **"Original Bid"**.
- For the Lowest Calculated Bidder, as part of Post Qualification Process, Equipment Demonstration or Site Installation Visit is required.
- Documents should be arranged chronologically according to the checklist issued.
- Typographical error is not acceptable.
- Color code for folders and envelopes:
  - GREEN**  
SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF:
    - MANUAL ROTARY MICROTOME
    - ULTRASONIC SCALER
    - PHYSICAL THERAPY EQUIPMENT
    - GENERAL HOSPITAL EQUIPMENT
- All other provisions on the bidding documents which are not affected shall remain in effect.
- The deadline for **Submission and Opening of Bids** is scheduled on **June 19, 2023, 9:00 AM, Monday** at the **Pag-Asa Hall**, National Center for Mental Health Compound, Mandaluyong City.

- I. Any bid submitted after the deadline for submission shall be declared "LATE" and shall NOT be accepted.
- J. The BAC shall open the bids immediately after the deadline for submission and receipt of bids

For the information and guidance of all concerned.

**-SGD-**

**ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA**  
Chairperson, BAC for Equipment CY 2023