



National Center for Mental Health

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC BIDDING FOR THE SUPPLY AND DELIVERY OF MEDICAL SUPPLIES AND DEVICES CY 2023

This Supplemental Bid Bulletin No. 1 dated **August 29, 2023**, is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

ISSUES	CLARIFICATION
<p>LIST OF ITEMS</p> <p>MSMD-003 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 6.0 (with at least 24 months expiration).</p> <p>MSMD-004 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 6.5 (with at least 24 months expiration).</p> <p>MSMD-005 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 7.0 (with at least 24 months expiration).</p> <p>MSMD-006 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 7.5 (with at least 24 months expiration).</p> <p>MSMD-007 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 8.0 (with at least 24 months expiration).</p> <p>MSMD-008 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 8.5 (with at least 24 months expiration).</p>	<p>LIST OF ITEMS</p> <p>MSMD-003 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 6.0 (with at least 24 months expiration).</p> <p>MSMD-004 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 6.5 (with at least 24 months expiration).</p> <p>MSMD-005 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 7.0 (with at least 24 months expiration).</p> <p>MSMD-006 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 7.5 (with at least 24 months expiration).</p> <p>MSMD-007 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 8.0 (with at least 24 months expiration).</p> <p>MSMD-008 Curved Reinforced Endotracheal Tube, Oral/Nasal, Murphy Eye, High-Volume, Low-Pressure Cuff, sterile, size 8.5 (with at least 24 months expiration).</p>

"There is no Health without Mental Health"

1

<p>MSMD-009 Laryngeal Mask Airway (LMA), sterile, single use, size 4 (with at least 24 months expiration).</p> <p>MSMD-010 Laryngeal Mask Airway (LMA), sterile, single use, size 5 (with at least 24 months expiration).</p> <p>MSMD-014 Spinal Anesthesia Needle, sterile, G. 22 x 3 1/2", must be sharp and sturdy (with at least 24 months expiration).</p> <p>MSMD-023 Corneal knife 2.75mm (6 pc/box)</p> <p>MSMD-024 Corneal Stab knife 15 degree (6 pc/box)</p> <p>MSMD-025 Crescent bevel up knife 2.3mm</p> <p>MSMD-028 Silk Suture, Reverse Cutting, Size 6-0, Black, 18", Needle HE-7, 1/2 Circle (with at least 24 months expiration).</p> <p>MSMD-030 Sodium Hyaluronate 1.5% viscoelastic eye gel, 1mL</p> <p>MSMD-031 Tryphan blue ophthalmic solution, 1mL</p> <p>MSMD-034 Gloves, Nitrile Examination, Disposable powder free, non-sterile, ambidextrous, single use, size MEDIUM, with at least 24 months expiration.</p> <p>QTY: 149,950 UOM: pair</p>	<p>MSMD-009 Laryngeal Mask Airway (LMA), sterile, single use, size 4, double lumen. (with at least 24 months expiration).</p> <p>MSMD-010 Laryngeal Mask Airway (LMA), sterile, single use, size 5, double lumen. (with at least 24 months expiration).</p> <p>MSMD-014 Spinal Anesthesia Needle, sterile, G. 22 x 3 1/2", must be sharp and sturdy (with at least 24 months expiration). <u>Note: Stick to the NCMH Specification</u></p> <p>MSMD-023 Corneal knife 2.75mm (6 pc/box)</p> <p>MSMD-024 Corneal Stab knife 15 degree (6 pc/box)</p> <p>MSMD-025 Crescent bevel up knife from 2.3mm to 2.5mm</p> <p>MSMD-028 Silk Suture, Reverse Cutting, Size 6-0, Black, 18", Needle HE-7, 3/8 - 1/2 Circle (with at least 24 months expiration).</p> <p>MSMD-030 Sodium Hyaluronate at least 1.5% viscoelastic eye gel, 1mL, pre-filled syringe</p> <p>MSMD-031 Tryphan blue ophthalmic solution, 1mL, vial</p> <p>MSMD-034 Gloves, Nitrile Examination, Disposable powder free, non-sterile, ambidextrous, single use, size MEDIUM, with at least 24 months expiration.</p> <p>QTY: 149,950 UOM: pair <u>Note: Stick to the NCMH Specification</u></p>
--	--

<p>MSMD-036 Blood Transfusion Set, sterilized by EO gas, non-toxic, no heavy metal & pyrogen free w/ vein needle G.18, double chamber w/ large mesh filter minimizing platelet loss & hemolysis, flexible table, individually packed into plastic pouch. (with at least 24 months expiration).</p>	<p>MSMD-036 Blood Transfusion Set, sterilized by EO gas, non-toxic, no heavy metal & pyrogen free w/ vein needle G.18, double chamber w/ large mesh filter minimizing platelet loss & hemolysis, flexible table, individually packed into plastic pouch. With standard length of tube of 180cm and dropping of 20 drops/mL. (with at least 24 months expiration).</p>
<p>MSMD-040 Infusion set, PEDIA, vented type (air cap) / 60 micro drops/mL/y-injection port/rubber/ 195cm length, with luer lock (with at least 24 months expiration).</p>	<p>MSMD-040 Infusion set, PEDIA, vented type (air cap) / 60 micro drops/mL/y-injection port/rubber/ 195cm length, with luer lock (with at least 24 months expiration). <u>Note: Stick to the NCMH Specification</u></p>
<p>MSMD-053 Syringes, disposable, 30cc, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).</p>	<p>MSMD-053 Syringes, with G. 18 needle, disposable, 30cc, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).</p>
<p>MSMD-055 Transparent Dressing with absorbent pad, size: D: 3 1/2 x 6 (9cm x 15cm) P: 1 3/4 x 4 (4.5cm x 10cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p>	<p>MSMD-055 Transparent Dressing with absorbent pad, size: D: 4 x 6 (10cm x 15cm) P: 2 x 4 3/8 (5cm x 11cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p>
<p>MSMD-056 Transparent Dressing with absorbent pad, size: D: 3 1/2 x 8 (10cm x 20cm) P: 1 3/4 x 6 (4.5cm x 15cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p>	<p>MSMD-056 Transparent Dressing with absorbent pad, size: D: 4 x 8 (10cm x 20cm) P: 2 x 6 5/16 (5cm x 16cm), waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p>
<p>MSMD-057 Transparent Dressing with absorbent pad, size: D: 3 1/2 x 10 (10cm x 25cm) P: 1 3/4 x 8 (4.5cm x 20cm),</p>	<p>MSMD-057 Transparent Dressing with absorbent pad, size: D: 4 x 10 (10cm x 25cm) P: 2 x 8 (5cm x 20cm), waterproof</p>

f

<p>waterproof and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p> <p>MSMD-062 Anesthesia Set Flex, latex free, adult, breathing circuit with fitting mask - disposable anesthesia set consisting of 2 hose, connector, fastening ring 2 latex free breathing bag</p> <p>MSMD-070 PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 36mm-40mm (with at least 24 months expiration).</p> <p>QTY: 24 UOM: piece</p> <p>MSMD-071 PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 3/0, 75-90cm, round needle, 1/2 circle, 36mm-40mm (with at least 24 months expiration).</p> <p>MSMD-072 'PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p> <p>MSMD-076 'Silk braided, non-absorbable, size 0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm (with at least 24 months expiration).</p> <p>MSMD-077 'Silk braided, non-absorbable, size 1, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm (with at least 24 months expiration).</p>	<p>and impermeable to bacteria and virus, pad provides excellent absorbency, non-adherent and absorbent contact layer to allow wound healing with less trauma and pain when dressing is removed, excellent adhesion balanced with gentleness to skin, no latex, reduces allergic possibility (with at least 18 to 24 months expiration)</p> <p>MSMD-062 Anesthesia Set Flex, latex free, adult, breathing circuit with fitting mask - disposable anesthesia set consisting of 2 hose, connector, fastening ring 3 LITERS latex free breathing bag</p> <p>MSMD-070 PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 2/0, 90cm, round needle, 1/2 circle, 36mm-40mm (with at least 24 months expiration).</p> <p>QTY: 24 UOM: piece <u>Note: Stick to the NCMH Specification</u></p> <p>MSMD-071 PGA (Polyglycolic Acid/Polyglactin), braided coated, braided synthetic absorbable, size 3/0, 75-90cm, round needle, 1/2 circle, 35mm-40mm (with at least 24 months expiration).</p> <p>MSMD-072 'PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 25mm - 40mm (with at least 24 months expiration).</p> <p>MSMD-076 'Silk braided, non-absorbable, size 0, 45cm - 76cm, cutting needle, 1/2 circle / 3/8 circle, 30mm - 45mm (with at least 24 months expiration).</p> <p>MSMD-077 'Silk braided, non-absorbable, size 1, 75cm, cutting needle, 1/2 circle / 3/8 circle, 37mm (with at least 24 months expiration). <u>Note: Stick to the NCMH Specification</u></p>
--	---

<p>MSMD-078 'Silk braided, non-absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with at least 24 months expiration).</p>	<p>MSMD-078 'Silk braided, non-absorbable, size 2/0, 75cm, cutting needle, 1/2 circle / 3/8 circle, 24mm / 26mm (with at least 24 months expiration). <u>Note: Stick to the NCMH Specification</u></p>
<p>MSMD-082 Arterio-Venous Fistula (AVF) kit, inclusive of 4 sterile cherry balls, drape, 2 sterile gauze, antiseptic pads, and 3 10cc syringes with needle, with at least 24 months expiration</p>	<p>MSMD-082 Arterio-Venous Fistula (AVF) kit, inclusive of 4pcs sterile cherry balls, 1pc drape, 2pcs sterile gauze, 2pcs antiseptic pads, 3pcs 10cc syringes with needle g.21, and 2pcs sterile gauze pillows, with at least 24 months expiration</p>
<p>MSMD-100 Fiber-Optic McIntosh (Curved5 blades) for ADULT Laryngoscope</p> <ul style="list-style-type: none"> ◆ LED Light Source ◆ Complete matt type-w/5 S.S Blades Size (0,00,1,2,3) or(0,1,2,3,4) -w/ carrying case Laryngoscope Bulb ◆ LED Light Source/Hallogen ◆ Completely autoclavable ◆ Can be assemble and disassemble ◆ Easy to assemble and disassemble for cleaning maintenance ◆ Easy to detach bulb socket for ease of bulb replacement ◆ The straight and curved blade are compatible to adult/pedia handle ◆ Set of blades are customize - either straight/curved blade Battery <p>➤ Type C Battery use</p> <p>Warranty Certificate 1 year for parts and services which include Preventive and Corrective Maintenance from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.</p>	<p>MSMD-100 Fiber-Optic McIntosh (Curved5 blades) for ADULT Laryngoscope</p> <ul style="list-style-type: none"> ◆ LED Light Source ◆ Complete matt type-w/5 S.S Blades Size (0,00,1,2,3) or(0,1,2,3,4) -w/ carrying case Laryngoscope Bulb ◆ LED Light Source/Hallogen ◆ Completely autoclavable ◆ Can be assemble and disassemble ◆ Easy to assemble and disassemble for cleaning maintenance ◆ Easy to detach bulb socket for ease of bulb replacement ◆ The straight and curved blade are compatible to adult/pedia handle ◆ Set of blades are customize - either straight/curved blade Battery <p>➤ Type C Battery use</p> <p>Warranty Certificate 1 year for parts and services which include Preventive and Corrective Maintenance from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit. <u>Note: Stick to the NCMH Specification</u></p>
<p>MSMD-101 Fiber-Optic McIntosh (Curved5 blades) for PEDIA Laryngoscope</p> <ul style="list-style-type: none"> ◆ LED Light Source ◆ Complete matt type-w/5 S.S Blades Size (0,00,1,2,3) or(0,1,2,3,4) -w/ carrying case Laryngoscope Bulb ◆ LED Light Source/Hallogen 	<p>MSMD-101 Fiber-Optic McIntosh/Miller (Curved/Miller 5 blades) for PEDIA Laryngoscope</p> <ul style="list-style-type: none"> ◆ LED Light Source ◆ Complete matt type-w/5 S.S Blades Size (0,00,1,2,3) or (0,1,2,3,4) -w/ carrying case Laryngoscope Bulb

- ◆ Completely autoclavable
 - ◆ Can be assemble and disassemble
 - ◆ Easy to assemble and disassemble for cleaning maintenance
 - ◆ Easy to detach bulb socket for ease of bulb replacement
 - ◆ The straight and curved blade are compatible to adult/pedia handle
 - ◆ Set of blades are customize - either straight/curved blade Battery
 - Double A/ (AA) Battery use
- Warranty Certificate 1 year for parts and services which include Preventive and Corrective Maintenance from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.

MSMD-113
ECG Trunk Cable with 3-lead wire. (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

**** "Can we provide brochure only?"**

MSMD-114
SPO2 Cable Trusignal with SPO2 Sensor (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

**** "Can we provide brochure only?"**

MSMD-115
NIBP Cuff and hose for ADULT (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

- ◆ LED Light Source/Hallogen
 - ◆ Completely autoclavable
 - ◆ Can be assemble and disassemble
 - ◆ Easy to assemble and disassemble for cleaning maintenance
 - ◆ Easy to detach bulb socket for ease of bulb replacement
 - ◆ The straight and curved blade are compatible to adult/pedia handle
 - ◆ Set of blades are customize - either straight/curved blade Battery
 - Double A/ (AA) Battery use
- Warranty Certificate 1 year for parts and services which include Preventive and Corrective Maintenance from the date of delivery, inspection and acceptance. Perform installation, demonstration/ training for end-user and maintenance personnel before handover of the unit.

MSMD-113
ECG Trunk Cable with 3-lead wire. (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

Note: Stick to the NCMH Specification

NO, the actual sample is needed to test the compatibility of the item. Stick to 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.

MSMD-114
SPO2 Cable Trusignal with SPO2 Sensor (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

Note: Stick to the NCMH Specification

NO, the actual sample is needed to test the compatibility of the item. Stick to 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.

MSMD-115
NIBP Cuff and hose for ADULT (must be compatible to GE B125 Cardiac Monitor) Provide 1 **(ONE) year warranty** certificate from the date of delivery, inspection and acceptance.

<p>** "Can we provide brochure only?"</p> <p>MSMD-116 NIBP Cuff and hose for PEDIA (must be compatible to GE B125 Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.</p>	<p><u>Note: Stick to the NCMH Specification</u></p> <p><i>NO, the actual sample is needed to test the compatibility of the item. Stick to 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.</i></p> <p>MSMD-116 NIBP Cuff and hose for PEDIA (must be compatible to GE B125 Cardiac Monitor) Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.</p> <p><u>Note: Stick to the NCMH Specification</u></p>
<p>SECTION VI SCHEDULE OF REQUIREMENTS</p> <p>MSMD-093 One Time Delivery upon issuance of NTD</p> <p>MSMD-095 One Time Delivery upon issuance of NTD</p> <p>MSMD-096 One Time Delivery upon issuance of NTD</p> <p>MSMD-097 One Time Delivery upon issuance of NTD</p> <p>MSMD-099 One Time Delivery upon issuance of NTD</p> <p>MSMD-100 One Time Delivery upon issuance of NTD</p>	<p>SECTION VI SCHEDULE OF REQUIREMENTS</p> <p>MSMD-093 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p> <p>MSMD-095 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p> <p>MSMD-096 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p> <p>MSMD-097 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.</p> <p>MSMD-099 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p> <p>MSMD-100 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p>

MSMD-101 One Time Delivery upon issuance of NTD	MSMD-101 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMD-102 One Time Delivery upon issuance of NTD	MSMD-102 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMSD-103 One Time Delivery upon issuance of NTD	MSMSD-103 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMD-104 One Time Delivery upon issuance of NTD	MSMD-104 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMSD-105 One Time Delivery upon issuance of NTD	MSMSD-105 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMD-106 One Time Delivery upon issuance of NTD.	MSMD-106 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.
MSMD-107 One Time Delivery upon issuance of NTD	MSMD-107 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.
MSMD-108 One Time Delivery upon issuance of NTD	MSMD-108 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.
MSMD-109 One Time Delivery upon issuance of NTD	MSMD-109 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.

f


<p>MSMD-110 One Time Delivery upon issuance of NTD</p> <p>MSMD-112 One Time Delivery upon issuance of NTD</p> <p>MSMD-117 One Time Delivery upon issuance of NTD</p> <p>MSMD-118 to MSMD-177 One Time Delivery upon issuance of NTD</p>	<p>MSMD-110 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.</p> <p>MSMD-112 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.</p> <p>MSMD-117 One Time Delivery upon issuance of NTD Delivery of Devices for 30 days upon issuance of NTD.</p> <p>MSMD-118 to MSMD-177 One Time Delivery upon issuance of NTD Delivery of Devices for 60 days upon issuance of NTD.</p>
<p>ADDITIONAL QUERIES:</p> <ol style="list-style-type: none"> 1) Pwede po ba yung Private Account sa SLCC? 2) Ang SLCC po ba ay yung item to bid lang o yung buong contract? 3) Sa SLCC po, yung supply and delivery of medical supplies and devices, pwede po ba na medical supplies lang ang similar? 4) Clarification lang po, meaning kasi may mga devices po kayo the part of the device po, so any of those na 25 %, correct? Kasi sa list of items niyo po kasi kinombine niyo po yung medical supplies and medical devices so it's a combination po ba, ok lang po ba? 5) So pwede po yung mga current na mga suction machines, so those with the devices tsaka mga equipment po? 	<ol style="list-style-type: none"> 1) YES, pwede po Private Account. 2) Ang SLCC po ay yung items lang na ibibid po. 3) Yes, pwede po. 4) Sa SLCC po yung mga devices po kasi pasok na po siya sa semi expendable supplies namin kaya ok na siya sa 25%, kahit supplies pasok na po. 5) Ang less than 50 thousand ay kinokonsider as semi expendable so medical devices po siya. Yung mga devices na yun ay pwede siyang ilagay na 25% para sa SLCC ng supplies, hindi siya

<p>6) Kapag po ba nagsubmit po ako ng SLCC ko na lahat is puro mga microscope at BP apparatus icoconsider niyo po kahit na meron akong supplies na kasama?</p> <p>7) Yung sample po ng medical devices, isasubmit po ba namin or kapag po nagging Lowest Calculated Bidder po kami?</p> <p>8) Kailan po kami magsasubmit ng sample?</p>	<p>kailangan na equipment ang ilagay niyo diyan.</p> <p>Medical devices at supplies, iisa na lang siya.</p> <p>6) Yes, considered po.</p> <p>7) Kapag kayo po yung Lowest Calculated Bidder.</p> <p>8) Kapag nakatanggap po kayo ng "Notice of Eligibility", meron po kayong five (5) calendar days para magsubmit ng samples.</p>
---	--

Other Matters:

- A. *Eligibility requirements and technical proposal* should be in *one folder* and *financial proposal* in a *separate folder*, with shoelace on top or ring bound instead of a fastener, table of contents, and index tabs in words, not numbers.
- B. *Folder of Eligibility requirements and technical proposal* should be placed in *one envelope*. And the *folder of the Financial proposal* should be in *another envelope*. Both envelopes shall then be placed in *one mother envelope* marked as "**Original Bid**"
- C. Documents should be arranged chronologically according to the checklist issued.
- D. Color code for folders and envelope:
YELLOW – Public Bidding for the Supply and Delivery of Medical Supplies and Devices CY 2023
- E. All other provisions on the bidding documents which are not affected shall remain in effect.
- F. The deadline for **Submission and Opening of Bids** is scheduled on **September 06, 2023 (Wednesday), 9:00 AM**, at the **Pag-Asa Hall**, National Center for Mental Health Compound, Mandaluyong City.
- G. Any bid submitted after the deadline for submission shall be declared "**LATE**" and shall NOT be accepted.
- H. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.

For the information and guidance of all concerned.


ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
 Chairperson, BAC for Goods