



National Center for Mental Health

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC RE-BIDDING FOR SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF FLOOR-MOUNTED DIGITAL X-RAY MACHINE CY2023

This Supplemental Bid Bulletin No. 1 dated, **September 25, 2023** is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

| ISSUES | CLARIFICATION |
|--|---|
| <p style="text-align: center;"><u>FROM</u></p> <p style="text-align: center;">SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF FLOOR-MOUNTED DIGITAL X-RAY MACHINE</p> <p style="text-align: center;">SECTION VII TECHNICAL SPECIFICATIONS</p> <p>DIGITAL DETECTOR 9. Auto detect location of detector</p> <p>13. Dust and water resistant (at least IP55)</p> <p>VERTICAL BUCKY STAND 15. Auto sync/tracking feature between tube and vertical bucky</p> <p>PATIENT TABLE 16. Auto tracking: Longitudinal tracking of detector with longitudinal travel and rotation of tube.</p> <p>STANDARD ACCESSORIES 24. One (1) upright gonad shield 25. One (1) upright gonad shield</p> <p>27. QA Phantom</p> <p>INCLUSIONS 28. Four (4) workstations with at least;</p> | <p style="text-align: center;"><u>TO</u></p> <p style="text-align: center;">SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND TESTING OF FLOOR-MOUNTED DIGITAL X-RAY MACHINE</p> <p style="text-align: center;">SECTION VII TECHNICAL SPECIFICATIONS</p> <p>DIGITAL DETECTOR 9. <u>The machine has a mechanism to detect the location of the detector</u></p> <p>13. <u>Dust and water resistant (at least IP54)</u></p> <p>VERTICAL BUCKY STAND 15. <u>AS IS</u></p> <p>PATIENT TABLE 16. <u>AS IS</u></p> <p>STANDARD ACCESSORIES 24. <u>One (1) upright gonad shield</u> 25. <u>Two (2) movable lead barriers</u> 26. <u>QA Phantom</u> a. <u>Collimation and beam alignment test</u> b. <u>Spatial Resolution test</u> c. <u>Contrast Resolution test</u> d. <u>Uniformity test</u></p> <p>26. <u>AS IS</u></p> <p>INCLUSIONS 27. <u>AS IS</u></p> |

"There is no Health without Mental Health"



| | |
|---|---|
| <p>STANDARD REQUIREMENTS</p> <p>55. Certification that the brand and model must be in the local market for a minimum of Five (5) years with at least five (5) Full digital X-ray of the same brand in tertiary government facilities (at least 1 must be DOH hospitals).</p> | <p>STANDARD REQUIREMENTS</p> <p>55. Certification that the brand and model must be in the local market for a minimum of Five (5) years with at least five (5) Full digital X-ray of the same brand <u>in both private and government facilities (at least one (1) must be a tertiary DOH hospital) in case a newer model is to be provided, the older model must be in the market for a minimum of five (5) years.</u></p> |
|---|---|

Other Matters:

- A. Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top or ring bound instead of fastener, table of contents and index tabs in words, not numbers.
- B. Folder of Eligibility requirements and technical proposal should be placed in one envelope. And folder of financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as **“Original Bid”**.
- C. Documents should be arranged chronologically according to the checklist issued.
- D. Typographical error must be countersigned by the authorized representative.
- E. Color code for folders and envelopes:
 - **WHITE**
SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING, AND TESTING OF FLOOR-MOUNTED DIGITAL X-RAY CY 2023
- F. All other provisions on the bidding documents which are not affected shall remain in effect.
- G. The deadline for **Submission and Opening of Bids** is scheduled on **October 2, 2023, 9:00 AM, Monday** at the **BAC Conference Room**, National Center for Mental Health Compound, Mandaluyong City.
- H. Any bid submitted after the deadline for submission shall be declared **“LATE”** and shall **NOT** be accepted.
- I. The BAC shall open the bids immediately after the deadline for submission and receipt of bids

For the information and guidance of all concerned.


ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
 Chairperson, BAC for Equipment CY 2023