



National Center for Mental Health

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW POLYSOMNOGRAPHY MACHINE FOR SLEEP CLINIC CY 2024

This Supplemental Bid Bulletin No. 1 dated **June 11, 2024** is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

A. Amendments / Clarifications:

| FROM | TO |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>Under Section VI: Schedule of Requirements</p> <p>On Delivery Period:</p> <p>a. One time delivery within (30 calendar days upon receipt of Notice to Deliver.</p> | <p>Under Section VI: Schedule of Requirements</p> <p>On Delivery Period:</p> <p>a. One time delivery within 30-60 <u>calendar days upon receipt of Notice to Deliver.</u></p> |
| <p>ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS</p> <p>Under Specifications:</p> <p>Included Accessories:</p> <p>4. Replacement Fuse for Base Unit</p> <p>5. USB-to-Ethernet Adaptor</p> <p>6. Quantum Nonin Oximeter Adapter Cable</p> <p><u>Request:</u></p> <p>➤ "To remove the above specifications"</p> <p><u>Justification:</u></p> <p>➤ "The amplifiers and modules of PSG models suited for the project have upgraded built-in capabilities and don't need the following since it will entail less cost for the end-users since the mentioned accessories needs to be replaced every year according to our man."</p> <p>➤ "Since this is being utilized in a specific brand."</p> | <p>ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS</p> <p>Under Specifications:</p> <p>Included Accessories:</p> <p>4. Replacement Fuse for Base Unit</p> <p>5. USB-to-Ethernet Adaptor</p> <p>6. Quantum Nonin Oximeter Adapter Cable</p> <p><u>Answer:</u></p> <p>➤ "Deleted."</p> |
| <p>7. Nonin – Reusable Soft Sensor, Medium, model 8000SM</p> | <p><u>7. Reusable Soft Sensor, Medium, model 8000SM</u></p> |

**Under List of Items: Standard Requirements /
Section III: Bid Data Sheet:**

- h. CTC of IEC 60601-1:2012 - General Safety Third Edition
- i. CTC of CAN / CSA-C22.2 No. 60601-1:08(R2013) +C2:2011
- j. CTC of IEC 60601-1-6:2010 – Usability Third Edition
- k. CTC of IEC 62366:2007, Edition 1.0
- l. CTC of IEC 60601-2-26:2012 – Electroencephalographs Third Edition
- m. CTC of IEC60601-2-61:2011 - Pulse Oximeters
- n. CTC of EN ISO 80601-2-61:2011, Edition 1
- o. CTC of IEC 60601-1-2:2014 – EMC Fourth Edition
- p. CTC of IEC 61000-3-2:2014, Fourth Edition
- q. CTC of IEC 61000-3-2 Harmonic emissions – Class A
- r. CTC of IEC 61000-3-3:2013, Third Edition Voltage Fluctuations/ Flicker emissions
- s. CTC of CISPR11, Edition 5.0 A1:2010 RF emissions – Group 1, Class A
- t. CTC of IEC 61000-4-2:2008, Second Edition
- u. CTC of IEC 61000-4-2 Electrostatic Discharge (ESD) ±8 kV contact (to the patient lead and exposed metal), ±15 kV air
- v. CTC of IEC 61000-4-3 Third Edition with A1:2007+A2:2010
- w. CTC of IEC 61000-4-3, 3 Vrms, 80 MHz to 2.7 GHz
- x. CTC of IEC 61000-4-4:2012, Third Edition, fast transient/burst ±2 kV power supply ±1 kV
- y. CTC of IEC 61000-4-5:2014, Third Edition
- z. CTC of IEC 61000-4-5 ±1 kV Surge differential mode ±2 kV common mode
- aa. CTC of IEC 61000-4-6 Second Edition with A1:2004 + A2:2006
- bb. CTC of IEC 61000-4-6, 150 kHz to 80 MHz
- cc. CTC of IEC 61000-4-8:2009, Second Edition
- dd. CTC of IEC 61000-4-8, Power frequency (50/60 Hz) magnetic field, 30 A/m

**Under List of Items: Standard Requirements /
Section III: Bid Data Sheet:**

- ~~h. CTC of IEC 60601-1:2012 – General Safety Third Edition~~
- ~~i. CTC of CAN / CSA-C22.2 No. 60601-1:08(R2013) +C2:2011~~
- ~~j. CTC of IEC 60601-1-6:2010 – Usability Third Edition~~
- ~~k. CTC of IEC 62366:2007, Edition 1.0~~
- ~~l. CTC of IEC 60601-2-26:2012 – Electroencephalographs Third Edition~~
- ~~m. CTC of IEC60601-2-61:2011 - Pulse Oximeters~~
- ~~n. CTC of EN ISO 80601-2-61:2011, Edition 1~~
- ~~o. CTC of IEC 60601-1-2:2014 – EMC Fourth Edition~~
- ~~p. CTC of IEC 61000-3-2:2014, Fourth Edition~~
- ~~q. CTC of IEC 61000-3-2 Harmonic emissions – Class A~~
- ~~r. CTC of IEC 61000-3-3:2013, Third Edition Voltage Fluctuations/ Flicker emissions~~
- ~~s. CTC of CISPR11, Edition 5.0 A1:2010 RF emissions – Group 1, Class A~~
- ~~t. CTC of IEC 61000-4-2:2008, Second Edition~~
- ~~u. CTC of IEC 61000-4-2 Electrostatic Discharge (ESD) ±8 kV contact (to the patient lead and exposed metal), ±15 kV air~~
- ~~v. CTC of IEC 61000-4-3 Third Edition with A1:2007+A2:2010~~
- ~~w. CTC of IEC 61000-4-3, 3 Vrms, 80 MHz to 2.7 GHz~~
- ~~x. CTC of IEC 61000-4-4:2012, Third Edition, fast transient/burst ±2 kV power supply ±1 kV~~
- ~~y. CTC of IEC 61000-4-5:2014, Third Edition~~
- ~~z. CTC of IEC 61000-4-5 ±1 kV Surge differential mode ±2 kV common mode~~
- ~~aa. CTC of IEC 61000-4-6 Second Edition with A1:2004 + A2:2006~~
- ~~bb. CTC of IEC 61000-4-6, 150 kHz to 80 MHz~~
- ~~cc. CTC of IEC 61000-4-8:2009, Second Edition~~
- ~~dd. CTC of IEC 61000-4-8, Power frequency (50/60 Hz) magnetic field, 30 A/m~~


| | |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p><u>Request:</u></p> <ul style="list-style-type: none"> ➤ "To provide a product conformity certificate from the manufacturer claiming that the following safety standards are complied with." <p>Justification:</p> <ul style="list-style-type: none"> ➤ "Since some of the certificates are revised and are the updated version of the required certificates." | <p><u>Answer:</u></p> <ul style="list-style-type: none"> ➤ "Deleted the above-mentioned requirements. Notarized certification of product conformity from the manufacturer claiming that the Safety Standards are complied with." |
| <p>mm. Notarized certification from the manufacturer authenticated by the Philippine consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor.</p> <p><u>Request:</u></p> <ul style="list-style-type: none"> ➤ "Local Notary Public can be applied in replacement for the authentication from the Philippine consulate." <p>Justification:</p> <ul style="list-style-type: none"> ➤ "Since the process will take some time to release, the schedules might need to meet the specific time frame for this project. And also the cost will be high for this." | <p>mm. Notarized certification from the manufacturer authenticated by the Philippine consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor. <u>(Local Notary Public can be applied)</u></p> |

Other Matters:

- A. Bid Security shall be 120 calendar days from the date of submission and opening of bids.
- B. Section VII: Technical Specifications - Please write "COMPLY" to indicate that requirements are met.
- C. *Eligibility requirements and technical proposal* should be in *one folder* and *financial proposal* in a *separate folder*, with shoelace on top instead of fastener, table of contents and index tabs in words, not numbers.
- D. *Folder of Eligibility requirements and technical proposal* should be placed in *one envelope*. And *folder of Financial proposal* should be in *another envelope*. Both envelopes shall then be placed in *one mother envelope* marked as "**Original Bid**"
- E. Documents should be arranged chronologically according to the checklist issued.
- F. Color code for folders and envelope: **GREEN**
- G. Typographical error must be counter signed by the authorized representative.
- H. All other provisions on the bidding documents which are not affected shall remain in effect.

- I. The deadline for **Submission and Opening of Bids** is on **June 18, 2024, 9:00 AM,** at the **BAC Conference Room**, National Center for Mental Health Compound, Mandaluyong City.
- J. Any bid submitted after the deadline for submission shall be declared "**LATE**" and shall NOT be accepted.
- K. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.

For the information and guidance of all concerned.


ALDEN C. CUYOS, MD, FPPA, IFAPA
Chairperson, BAC for Equipment CY 2024