



**SUPPLEMENTAL BID BULLETIN**

**ADDENDUM NO. 1**

**PUBLIC BIDDING FOR THE SUPPLY AND DELIVERY OF MEDICAL SUPPLIES CY 2025  
(EARLY PROCUREMENT ACTIVITY)**

This Supplemental Bid Bulletin No. 1 dated **November 5, 2024**, is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

| ISSUES   | CLARIFICATION  |
|--|--|
| <p><b>SECTION VII</b><br/><b>TECHNICAL SPECIFICATIONS</b></p> <p>ITEM CODE (MS25-09)<br/>Aneroid Sphygmomanometer with Stethoscope Manual BP Monitor - Aneroid Sphygmomanometer. • - Measures blood pressure with an accuracy of - or - 3mmHg. •Luminous gauge • Index and range markings on the dial for clear readings. •Durable and stain resistant natural cotton fiber cuff (Velcro closure) •Latex-free inflation system. • Inflation Bulb - Valve: Chrome-plated brass screw-type valve delivers precise deflation control rate. The inflation bulb is constructed of crack-resistant non-sticking high-density hypoallergenic latex-free PVC. Stethoscope has double-sided non-chill chest piece. This specification shall be reflected or visible in the product manufacturer's specification details and/or manual. Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.</p> <p>ITEM CODE (MS25-28)<br/>COVERALL SUITS, XXL (Height 186-194cm, Chest 115-124cm) - XXXL (Height 194-200cm, Chest 124-132cm) - 60 GSM with <u>Shoe Cover</u>. Made from non-woven polypropelene Fabric Material, breathable film.</p> <p><i>** Is it okay to remove the "shoe cover"?</i></p> | <p>ITEM CODE (MS25-09)<br/>Aneroid Sphygmomanometer with Stethoscope Manual BP Monitor - Aneroid Sphygmomanometer. • -Measures blood pressure with an accuracy of - or - <b>3mmHg</b>. •Luminous gauge • Index and range markings on the dial for clear readings. •Durable and stain resistant natural cotton fiber cuff (Velcro closure) •Latex-free inflation system. • Inflation Bulb - Valve: Chrome-plated brass screw-type valve delivers precise deflation control rate. The inflation bulb is constructed of crack-resistant non-sticking high-density hypoallergenic latex-free PVC. Stethoscope has double-sided non-chill chest piece. This specification shall be reflected or visible in the product manufacturer's specification details and/or manual. Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance. <b>With Calibration Certificate.</b></p> <p>ITEM CODE (MS25-28)<br/>COVERALL SUITS, XXL (Height 186-194cm, Chest 115-124cm) - XXXL (Height 194-200cm, Chest 124-132cm) - 60 GSM with Shoe Cover. Made from non-woven polypropelene Fabric Material, breathable film.</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |

*"There is no Health without Mental Health"*



*Jpa*

|  |  |
|--|--|
| <p>ITEM CODE (MS25-31)</p> <p>Disposable ECG Electrodes, <u>ADULT</u>, at least 36x40mm (oval) / circular snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. <u>(with at least 24 months expiration)</u>.</p> <p><i>** Is it okay if at least 18 months expiration?</i></p> <p><i>** Range Electrode Adult to Adult/Pedia</i></p> | <p>ITEM CODE (MS25-31)</p> <p>Disposable ECG Electrodes, ADULT, at least 36x40mm (oval) / circular snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-62)</p> <p>IV CANNULA G.18 <u>with transparent or white plastic safety protection device</u>, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 45mm length. (with at least 24 months expiration)</p> <p><i>** with safety protection device (no color required)</i></p>                     | <p>ITEM CODE (MS25-62)</p> <p>IV CANNULA G.18 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 45mm length. (with at least 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>               |
| <p>ITEM CODE (MS25-63)</p> <p>IV CANNULA G.20 <u>with transparent or white plastic safety protection device</u>, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 32mm length. (with at least 24 months expiration)</p> <p><i>** with safety protection device (no color required)</i></p>                     | <p>ITEM CODE (MS25-63)</p> <p>IV CANNULA G.20 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 32mm length. (with at least 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>               |
| <p>ITEM CODE (MS25-64)</p> <p>IV CANNULA G.22 <u>with transparent or white plastic safety protection device</u>, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous</p>  | <p>ITEM CODE (MS25-64)</p> <p>IV CANNULA G.22 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter,</p>   |

*fu*

|  |  |
|--|--|
| <p>hydrophobic filter, 25 mm length. (with at least 24 months expiration)</p> <p><b>** with safety protection device (no color required)</b></p>   | <p>25 mm length. (with at least 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-65)</p>   | <p>ITEM CODE (MS25-65)</p>   |
| <p>IV CANNULA G.24 <u>with transparent or white plastic safety protection device</u>, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length. (with at least 24 months expiration)</p> <p><b>** with safety protection device (no color required)</b></p> | <p>IV CANNULA G.24 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length. (with at least 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-66)</p>   | <p>ITEM CODE (MS25-66)</p>   |
| <p>IV CANNULA G.26 <u>with transparent or white plastic safety protection device</u>, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length. (with at least 24 months expiration)</p> <p><b>** with safety protection device (no color required)</b></p> | <p>IV CANNULA G.26 with transparent or white plastic safety protection device, injection port and wide wings with ultra lancet point and ultra sharp triple bevel needle, silicon coated, made of polyurethane/FEP radio-opaque strip on catheter with removable micro-porous hydrophobic filter, 19 mm length. (with at least 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-76)</p>   | <p>ITEM CODE (MS25-76)</p>   |
| <p>Mask, surgical disposable, 3 PLY, Elastic without glass fibres, hypoallergenic, very low resistance to breathing, nose baradaptable, high filtration capacity, perfect fitting, EAR-LOOP</p> <p><b>** Do you have a preferred color for the mask?</b></p>   | <p>Mask, surgical disposable, 3 PLY, Elastic without glass fibres, hypoallergenic, very low resistance to breathing, nose baradaptable, high filtration capacity, perfect fitting, EAR-LOOP</p> <p><b>NONE.</b></p>  |
| <p>ITEM CODE (MS25-83)</p>   | <p>ITEM CODE (MS25-83)</p>   |
| <p>N95 mask, particulate respirator contains no components made from natural rubber latex, Collapse resistant cup shape design. Two-strap design with</p>  | <p>N95 mask, particulate respirator contains no components made from natural rubber latex, Collapse resistant cup shape design. Two-strap design with welded dual point</p>  |

*fw*

welded dual point attachment, braided headbands, cushioning nose foam, and lightweight construction for comfortable wear, easy breathing/ speaking. Made from synthetic plastic fibers, usually polypropylene (PP). They also contain rubber and metal for the straps, staples, and nose bar. with NIOSH Certificate (with at least 24 months expiration)

**\*\* Do you require fitness of the mask?**

ITEM CODE (MS25-96)

Pre-filled syringe 11mL containing sterile lubricating jelly, 2% lidocaine and 0.255 chlorhexidine gluconate. Syringe with long nozzle and blunt edge for easy insertion to urethra and to avoid mucosal injury. (with at least 24 months expiration).

ITEM CODE (MS25-98)

Pulse Oximeter, fingertip, SPO2 and pulse rate (PR) display, PR waveform and bargraph display, built in alarm can be set to high and low, adjustable auto alarm range, simple operation, low power consumption, four directions and eight display modes, color OLED, automatic power off, with USB rechargeable cord. with Certificate of Calibration. Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.

ITEM CODE (MS25-102)

Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 1 inch x 10 yards (with at least 24 months expiration).

ITEM CODE (MS25-103)

Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 2 inches x 10 yards (with at least 24 months expiration).

attachment, braided headbands, cushioning nose foam, and lightweight construction for comfortable wear, easy breathing/ speaking. Made from synthetic plastic fibers, usually polypropylene (PP). They also contain rubber and metal for the straps, staples, and nose bar. with NIOSH Certificate (with at least 24 months expiration)

**Not required.**

ITEM CODE (MS25-96)

Pre-filled syringe 11mL containing sterile lubricating jelly, 2% lidocaine and **0.25** chlorhexidine gluconate. Syringe with long nozzle and blunt edge for easy insertion to urethra and to avoid mucosal injury. (with at least 24 months expiration).

ITEM CODE (MS25-98)

Pulse Oximeter, fingertip, SPO2 and pulse rate (PR) display, PR waveform and bargraph display, built in alarm can be set to high and low, adjustable auto alarm range, simple operation, low power consumption, four directions and **at least six (6)** display modes, color OLED, automatic power off, with USB rechargeable cord. with Certificate of Calibration. Provide 1 (ONE) year warranty certificate from the date of delivery, inspection and acceptance.

ITEM CODE (MS25-102)

Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size **at least** 1 inch x 10 yards (with at least 24 months expiration).

ITEM CODE (MS25-103)

Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size **at least** 2 inches x 10 yards (with at least 24 months expiration).

|   |   |
|---|---|
| <p>ITEM CODE (MS25-104)</p> <p>Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size 3 inches x 10 yards (with at least 24 months expiration).</p>   | <p>ITEM CODE (MS25-104)</p> <p>Silk-like cloth tape, Convenient bi-directional tear, white color, hypoallergenic and latex free, size <b>at least</b> 3 inches x 10 yards (with at least 24 months expiration).</p>   |
| <p>ITEM CODE (MS25-111)</p> <p>Surface Disinfectant, Towelettes / Wipes. Cleaner and disinfectant in one. Kill clinically relevant organisms such as bacteria, viruses, fungi and TB in just 3 minutes. Ready-to-use with no dilution or mixing required. Clean, disinfect and decontaminate the hard, non-porous surfaces with speed and confidence. Durable, nonwoven towelettes will not bunch up during use. 220 pulls per cannister. With at least 24 months expiration.</p> <p>UOM: <u>cannister</u></p> <p><b>** Can we offer pack instead of cannister?</b></p> | <p>ITEM CODE (MS25-111)</p> <p>Surface Disinfectant, Towelettes / Wipes. Cleaner and disinfectant in one. Kill clinically relevant organisms such as bacteria, viruses, fungi and TB in just 3 minutes. Ready-to-use with no dilution or mixing required. Clean, disinfect and decontaminate the hard, non-porous surfaces with speed and confidence. Durable, nonwoven towelettes will not bunch up during use. 220 pulls per cannister. With at least 24 months expiration.</p> <p><b>NO. Cannister only. Stick to NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-116)</p> <p>Syringe, disposable, Slip Tip 1cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.25 x 1" (with at least 24 months expiration).</p> <p><b>** A separate and detached Syringe Barrel and Needle is acceptable</b></p>                               | <p>ITEM CODE (MS25-116)</p> <p>Syringe, disposable, Slip Tip 1cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.25 x 1" (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-117)</p> <p>Syringe 3cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, <u>with retractable needle safety feature</u>, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p>   | <p>ITEM CODE (MS25-117)</p> <p>Syringe 3cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, with retractable needle safety feature, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p>  |

*Qu*

|   |   |
|---|---|
| <p><b>** Can you remove the word "retractable" and retain the safety feature?</b></p> <p><b>** with retractable <u>or</u> needle safety feature?</b></p> <p>ITEM CODE (MS25-118)</p> <p>Syringe, disposable, Luer-Lok, Tip 3cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.23 x 1" (with at least 24 months expiration).</p> <p><b>** separate and detached Syringe Barrel and Needle is acceptable</b></p> | <p><b>NO. Stick with NCMH Specifications.</b></p> <p><b>NO. Stick with NCMH Specifications.</b></p> <p>ITEM CODE (MS25-118)</p> <p>Syringe, disposable, Luer-Lok, Tip 3cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.23 x 1" (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-119)</p> <p>Syringe 3cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, <u>with retractable needle safety feature</u>, Gauge.23 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>** with retractable <u>or</u> needle safety feature</b></p>   | <p>ITEM CODE (MS25-119)</p> <p>Syringe 3cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, with retractable needle safety feature, Gauge.23 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-120)</p> <p>Syringe 5cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, <u>with retractable needle safety feature</u>, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>** with retractable <u>or</u> needle safety feature</b></p>   | <p>ITEM CODE (MS25-120)</p> <p>Syringe 5cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, with retractable needle safety feature, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |

|  |  |
|--|--|
| <p>ITEM CODE (MS25-121)</p> <p>Syringe, disposable, Luer-Lok, Tip 5cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.23 x 1" (with at least 24 months expiration).</p> <p><b>** A separate and detached Syringe Barrel and Needle is acceptable.</b></p>  | <p>ITEM CODE (MS25-121)</p> <p>Syringe, disposable, Luer-Lok, Tip 5cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.23 x 1" (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-122)</p> <p>Syringe, disposable, Luer-Lok, Tip 10cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.21 x 1" (with at least 24 months expiration).</p> <p><b>** A separate and detached Syringe Barrel and Needle is acceptable.</b></p> | <p>ITEM CODE (MS25-122)</p> <p>Syringe, disposable, Luer-Lok, Tip 10cc, non-toxic, non-pyrogenic. Clear transparent barrel, smooth plunger action - clear scale markings, accurate and readable calibration in cc or mL, marking cannot be erased by any liquid and with plastic packaging, with plunger retention ring. With hypoallergenic, ultra sharp, smooth and glides easily Safety cap/lock Needle G.21 x 1" (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-123)</p> <p>Syringe 10cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, <u>with retractable needle safety feature</u>, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>** with retractable <u>or</u> needle safety feature</b></p>   | <p>ITEM CODE (MS25-123)</p> <p>Syringe 10cc, sterile, siliconized and transparent barrel, firm and definite end stop plunger, clear and visible scale markings, latex and PVC free, with visible back flow on luer lock, attached ultra sharp honed and beveled hypodermic needle, with retractable needle safety feature, Gauge.21 x 1" (100pieces/box). With at least 24 months expiration. (for Clinical Laboratory use)</p> <p><b>NO. Stick with NCMH Specifications.</b></p>                                |
| <p>ITEM CODE (MS25-130)</p> <p>Thermometer infrared, with memory storage, non-contact, fast measurement, beep alarm, measure range: 32C to 43C, temperature unit: celsius, display screen: backlight display, measure position: forehead, measure distance: 3-7cm, with AAA batteries</p>  | <p>ITEM CODE (MS25-130)</p> <p>Thermometer infrared, with memory storage, non-contact, fast measurement, beep alarm, measure range: 32C to 43C, temperature unit: celsius, display screen: backlight display, measure position: forehead, measure distance: 3-7cm, with AAA batterie. <b>With Calibration Certificate.</b></p>   |

|   |  |
|---|--|
| <p>ITEM CODE (MS25-153)</p> <p><u>Transpore Plaster</u> 1 inch</p>  | <p>ITEM CODE (MS25-153)</p> <p><b>Transparent Tape</b> 1 inch</p>  |
| <p>ITEM CODE (MS25-154)</p> <p>Ventilator Breathing Circuit - 2 limbs (with extra 150cc corrugated tube for the ventilator and humidifier connection), 2 H2O traps, non-heated, with nebulizer kit, disposable (single patient only) with at least 24 months expiration.</p>  | <p>ITEM CODE (MS25-154)</p> <p>Ventilator Breathing Circuit <b>ADULT</b> - 2 limbs (with extra 150cc corrugated tube for the ventilator and humidifier connection), 2 H2O traps, non-heated, with nebulizer kit, disposable (single patient only) with at least 24 months expiration.</p>  |
| <p>ITEM CODE (MS25-159)</p> <p>Autoclave tape indicator, high-tack adhesive, clean removal, saturated crepe paper backing with pressure-sensitive adhesive, for steam and chemical vapor sterilizers, lead free, latex free, <u>1 inch x 35 yards</u> (with at least 24 months expiration).</p> <p><b>** Range 1 to 1.9 inches 35m to 50m</b></p> | <p>ITEM CODE (MS25-159)</p> <p>Autoclave tape indicator, high-tack adhesive, clean removal, saturated crepe paper backing with pressure-sensitive adhesive, for steam and chemical vapor sterilizers, lead free, latex free, 1 inch x 35 yards (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-162)</p> <p>Cautery Pad, adult disposable, solid gel, foam support 202x101mm, bipolar (REM Control), <u>300m</u> cable and valleylab connector (with at least 24 months expiration).</p> <p><b>** Is it 300 meters or 300 centimeters?</b></p> <p><b>** Do you accept at least 270 cm?</b></p>                                 | <p>ITEM CODE (MS25-162)</p> <p>Cautery Pad, adult disposable, solid gel, foam support 202x101mm, bipolar (REM Control), <b>300cm</b> cable and valleylab connector (with at least 24 months expiration).</p> <p><b>300 centimeters.</b></p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-163)</p> <p>Cautery Pencil - Hand controlled with ultra clean blade electrode, safety holster, <u>Teflon coated tip</u>. (with at least 24 months expiration).</p> <p><b>** Consider if "with or without" Teflon</b></p>   | <p>ITEM CODE (MS25-163)</p> <p>Cautery Pencil - Hand controlled with ultra clean blade electrode, safety holster, Teflon coated tip. (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p>   |



|  |   |
|--|---|
| <p>ITEM CODE (MS25-199)</p> <p>PGA (Polyglycolic / Polyglactic Acid), braided, coated, braided synthetic absorbable, size 0, 90cm, <u>37mm</u>, taper point / round needle, 1/2 circle (with at least 24 months expiration).</p> <p><b>** 35 – 37mm?</b></p>   | <p>ITEM CODE (MS25-199)</p> <p>PGA (Polyglycolic / Polyglactic Acid), braided, coated, braided synthetic absorbable, size 0, 90cm, <u>37mm</u>, taper point / round needle, 1/2 circle (with at least 24 months expiration).</p> <p><b>NO. Stick with NCMH Specifications.</b></p>  |
| <p>ITEM CODE (MS25-226)</p> <p>Transparent + Pad Film Dressing - Sterile, waterproof breathable transparent dressing with absorbent Non-Adherent Pad. With viral and bacterial barrier. Dressing Size: <u>9 cm</u> x 15 cm, Pad size: 4.5 cm x 10 cm. (with at least 18 to 24 months expiration)</p> <p><b>**Is it okay if 8 – 9cm?</b></p>  | <p>ITEM CODE (MS25-226)</p> <p>Transparent + Pad Film Dressing - Sterile, waterproof breathable transparent dressing with absorbent Non-Adherent Pad. With viral and bacterial barrier. Dressing Size: 9 cm x 15 cm, Pad size: 4.5 cm x 10 cm. (with at least 18 to 24 months expiration)</p> <p><b>NO. Stick with NCMH Specifications.</b></p> |
| <p>ITEM CODE (MS25-227)</p> <p>Transparent + Pad Film Dressing - Sterile, waterproof breathable transparent dressing with absorbent Non-Adherent Pad. With viral and bacterial barrier. Dressing Size: <u>9 cm</u> x 20 cm, Pad size: 4.5 cm x 15 cm. (with at least 18 to 24 months expiration)</p> <p><b>**Is it okay if 9 – 10cm?</b></p> | <p>ITEM CODE (MS25-227)</p> <p>Transparent + Pad Film Dressing - Sterile, waterproof breathable transparent dressing with absorbent Non-Adherent Pad. With viral and bacterial barrier. Dressing Size: <b>at least</b> 9 cm x 20 cm, Pad size: 4.5 cm x 15 cm. (with at least 18 to 24 months expiration)</p>                                   |
| <p>ITEM CODE (MS25-266)</p> <p>Dialysis Needle for Fistula and Graft, G15, <u>venous</u>, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration</p> <p><b>** Is venous with back-eye?</b></p>   | <p>ITEM CODE (MS25-266)</p> <p>Dialysis Needle for Fistula and Graft, G15, venous, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration</p> <p><b>Stick with indicates specifications. Back eye is not needed for this item.</b></p>  |
| <p>ITEM CODE (MS25-268)</p> <p>Dialysis Needle for Fistula and Graft, G16, <u>venous</u>, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration</p>   | <p>ITEM CODE (MS25-268)</p> <p>Dialysis Needle for Fistula and Graft, G16, venous, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration</p>   |

*fm*


|   |   |
|---|---|
| <p><b>** Is venous with back-eye?</b></p> <p>ITEM CODE (MS25-270)</p> <p>Dialysis Needle for Fistula and Graft, G17, <u>venous</u>, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration.</p> <p>ITEM CODE (MS25-272)</p> <p>Bicarbonate Concentrate powder, <u>8.4kg/100L concentrate (8.4%)</u>, each 100 ml of this concentrate contain 84 grams sodium hydrogen carbonate, with at least 24 months expiration, will not require technical adjustments/service prior usage. Must be compatible with 4008s hemodialysis machine.</p> <p><i>** We do not have per kilo quantity of Bicarb powder, instead can we provide a total of 10 packs of bicarb powder to match the requirement you were requested</i></p> <p>ITEM CODE (MS25-277)</p> <p>Bloodlines AV-Set <u>with drain bag</u> and macroset, blood pump segment diameter (ID) 8mm x (OD),12mm, chamber size 22mm, with at least 24 months expiration, must be compatible with the technical requirements of the existing hemodialysis machine (Fresenius), will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier, must be compatible with 4008s hemodialysis machine.</p> <p><i>**Is it ok if products we will be submitting has no drain bag? (BIOTEQUE)</i></p> | <p><b>Stick with indicates specifications. Back eye is not needed for this item.</b></p> <p>ITEM CODE (MS25-270)</p> <p>Dialysis Needle for Fistula and Graft, G17, venous, 1.6 x 25mm, with metal stainless needle, smooth insertion, with fixed wing, with at least 24 months expiration.</p> <p><b>Stick with indicates specifications. Back eye is not needed for this item.</b></p> <p>ITEM CODE (MS25-272)</p> <p>Bicarbonate Concentrate powder, 8.4kg/100L concentrate (8.4%), each 100 ml of this concentrate contain 84 grams sodium hydrogen carbonate, with at least 24 months expiration, will not require technical adjustments/service prior usage. Must be compatible with 4008s hemodialysis machine.</p> <p><b>Preparation in grams is allowed but must be equivalent to the indicated measurement in the specification.</b></p> <p>ITEM CODE (MS25-277)</p> <p>Bloodlines AV-Set with drain bag and macroset, blood pump segment diameter (ID) 8mm x (OD),12mm, chamber size 22mm, with at least 24 months expiration, must be compatible with the technical requirements of the existing hemodialysis machine (Fresenius), will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier, must be compatible with 4008s hemodialysis machine.</p> <p><b>Stick with indicated specifications. Bloodlines with drain bag is preferred.</b></p> |
|---|---|

**Other Matters:**

- A. *Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top or ring bound instead of a fastener, table of contents, and index tabs in words, not numbers.*

- B. Folder of Eligibility requirements and technical proposal should be placed in one envelope. And the folder of the Financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as **"Original Bid"**
- C. Documents should be arranged chronologically according to the checklist issued.
- D. Color code for folders and envelope:  
**YELLOW** – Public Bidding for the Supply and Delivery of Medical Supplies CY 2025 (Early Procurement Activity)
- E. All other provisions on the bidding documents which are not affected shall remain in effect.
- F. The deadline for **Submission and Opening of Bids** is scheduled on **November 13, 2024 (Wednesday), 9:00 AM**, at the **BAC Conference Room**, National Center for Mental Health Compound, Mandaluyong City.
- G. Any bid submitted after the deadline for submission shall be declared **"LATE"** and shall NOT be accepted.
- H. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.

For the information and guidance of all concerned.

  
**ALDEN C. CUYOS, MD, FRPA, IFAPA, MMHoA**  
Chairperson, BAC for Goods 