



National Center for Mental Health

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC BIDDING FOR THE SUPPLY AND DELIVERY OF MEDICAL SUPPLIES CY 2024

This Supplemental Bid Bulletin No. 1 dated **November 21, 2023**, is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

ISSUES	CLARIFICATION
<p>LIST OF ITEMS</p> <p>ITEM CODE (MS24-027) Disposable ECG Electrodes, PEDIATRIC, 32x36mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-073) Infusion set, ADULT, 20 macrodrops/mL/y-injection port/rubber/ 215 cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-074) Infusion set, PEDIA - 60 microdrops/mL/y-injection port/rubber/ 195cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-083) IV Starter Kit, includes latex free IV constricting band (18 in. x 1in.), clear adhesive hypoallergenic tape (4 in. x 4.75in.), 2x alcohol pads; 2x 2x gauze pads (2in. X 2in.); Tegaderm Transparent Film (Pav 2 ER use)</p>	<p>LIST OF ITEMS</p> <p>ITEM CODE (MS24-027) Disposable ECG Electrodes, PEDIATRIC, 32x36mm (oval) snap connection, foam support 1mm thick, hypoallergenic medical grade adhesive, pre-gelled that is not sticky to skin and with silver chloride stainless steel stud and defibrillation resistant. (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-073) Infusion set, ADULT, 20 macrodrops/mL/y-injection port/rubber/ 215 cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-074) Infusion set, PEDIA - 60 microdrops/mL/y-injection port/rubber/ 195cm length, with luer lock adapter, air vented, disposable, E.O. gas sterile / pyrogen -free (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-083) IV Starter Kit, includes 1pc latex free IV constricting band (18 in. x 1in.) / Flat Tourniquet, 1pc clear adhesive hypoallergenic tape (1/2"), 2pc alcohol pads (2inx2in); gauze pad packs of 5's, sterile; 1pc IV Film 6x8 cm, sterile, 1pc Change Label (Pav 2 ER use)</p>

"There is no Health without Mental Health"



<p>ITEM CODE (MS24-109) NEEDLE G.18 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-109) NEEDLE G.18 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam / EO sterilization. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>
<p>ITEM CODE (MS24-110) NEEDLE G.20 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-110) NEEDLE G.20 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam / EO sterilization. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>
<p>ITEM CODE (MS24-111) NEEDLE G.22 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-111) NEEDLE G.22 - Hypodermic Needles. Ultra Sharps needle, unique double bevel cut design, ultra-thin wall, Tamper Evident peel packs, sterilized by electron beam / EO sterilization. High quality, stainless-steel structure assures smooth puncture. (with at least 24 months expiration).</p>
<p>ITEM CODE (MS24-137) Syringes, disposable, 10cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-137) Syringes, disposable, 10cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-140) Syringes, disposable, 3cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-140) Syringes, disposable, 3cc, with needle G.21 x 1 inch, with luer lock, sterilized by EO gas, non-toxic, non-heavy metal and pyrogen-free, no air-leakage, needle must be sharp and sturdy. (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-173) Antiseptic surgical scrub solution. Non-ionic Surfactants, has a pleasant perfume, and Tartrazine as colours. Easy to wash off. Can be used for neonates and thyroid patients. Not degraded due to bioburden. Effective in presence of blood, serious fluid, sweat etc. Non-staining and skin friendly. (5liters/gallon) (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-173) Antiseptic surgical scrub solution. Non-ionic Surfactants, has a pleasant perfume, and Tartrazine as colours. Easy to wash off. Can be used for neonates and thyroid patients. Not degraded due to bioburden. Effective in presence of blood, serious fluid, sweat etc. Non-staining and skin friendly. (5liters/gallon) (with at least 24 months expiration). Note: RED (FD & C Red 40) coloured solution ACCEPTED. May consider our actual Unit of Measurement to its equivalent.</p>

<p>ITEM CODE (MS24-223) Liquid Hand Rub Antiseptic with 70% Ethanol and 0.5% Chlorhexidine Gluconate skin disinfectant, Bactericidal, Fungicidal and Virucidal, 5 Liters / containers. (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-223) Liquid Hand Rub Antiseptic with 70% Ethanol and 0.5% Chlorhexidine Gluconate skin disinfectant, Bactericidal, Fungicidal and Virucidal, 5 Liters / containers. (with at least 24 months expiration). *Stick with the NCMH Specifications Note: May consider our actual Unit of Measurement to its equivalent.</p>
<p>ITEM CODE (MS24-226) Nylon Cutting 5-0, P3, 26mm, 1/2c, 75cm.</p>	<p>ITEM CODE (MS24-226) Nylon Cutting 5-0, P3, 26mm, 1/2c, 75cm. *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-227) Nylon Cutting 8-0, 6.5mm, 1/4c, 30cm, spatula</p>	<p>ITEM CODE (MS24-227) Nylon Cutting 8-0, 6.0mm to 6.5mm, 1/4c to 3/8c, 30cm to 40cm, spatula or Advanced Micropoint Spatula</p>
<p>ITEM CODE (MS24-229) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-229) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-230) 'PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-230) 'PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 0, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-231) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-231) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-232) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-232) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 1, 75cm, cutting needle, 3/8 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-233) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-233) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>

<p>ITEM CODE (MS24-235) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-236) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-237) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm / 40mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-238) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm / 40mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-239) PGA (Polyglycolic Acid) 5-0 P3 Needle cutting (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-243) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-244) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-245) PGA (Polyglycolic Acid) 8-0 Round (with at least 24 months expiration). ** For item MS24-245, is it for Ophtha?</p>	<p>ITEM CODE (MS24-235) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-236) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-237) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-238) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 75cm, cutting needle, 3/8 circle, 24mm / 26mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-239) PGA (Polyglycolic Acid) 5-0 P3 Needle cutting (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-243) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 3/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-244) PGA (Polyglycolic Acid) braided, coated, braided synthetic absorbable, size 4/0, 90cm, round needle, 1/2 circle, 37mm / 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-245) PGA (Polyglycolic Acid) 8-0 Round (with at least 24 months expiration). ** Yes, item MS24-245 is for Ophtha.</p>
--	--

<p>ITEM CODE (MS24-247) Polypropylene knitted non-absorbable Mesh, undyed, 7.6cm x 15cm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-247) Polypropylene knitted non-absorbable Mesh, undyed, 7.6cm x 15cm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-248) Polypropylene monofilament, non-absorbable, size 0, 75 cm, cutting needle, 1/2 circle, 37mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-248) Polypropylene monofilament, non-absorbable, size 0, 75 cm, cutting needle, 1/2 circle, 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-249) Polypropylene monofilament, non-absorbable, size 1, 75 cm, cutting needle, 1/2 circle, 37mm – 40mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-249) Polypropylene monofilament, non-absorbable, size 1, 75 cm, cutting needle, 1/2 circle, 37mm – 40mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-250) Polypropylene monofilament, non-absorbable, size 2/0, 75 cm, cutting needle, 1/2 circle / 3/8 circle, 26mm - 37mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-250) Polypropylene monofilament, non-absorbable, size 2/0, 75 cm, cutting needle, 1/2 circle / 3/8 circle, 26mm - 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-251) Polypropylene monofilament, non-absorbable, size 3/0, 75 cm, cutting needle, 1/2 circle, 24mm / 37mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-251) Polypropylene monofilament, non-absorbable, size 3/0, 75 cm, cutting needle, 1/2 circle, 24mm / 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-252) Polypropylene monofilament, non-absorbable, size 3/0, 75 cm, cutting needle, 3/8 circle, 24mm / 37mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-252) Polypropylene monofilament, non-absorbable, size 3/0, 75 cm, cutting needle, 3/8 circle, 24mm / 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-253) Polypropylene monofilament, non-absorbable, size 4/0, 45cm-75 cm, cutting needle, 1/2 circle, 19mm - 24mm (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-253) Polypropylene monofilament, non-absorbable, size 4/0, 45cm-75 cm, cutting needle, 1/2 circle, 19mm - 24mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p>
<p>ITEM CODE (MS24-268) Silk braided, non-absorbable, size 0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).</p>	<p>ITEM CODE (MS24-268) Silk braided, non-absorbable, size 0, 60cm, 12-15 threads, no needles (with at least 24 months expiration). *Stick with the NCMH Specifications</p>

<p>ITEM CODE (MS24-270) Silk braided, non-absorbable, size 2/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-271) Silk braided, non-absorbable, size 3/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-274) Silk braided, non-absorbable, size 0, 75cm, cutting needle, 3/8 circle, 37mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-275) Silk braided, non-absorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-277) Silk braided, non-absorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-280) Silk braided, non-absorbable, size 3/0, 45cm - 75cm, cutting needle, 3/8 circle, 26mm (with at least 24 months expiration). (For OR and Dental Clinic use)</p> <p>ITEM CODE (MS24-287) Silk braided, non-absorbable, size 4/0, 75cm, round needle, 1/2 circle, 22mm / 26mm (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-339) Polyester Suture, 2-0 - 30"(75cm) 17mm, 1/2c, Tapercut</p> <p>ITEM CODE (MS24-346) Corneal knife 2.75mm Qty: 100 UOM: Box</p>	<p>ITEM CODE (MS24-270) Silk braided, non-absorbable, size 2/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-271) Silk braided, non-absorbable, size 3/0, 60cm, 12-15 threads, no needles (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-274) Silk braided, non-absorbable, size 0, 75cm, cutting needle, 3/8 circle, 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-275) Silk braided, non-absorbable, size 1, 75cm, cutting needle, 1/2 circle, 37mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-277) Silk braided, non-absorbable, size 2/0, 75cm, cutting needle, 1/2 circle, 24mm / 26mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-280) Silk braided, non-absorbable, size 3/0, 45cm - 75cm, cutting needle, 3/8 circle, 26mm (with at least 24 months expiration). (For OR and Dental Clinic use) *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-287) Silk braided, non-absorbable, size 4/0, 75cm, round needle, 1/2 circle, 22mm / 26mm (with at least 24 months expiration). *Stick with the NCMH Specifications</p> <p>ITEM CODE (MS24-339) Polyester Suture, 2-0 - 30"(75cm) 17mm to 20mm, 1/2c, Tapercut</p> <p>ITEM CODE (MS24-346) Corneal knife 2.75mm Qty: 100 UOM: Piece</p>
--	---

<p>ITEM CODE (MS24-347) Corneal Stab knife 15 degree Qty: 100 UOM: Box</p> <p>ITEM CODE (MS24-351) PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 6/0, 75cm, cutting needle, 0.7 metric, 18", 45cm-76cm, 1/4c, spatulated (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-353) Sodium Hyaluronate at least 1.5% viscoelastic eye gel, 1mL Pre-filled syringe</p> <p>ITEM CODE (MS24-354) Sodium Hyaluronate 1.5% viscoelastic eye gel, 1mL</p> <p>ITEM CODE (MS24-356) Phacoemulsification basic pack (must be compatible with Stellaris Phaco machine)</p> <p>** For Phacoemulsification basic pack, is it consumables catpack or cassette for machine of sterallis?</p> <p>ITEM CODE (MS24-364) Bicarbonate Concentrate granules, 8.4kg/100L concentrate (8.4%), each 100 ml of this concentrate contain 84 grams sodium hydrogen carbonate, with at least 24 months expiration, will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine) QTY: 610 UOM: packs</p> <p>** We only serve 8.4 grams per pack of Bicarbonate Concentrate. Is it ok if we will provide ten (10) packs so that we will match your requirement.</p> <p>** Can NCMH adjust the quantity of Bicarbonate Concentrates granules from 610 to 612 packs?</p>	<p>ITEM CODE (MS24-347) Corneal Stab knife 15 degree Qty: 100 UOM: Piece</p> <p>ITEM CODE (MS24-351) PGA (Polyglycolic Acid/Polyglactin) braided, coated, braided synthetic absorbable, size 6/0, 75cm, cutting needle, 8mm to 8.6mm, 0.7 metric, 18", 30cm-45cm, 1/4c, spatulated (with at least 24 months expiration).</p> <p>ITEM CODE (MS24-353) Sodium Hyaluronate at least 3% Chondroitin sulfate 4% viscoelastic eye gel, 0.5mL</p> <p>ITEM CODE (MS24-354) Sodium Hyaluronate 1.5% to 1.8% viscoelastic eye gel, 1mL Pre-filled syringe</p> <p>ITEM CODE (MS24-356) Phacoemulsification basic pack (must be compatible with Stellaris Phaco machine)</p> <p>** Consumables Phaco pack for Sterallis.</p> <p>ITEM CODE (MS24-364) Bicarbonate Concentrate granules, 8.4kg/100L concentrate (8.4%), each 100 ml of this concentrate contain 84 grams sodium hydrogen carbonate, with at least 24 months expiration, will not require technical adjustments/service prior usage, so as not to void warranty of equipment supplier. (must be compatible with 4008s Hemodialysis Machine) QTY: 610 UOM: packs</p> <p>**Stick with the NCMH Specifications and quantity. <u>May consider our actual Unit of Measurement to its equivalent.</u></p>
---	--


<p>** We have newly distributed products but we are not sure if those brands are already circulating in the Philippine market for two (2) years. May I ask what are the necessary requirements for us to submit in order for us to join.</p> <p>ITEM CODE (MS24-395) BATHROOM WEIGHING SCALE HEAVY DUTY - made with non-skid vinyl mat features white, baked enamel finish battery-free mechanized operation analogue. Capability: 150kg maximum 500g graduation, sturdy, big metallic non-slip podium, easy-to-read dial, used to monitor body weight. Dimension: 32cm x 17cm x 40cm. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.</p>	<p>** NO required two (2) years in the Philippine market. Please see Section III BID DATA SHEET for the list of requirements or additional documents relevant to the Project.</p> <p>ITEM CODE (MS24-395) BATHROOM WEIGHING SCALE HEAVY DUTY - made with non-skid vinyl mat features white, baked enamel finish battery-free mechanized operation analogue. Capability: 136kg maximum 500g graduation, sturdy, big metallic non-slip podium, easy-to-read dial, used to monitor body weight. Dimension: 32cm x 17cm x 40cm. Provide Calibration Certificate for each unit and Warranty Certificate 1 (ONE) year for parts and services which include Preventive, Corrective Maintenance and Calibration from the date of delivery, inspection and acceptance.</p>
<p>Section III BID DATA SHEET</p> <p>ITB Clause 5.3 For this purpose, contracts similar to the Project shall be:</p> <ul style="list-style-type: none"> a. <i>Supply and Delivery of Medical Supplies</i> b. Completed within <i>two (2) years</i> prior to the deadline for the submission and receipts of bids. <p>***"If SLCC can be change from 2 years to 3 years prior to the deadline of submission and receipt of bids?"</p>	<p>Section III BID DATA SHEET</p> <p>ITB Clause 5.3 For this purpose, contracts similar to the Project shall be:</p> <ul style="list-style-type: none"> a. <i>Supply and Delivery of Medical Supplies</i> b. Completed within <i>two (2) years</i> prior to the deadline for the submission and receipts of bids. <p>** <u>Stick to two (2) years.</u></p>
<p>Section II INSTRUCTION TO BIDDERS</p> <p>xx 5. Eligible Bidders xx d.</p> <ul style="list-style-type: none"> i. Completed at least two (2) similar contracts, the aggregate amount of which should be equivalent to at least fifty percent (50%) in the case of non-expendable supplies and services of twenty-five (25%) percent in the case of expendable supplies of the ABC for this Project <p>xx</p>	<p>Section II INSTRUCTION TO BIDDERS</p> <p>xx 5. Eligible Bidders xx d.</p> <ul style="list-style-type: none"> i. Completed <u>at least two (2) similar contracts</u>, the aggregate amount of which should be equivalent to at least fifty percent (50%) in the case of non-expendable supplies and services of twenty-five (25%) percent in the case of expendable supplies of the ABC for this Project <p>xx</p>

<p>** "Is it allowed to submit more than two similar contracts aggregate to the 25% of the ABC?"</p>	<p>In accordance with Section 23.4.1.3 of RA 9184 and its 2016 revised IRR:</p> <p>"The bidder must have completed, within the period specified in the Invitation to Bid, an SLCC that is similar to the contract to be bid, and whose value, adjusted to current prices using the Philippine Statistics Authority (PSA) consumer price indices, must be at least fifty percent (50%) of the ABC. However, in the case of Expendable Supplies, said SLCC must be at least twenty five percent (25%) of the ABC.</p> <p>xx</p> <p>(a) <u>The bidder should have completed at least two (2) similar contracts and the aggregate contract amounts should be equivalent to at least the percentage of the ABC as required above; and</u></p> <p>(b) <u>The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.</u></p>
<p>SECTION I INVITATION TO BID</p> <p>8. Bid opening shall be on November 29, 2023 (Wednesday), 1:30 pm at the Pag-asa Hall of the National Center for Mental Health. Bids will be opened in the presence of bidders' representatives who choose to attend the activity.</p>	<p>SECTION I INVITATION TO BID</p> <p>8. Bid opening shall be on November 29, 2023 (Wednesday), 9:00 AM at the Pag-asa Hall of the National Center for Mental Health. Bids will be opened in the presence of bidders' representatives who choose to attend the activity.</p>
<p>CHECKLIST FOR BIDDERS</p> <p>I. TECHNICAL COMPONENT ENVELOPE</p> <p>B. TECHNICAL DOCUMENTS</p> <p>xx</p> <p>(6) Statement of the Bidder's SINGLE LARGEST COMPLETED CONTRACT (SLCC)* <u>similar</u> to the contract to be bid, in accordance with ITB Clause 5.4 and using the form prescribed in Annex: Bidding Forms.</p> <p>** We would like to ask if we can possibly use a Consignment for Medical / Surgical supplies contract which is already completed for the SLCC.</p>	<p>CHECKLIST FOR BIDDERS</p> <p>I. TECHNICAL COMPONENT ENVELOPE</p> <p>B. TECHNICAL DOCUMENTS</p> <p>xx</p> <p>(6) Statement of the Bidder's SINGLE LARGEST COMPLETED CONTRACT (SLCC)* <u>similar</u> to the contract to be bid, in accordance with ITB Clause 5.4 and using the form prescribed in Annex: Bidding Forms.</p> <p>** <u>The Bids and Awards Committee is precluded from pre-qualifying the Bidders at this stage.</u></p>

Other Matters:

- A. *Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top or ring bound instead of a fastener, table of contents, and index tabs in words, not numbers.*
- B. *Folder of Eligibility requirements and technical proposal should be placed in one envelope. And the folder of the Financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as **"Original Bid"***
- C. Documents should be arranged chronologically according to the checklist issued.
- D. Color code for folders and envelope:
YELLOW – Public Bidding for the Supply and Delivery of Medical Supplies CY 2024
- E. All other provisions on the bidding documents which are not affected shall remain in effect.
- F. The deadline for **Submission and Opening of Bids** is scheduled on **November 29, 2023 (Wednesday), 9:00 AM,** at the **Pag-Asa Hall**, National Center for Mental Health Compound, Mandaluyong City.
- G. Any bid submitted after the deadline for submission shall be declared **"LATE"** and shall NOT be accepted.
- H. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.

For the information and guidance of all concerned.


ALDEN C. CUYOS, MD, FPPA, IFAPA, MMHoA
Chairman, BAC for Goods 