

SUPPLEMENTAL BID BULLETIN

ADDENDUM NO. 1

PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONG OF BRAND-NEW OPERATING ROOM TABLE FOR PAVILION 7-GENERAL HOSPITAL CY 2024

This Supplemental Bid Bulletin No. 1 dated **December 04, 2024** is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

A. Amendments / Clarifications:

A. Amendments / Clarifications:	
FROM	TO
ON SECTION III: BID DATA SHEET / LIST OF ITEMS	ON SECTION III: BID DATA SHEET / LIST OF ITEMS
p. Notarized certification from the manufacturer authenticated by the Philippines consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor.	 p. Notarized certification from the manufacturer authenticated by the Philippines consulate from the country of origin of the unit that the warranty should not be affected with a change of distributor. > "Provision will be retained"
ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS	ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS
Under OR-01-2024: Supply, Delivery, Installation, Testing and Commissioning of Brand-New Operating Room Table for Pavilion 7 CY 2024	Under OR-01-2024: Supply, Delivery, Installation, Testing and Commissioning of Brand-New Operating Room Table for Pavilion 7 CY 2024
OPERATING ROOM TABLE FOR VARIOUS SURGICAL DISCIPLINE WITH ORTHOPEDIC TRACTION DEVICE	OPERATING ROOM TABLE FOR VARIOUS SURGICAL DISCIPLINE WITH ORTHOPEDIC TRACTION DEVICE
GENERAL DESCRIPTION ■ Safe working load of at least 450kg guaranteeing highest safety and stability.	GENERAL DESCRIPTION ■ Safe working load of at least 400kg, guaranteeing highest safety and stability.
 Equip with system that shall recognize the type of the table component and match the anti-collision data automatically. 	 Equip with system that shall recognize the type of the table component and match the anti-collision data automatically. (Optional)
 Equip with warning/notification indication technology to show the angle and issue a timely warning when it comes to excessive Trendelenburg and reverse Trendelenburg positions. 	 Equip with warning/notification <u>technology</u> <u>to prevent excessive angle movement in</u> <u>the Trendelenburg</u> and reverse Trendelenburg positions.

"There is no Health without Mental Health"












The table pad should be double layered and The table pad should be thick if (single) or not soft but can be molded by the figure of the double layered and not soft but can be patient to deliver even counterforce and molded by the figure of the patient to deliver reduce the possibility of ulcer, and it should even counterforce and reduce the possibility be water-proof and anti-static material and of ulcer, and it should be water-proof and can be washed by water directly; each joint anti-static material and can be washed by should be sealed by ultrasonic not glue and water directly; each joint should be sealed by sewing. ultrasonic not glue and sewing. It shall be battery powered and electric It shall be battery powered and electric /electro-hydraulic or electro mechanical /electro-hydraulic operating table with battery operating table with battery inside the table, inside the table, which can work at least 50which can work at least 50-80 operations for 80 operations for two weeks and the battery two weeks and the battery should be should be standard configuration. standard configuration. Equip with table positions memory that can with table memory position Equip save at least 15 positions, so the user can technology or its equivalent (Optional). select the memorized positions for the specific surgery freely without complex adjustment. Equip with electric-hydraulic system that Equip with electro-hydraulic or electro includes table up and down, back up and mechanical system that includes table up down, turn left and right, Trendelenburg and and down, back up and down, turn left and reverse Trendelenburg, leg plate up & down, Trendelenburg right, and reverse which shall be controlled by hand control Trendelenburg, leg plate up & down, which panel. shall be controlled by hand control panel and/or remote control. **TECHNICAL SPECIFICATION TECHNICAL SPECIFICATION Basic Information Basic Information** Width of the table with rails at least 570 Width of the table with rails at least 570mm. mm to 590 mm. **Electric Function Electric Function** The lowest position at least 595 mm to 600 The lowest position at 660 mm or lower. Tilting position for both left and right must be Tilting position for both left and right must be minimum of 26°. minimum of 25°. Trendelenburg position Trendelenburg position and Reverse and Reverse Trendelenburg position must be minimum of Trendelenburg position must be minimum of 36°. 30°. Back plate up position 75° to 90°. Back plate up position at least 75°. Back plate down position 40° to 45°. Back plate down position at least 40°. Leg plate down 90° to 105°. Leg plate down at least 90°.

Other Matters:

- A. Bid Security shall be 120 calendar days from the date of submission and opening of bids.
- B. Section VII: Technical Specifications Please write "COMPLY" to indicate that requirements are met.
- C. Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with <u>shoelace</u> on top instead of fastener, <u>table of contents and index tabs in words</u>, not numbers.



- D. Folder of Eligibility requirements and technical proposal should be placed in one envelope. And folder of Financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as "Original Bid"
- E. Documents should be arranged chronologically according to the checklist issued.
- F. Color code for folders and envelope: GREEN
- G. Typographical error must be counter signed by the authorized representative.
- H. All other provisions on the bidding documents which are not affected shall remain in effect.
- I. The deadline for **Submission and Opening of Bids** is scheduled on **December 11, 2024**, <u>1:30 PM</u>, at the **BAC Conference Room**, National Center for Mental Health Compound, Mandaluyong City.
- J. Any bid submitted after the deadline for submission shall be declared "LATE" and shall NOT be accepted.
- K. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.
- L. Bidder is required for equipment demonstration within 10 calendar days upon receipt of Notice of Eligibility.

For the information and guidance of all concerned.

JERRY C. RODRIGUEZ, MGM-ESP

Chairperson, Bids & Awards Committee CY2024