



**SUPPLEMENTAL BID BULLETIN**

**ADDENDUM NO. 1**

**PUBLIC BIDDING FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF BRAND-NEW ANESTHESIA MACHINE FOR PAVILION 7- GENERAL HOSPITAL AND ECT MACHINE OF PAVILION 1 & 3 CY 2024**

This Supplemental Bid Bulletin No. 1 dated **December 16, 2024** is issued to clarify, modify or amend items in the Bid Documents. This shall form an integral part of the Bid Documents.

**A. Amendments / Clarifications:**

ISSUE	CLARIFICATION
<p><b>ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS</b></p> <p><b>Under AM-01-2024: ANESTHESIA MACHINE WITH PATIENT MONITOR AND PRONE POSITIONING SYSTEM</b></p> <p><b>A. MINIMUM FEATURES</b></p> <ul style="list-style-type: none"> <li>• <b>Required Parameters for Monitoring</b> <ul style="list-style-type: none"> <li>○ A hypoxic guard system must protect against less than 21% inspired oxygen if nitrous oxide is in use.</li> </ul> </li> </ul>	<p><b>ON LIST OF ITEMS / TECHNICAL SPECIFICATIONS</b></p> <p><b>Under AM-01-2024: ANESTHESIA MACHINE WITH PATIENT MONITOR AND PRONE POSITIONING SYSTEM</b></p> <p><b>B. MINIMUM FEATURES</b></p> <ul style="list-style-type: none"> <li>• <b>Required Parameters for Monitoring</b> <ul style="list-style-type: none"> <li>○ A hypoxic guard system must protect against less than 21% inspired oxygen if nitrous oxide is in use.</li> </ul> </li> </ul> <p><b>"Provision will be retained"</b></p>
<ul style="list-style-type: none"> <li>○ Pulse oximetry, blood pressure monitoring, and EKG are required.</li> </ul>	<p><del>○ Pulse oximetry, blood pressure monitoring, and EKG are required.</del></p> <p><b>"Not required"</b></p>
<ul style="list-style-type: none"> <li>• Pressure in the breathing circuit is limited to <b><u>12.5 kPa (125 cm water)</u></b>.</li> </ul>	<ul style="list-style-type: none"> <li>• Pressure in the breathing circuit is limited to <b><u>at least 120 cmH2O to 125 cmH2O.</u></b></li> </ul>
<ul style="list-style-type: none"> <li>• The hanger yoke must be pin-indexed, have a clamping device that resists leaks, and contain a filter. It must have a check valve to prevent transfilling, and a cylinder pressure gauge. There must be cylinder pressure regulators. The machine must use pipeline gas as long as pipeline pressure is <b><u>greater than 345 kPa (50 psi)</u></b>.</li> </ul>	<ul style="list-style-type: none"> <li>• The hanger yoke must be pin-indexed, have a clamping device that resists leaks, and contain a filter. It must have a check valve to prevent transfilling, and a cylinder pressure gauge. There must be cylinder pressure regulators. The machine must use pipeline gas as long as pipeline pressure is <b><u>at least 280 Kpa to 345 Kpa or greater.</u></b></li> </ul>

*"There is no Health without Mental Health"*

<ul style="list-style-type: none"> <li>An oxygen flush is required, capable of 35-75 L/min flow which does not proceed through any vaporizers.</li> </ul>	<ul style="list-style-type: none"> <li>An oxygen flush is required, capable of 35-75 L/min flow which does not proceed through any vaporizers.</li> </ul> <p style="text-align: center;"><b><i>“Provision will be retained”</i></b></p>
<p><b>ANESTHESIA MACHINE WITH PATIENT MONITOR AND <u>PRONE POSITIONING SYSTEM</u></b></p>	<p><b>ANESTHESIA MACHINE WITH PATIENT MONITOR AND <u>PRONE POSITIONING SYSTEM</u></b></p> <p><b><u>PRONE POSITIONING SYSTEM (manufacturer's standards)</u></b></p>

**Other Matters:**

- A. Bid Security shall be 120 calendar days from the date of submission and opening of bids.
- B. Section VII: Technical Specifications - Please write “COMPLY” to indicate that requirements are met.
- C. *Eligibility requirements and technical proposal should be in one folder and financial proposal in a separate folder, with shoelace on top instead of fastener, table of contents and index tabs in words, not numbers.*
- D. *Folder of Eligibility requirements and technical proposal should be placed in one envelope. And folder of Financial proposal should be in another envelope. Both envelopes shall then be placed in one mother envelope marked as “Original Bid”*
- E. Documents should be arranged chronologically according to the checklist issued.
- F. Color code for folders and envelope: **GREEN**
- G. Typographical error must be counter signed by the authorized representative.
- H. All other provisions on the bidding documents which are not affected shall remain in effect.
- I. The deadline for **Submission and Opening of Bids** is scheduled on **December 23, 2024, 9:00 AM**, at the **BAC Conference Room**, National Center for Mental Health Compound, Mandaluyong City.
- J. Any bid submitted after the deadline for submission shall be declared “LATE” and shall NOT be accepted.
- K. The BAC shall open the bids immediately after the deadline for submission and receipt of bids.
- L. Bidder is required for equipment demonstration within 10 calendar days upon receipt of Notice of Eligibility.

For the information and guidance of all concerned.

  
**JERRY C. RODRIGUEZ, MGM-ESP**  
 - Chairperson, Bids & Awards Committee CY2024 *gr*